
PyMeasure Documentation

Release 0.7.0


PyMeasure Developers

Jan 18, 2020

LEARNING PYMEASURE

1	Introduction	3
1.1	Instrument ready	3
1.2	Graphical displays	3
2	Quick start	5
2.1	Setting up Python	5
2.2	Installing PyMeasure	5
3	Tutorials	7
3.1	Connecting to an instrument	7
3.2	Making a measurement	8
3.3	Using a graphical interface	15
4	pymeasure.adapters	25
4.1	Adapter base class	25
4.2	Fake adapter	26
4.3	Serial adapter	27
4.4	Prologix adapter	28
4.5	VISA adapter	29
4.6	VXI-11 adapter	31
5	pymeasure.experiment	33
5.1	Experiment class	33
5.2	Listener class	34
5.3	Procedure class	35
5.4	Parameter classes	36
5.5	Worker class	38
5.6	Results class	39
6	pymeasure.display	41
6.1	Browser classes	41
6.2	Curves classes	41
6.3	Inputs classes	42
6.4	Listeners classes	43
6.5	Log classes	44
6.6	Manager classes	44
6.7	Plotter class	45
6.8	Qt classes	45
6.9	Thread classes	46
6.10	Widget classes	46
6.11	Windows classes	47

7	pymeasure.instruments	49
7.1	Instrument classes	49
7.2	Validator functions	52
7.3	Comedi data acquisition	53
7.4	Resource Manager	53
7.5	Advantest	54
7.6	Agilent	54
7.7	AMI	72
7.8	Anritsu	73
7.9	Danfysik	76
7.10	F.W. Bell	79
7.11	Hewlett Packard	80
7.12	Keithley	81
7.13	Lake Shore Cryogenics	92
7.14	Newport	95
7.15	Oxford Instruments	96
7.16	Parker	98
7.17	Signal Recovery	100
7.18	Stanford Research Systems	101
7.19	Tektronix	104
7.20	Thorlabs	104
7.21	Yokogawa	105
8	Contributing	107
8.1	Using the development version	107
8.2	Working on a new feature	108
8.3	Making a pull-request	108
8.4	Unit testing	108
9	Reporting an error	111
10	Adding instruments	113
10.1	File structure	113
10.2	Instrument file	114
10.3	Writing properties	115
10.4	Advanced properties	115
11	Coding Standards	119
11.1	Python style guides	119
11.2	Documentation	119
11.3	Usage of getter and setter functions	119
12	Authors	121
13	License	123
	Python Module Index	125
	Index	127



PyMeasure

PyMeasure makes scientific measurements easy to set up and run. The package contains a repository of instrument classes and a system for running experiment procedures, which provides graphical interfaces for graphing live data and managing queues of experiments. Both parts of the package are independent, and when combined provide all the necessary requirements for advanced measurements with only limited coding.

Installing Python and PyMeasure are demonstrated in the [Quick Start guide](#). From there, checkout the existing [instruments that are available for use](#).

PyMeasure is currently under active development, so please report any issues you experience on our [Issues page](#).

The main documentation for the site is organized into a couple sections:

- [Learning PyMeasure](#)
- [API References](#)
- [About PyMeasure](#)

Information about development is also available:

- [Getting involved](#)

INTRODUCTION

PyMeasure uses an object-oriented approach for communicating with scientific instruments, which provides an intuitive interface where the low-level SCPI and GPIB commands are hidden from normal use. Users can focus on solving the measurement problems at hand, instead of re-inventing how to communicate with instruments.

Instruments with VISA (GPIB, Serial, etc) are supported through the [PyVISA package](#) under the hood. [Prologix GPIB](#) adapters are also supported. Communication protocols can be swapped, so that instrument classes can be used with all supported protocols interchangeably.

Before using PyMeasure, you may find it helpful to be acquainted with [basic Python programming for the sciences](#) and understand the concept of objects.

1.1 Instrument ready

The package includes a number of *instruments already defined*. Their definitions are organized based on the manufacturer name of the instrument. For example the class that defines the *Keithley 2400 SourceMeter* can be imported by calling:


```
from pymeasure.instruments.keithley import Keithley2400
```

The *Tutorials* section will go into more detail on *connecting to an instrument*. If you don't find the instrument you are looking for, but are interested in contributing, see the documentation on *adding an instrument*.

1.2 Graphical displays

Graphical user interfaces (GUIs) can be easily generated to manage execution of measurement procedures with PyMeasure. This includes live plotting for data, and a queue system for managing large numbers of experiments.

These features are explored in the *Using a graphical interface* tutorial.



QUICK START

This section provides instructions for getting up and running quickly with PyMeasure.

2.1 Setting up Python

The easiest way to install the necessary Python environment for PyMeasure is through the [Anaconda distribution](#), which includes 720 scientific packages. The advantage of using this approach over just relying on the `pip` installer is that it Anaconda correctly installs the required Qt libraries.

Download and install the appropriate Python 3.5 version of [Anaconda](#) for your operating system.

2.2 Installing PyMeasure

2.2.1 Install with conda

If you have the [Anaconda distribution](#) you can use the conda package mangager to easily install PyMeasure and all required dependencies.

Open a terminal and type the following commands (on Windows look for the *Anaconda Prompt* in the Start Menu):

```
conda config --add channels conda-forge
conda install pymeasure
```

This will install PyMeasure and all the required dependencies.

2.2.2 Install with pip

PyMeasure can also be installed with `pip`.

```
pip install pymeasure
```

Depending on your operating system, using this method may require additional work to install the required dependencies, which include the Qt libraries.

2.2.3 Checking the version

Now that you have Python and PyMeasure installed, open up a “Jupyter Notebook” to test which version you have installed. Execute the following code into a notebook cell.

```
import pymeasure
pymeasure.__version__
```

You should see the version of PyMeasure printed out. At this point you have PyMeasure installed, and you are ready to start using it! Are you ready to *connect to an instrument*?

TUTORIALS

The following sections provide instructions for getting started with PyMeasure.

3.1 Connecting to an instrument

After following the *Quick Start* section, you now have a working installation of PyMeasure. This section describes connecting to an instrument, using a Keithley 2400 SourceMeter as an example. To follow the tutorial, open a command prompt, IPython terminal, or Jupyter notebook.

First import the instrument of interest.

```
from pymeasure.instruments.keithley import Keithley2400
```

Then construct an object by passing the GPIB address. For this example we connect to the instrument over GPIB (using VISA) with an address of 4. See the *adapters* section below for more details.

```
sourcemeter = Keithley2400("GPIB::4")
```

For instruments with standard SCPI commands, an `id` property will return the results of a `*IDN?` SCPI command, identifying the instrument.

```
sourcemeter.id
```

This is equivalent to manually calling the SCPI command.

```
sourcemeter.ask("*IDN?")
```

Here the `ask` method writes the SCPI command, reads the result, and returns that result. This is further equivalent to calling the methods below.

```
sourcemeter.write("*IDN?")  
sourcemeter.read()
```

This example illustrates that the top-level methods like `id` are really composed of many lower-level methods. Both can be called depending on the operation that is desired. PyMeasure hides the complexity of these lower-level operations, so you can focus on the bigger picture.

3.1.1 Using adapters

PyMeasure supports a number of adapters, which are responsible for communicating with the underlying hardware. In the example above, we passed the string “GPIB::4” when constructing the instrument. By default this constructs a `VISAAdapter` class to connect to the instrument using VISA. Instead of passing a string, we could equally pass an adapter object.

```
from pymeasure.adapters import VISAAdapter

adapter = VISAAdapter("GPIB::4")
sourcemeter = Keithley2400(adapter)
```

To instead use a Prologix GPIB device connected on `/dev/ttyUSB0` (proper permissions are needed in Linux, see [PrologixAdapter](#)), the adapter is constructed in a similar way. Unlike the VISA adapter which is specific to each instrument, the Prologix adapter can be shared by many instruments. Therefore, they are addressed separately based on the GPIB address number when passing the adapter into the instrument construction.

```
from pymeasure.adapters import PrologixAdapter

adapter = PrologixAdapter('/dev/ttyUSB0')
sourcemeter = Keithley2400(adapter.gpib(4))
```

For instruments using serial communication that have particular settings that need to be matched, a custom `Adapter` sub-class can be made. For example, the LakeShore 425 Gaussmeter connects via USB, but uses particular serial communication settings. Therefore, a `LakeShoreUSBAdapter` class enables these requirements in the background.

```
from pymeasure.instruments.lakeshore import LakeShore425

gaussmeter = LakeShore425('/dev/lakeshore425')
```

Behind the scenes the `/dev/lakeshore425` port is passed to the `LakeShoreUSBAdapter`.

Some equipment may require the vxi-11 protocol for communication. An example would be a Agilent E5810B ethernet to GPIB bridge. To use this type equipment the `python-vxi11` library has to be installed which is part of the extras package requirements.

```
from pymeasure.adapters import VXI11Adapter
from pymeasure.instruments import Instrument

adapter = VXI11Adapter("TCPIP::192.168.0.100::inst0::INSTR")
instr = Instrument(adapter, "my_instrument")
```

The above examples illustrate different methods for communicating with instruments, using adapters to keep instrument code independent from the communication protocols. Next we present the methods for setting up measurements.

3.2 Making a measurement

This tutorial will walk you through using PyMeasure to acquire a current-voltage (IV) characteristic using a Keithley 2400. Even if you don’t have access to this instrument, this tutorial will explain the method for making measurements with PyMeasure. First we describe using a simple script to make the measurement. From there, we show how `Procedure` objects greatly simplify the workflow, which leads to making the measurement with a graphical interface.

3.2.1 Using scripts

Scripts are a quick way to get up and running with a measurement in PyMeasure. For our IV characteristic measurement, we perform the following steps:

- 1) Import the necessary packages
- 2) Set the input parameters to define the measurement
- 3) Connect to the Keithley 2400
- 4) Set up the instrument for the IV characteristic
- 5) Allocate arrays to store the resulting measurements
- 6) Loop through the current points, measure the voltage, and record
- 7) Save the final data to a CSV file
- 8) Shutdown the instrument

These steps are expressed in code as follows.

```
# Import necessary packages
from pymeasure.instruments.keithley import Keithley2400
import numpy as np
import pandas as pd
from time import sleep

# Set the input parameters
data_points = 50
averages = 50
max_current = 0.01
min_current = -max_current

# Connect and configure the instrument
sourceterm = Keithley2400("GPIB::4")
sourceterm.reset()
sourceterm.use_front_terminals()
sourceterm.measure_voltage()
sourceterm.config_current_source()
sleep(0.1) # wait here to give the instrument time to react
sourceterm.set_buffer(averages)

# Allocate arrays to store the measurement results
currents = np.linspace(min_current, max_current, num=data_points)
voltages = np.zeros_like(currents)
voltage_stds = np.zeros_like(currents)

# Loop through each current point, measure and record the voltage
for i in range(data_points):
    sourceterm.current = currents[i]
    sourceterm.reset_buffer()
    sleep(0.1)
    sourceterm.start_buffer()
    sourceterm.wait_for_buffer()

    # Record the average and standard deviation
    voltages[i] = sourceterm.means
    voltage_stds[i] = sourceterm.standard_devs
```

(continues on next page)

(continued from previous page)

```
# Save the data columns in a CSV file
data = pd.DataFrame({
    'Current (A)': currents,
    'Voltage (V)': voltages,
    'Voltage Std (V)': voltage_stds,
})
data.to_csv('example.csv')

sourcemeter.shutdown()
```

Running this example script will execute the measurement and save the data to a CSV file. While this may be sufficient for very basic measurements, this example illustrates a number of issues that PyMeasure solves. The issues with the script example include:

- The progress of the measurement is not transparent
- Input parameters are not associated with the data that is saved
- Data is not plotted during the execution (nor at all in this case)
- Data is only saved upon successful completion, which is otherwise lost
- Canceling a running measurement causes the system to end in an undetermined state
- Exceptions also end the system in an undetermined state

The *Procedure* class allows us to solve all of these issues. The next section introduces the *Procedure* class and shows how to modify our script example to take advantage of these features.

3.2.2 Using Procedures

The Procedure object bundles the sequence of steps in an experiment with the parameters required for its successful execution. This simple structure comes with huge benefits, since a number of convenient tools for making the measurement use this common interface.

Let's start with a simple example of a procedure which loops over a certain number of iterations. We make the SimpleProcedure object as a sub-class of Procedure, since SimpleProcedure *is a* Procedure.

```
from time import sleep
from pymeasure.experiment import Procedure
from pymeasure.experiment import IntegerParameter

class SimpleProcedure(Procedure):

    # a Parameter that defines the number of loop iterations
    iterations = IntegerParameter('Loop Iterations')

    # a list defining the order and appearance of columns in our data file
    DATA_COLUMNS = ['Iteration']

    def execute(self):
        """ Loops over each iteration and emits the current iteration,
        before waiting for 0.01 sec, and then checking if the procedure
        should stop
        """
        for i in range(self.iterations):
            self.emit('results', {'Iteration': i})
            sleep(0.01)
```

(continues on next page)

(continued from previous page)

```

if self.should_stop():
    break

```

At the top of the `SimpleProcedure` class we define the required Parameters. In this case, `iterations` is a `IntegerParameter` that defines the number of loops to perform. Inside our Procedure class we reference the value in the `iterations` Parameter by the class variable where the Parameter is stored (`self.iterations`). PyMeasure swaps out the Parameters with their values behind the scene, which makes accessing the values of parameters very convenient.

We define the data columns that will be recorded in a list stored in `DATA_COLUMNS`. This sets the order by which columns are stored in the file. In this example, we will store the Iteration number for each loop iteration.

The `execute` methods defines the main body of the procedure. Our example method consists of a loop over the number of iterations, in which we emit the data to be recorded (the Iteration number). The data is broadcast to any number of listeners by using the `emit` method, which takes a topic as the first argument. Data with the `'results'` topic and the proper data columns will be recorded to a file. The sleep function in our example provides two very useful features. The first is to delay the execution of the next lines of code by the `time` argument in units of seconds. The second is that during this delay time, the CPU is free to perform other code. Successful measurements often require the intelligent use of sleep to deal with instrument delays and ensure that the CPU is not hogged by a single script. After our delay, we check to see if the Procedure should stop by calling `self.should_stop()`. By checking this flag, the Procedure will react to a user canceling the procedure execution.

This covers the basic requirements of a Procedure object. Now let's construct our `SimpleProcedure` object with 100 iterations.

```

procedure = SimpleProcedure()
procedure.iterations = 100

```

Next we will show how to run the procedure.

Running Procedures

A Procedure is run by a Worker object. The Worker executes the Procedure in a separate Python thread, which allows other code to execute in parallel to the procedure (e.g. a graphical user interface). In addition to performing the measurement, the Worker spawns a Recorder object, which listens for the `'results'` topic in data emitted by the Procedure, and writes those lines to a data file. The Results object provides a convenient abstraction to keep track of where the data should be stored, the data in an accessible form, and the Procedure that pertains to those results.

We first construct a Results object for our Procedure.

```

from pymeasure.experiment import Results

data_filename = 'example.csv'
results = Results(procedure, data_filename)

```

Constructing the Results object for our Procedure creates the file using the `data_filename`, and stores the Parameters for the Procedure. This allows the Procedure and Results objects to be reconstructed later simply by loading the file using `Results.load(data_filename)`. The Parameters in the file are easily readable.

We now construct a Worker with the Results object, since it contains our Procedure.

```

from pymeasure.experiment import Worker

worker = Worker(results)

```

The Worker publishes data and other run-time information through specific queues, but can also publish this information over the local network on a specific TCP port (using the optional `port` argument). Using TCP communication

allows great flexibility for sharing information with Listener objects. Queues are used as the standard communication method because they preserve the data order, which is of critical importance to storing data accurately and reacting to the measurement status in order.

Now we are ready to start the worker.

```
worker.start()
```

This method starts the worker in a separate Python thread, which allows us to perform other tasks while it is running. When writing a script that should block (wait for the Worker to finish), we need to join the Worker back into the main thread.

```
worker.join(timeout=3600) # wait at most 1 hr (3600 sec)
```

Let's put all the pieces together. Our SimpleProcedure can be run in a script by the following.

```
from time import sleep
from pymeasure.experiment import Procedure, Results, Worker
from pymeasure.experiment import IntegerParameter

class SimpleProcedure(Procedure):

    # a Parameter that defines the number of loop iterations
    iterations = IntegerParameter('Loop Iterations')

    # a list defining the order and appearance of columns in our data file
    DATA_COLUMNS = ['Iteration']

    def execute(self):
        """ Loops over each iteration and emits the current iteration,
        before waiting for 0.01 sec, and then checking if the procedure
        should stop
        """
        for i in range(self.iterations):
            self.emit('results', {'Iteration': i})
            sleep(0.01)
            if self.should_stop():
                break

if __name__ == "__main__":
    procedure = SimpleProcedure()
    procedure.iterations = 100

    data_filename = 'example.csv'
    results = Results(procedure, data_filename)

    worker = Worker(results)
    worker.start()

    worker.join(timeout=3600) # wait at most 1 hr (3600 sec)
```

Here we have included an if statement to only run the script if the `__name__` is `__main__`. This precaution allows us to import the SimpleProcedure object without running the execution.

Using Logs

Logs keep track of important details in the execution of a procedure. We describe the use of the Python logging module with PyMeasure, which makes it easy to document the execution of a procedure and provides useful insight when diagnosing issues or bugs.

Let's extend our SimpleProcedure with logging.

```
import logging
log = logging.getLogger(__name__)
log.addHandler(logging.NullHandler())

from time import sleep
from pymeasure.log import console_log
from pymeasure.experiment import Procedure, Results, Worker
from pymeasure.experiment import IntegerParameter

class SimpleProcedure(Procedure):

    iterations = IntegerParameter('Loop Iterations')

    DATA_COLUMNS = ['Iteration']

    def execute(self):
        log.info("Starting the loop of %d iterations" % self.iterations)
        for i in range(self.iterations):
            data = {'Iteration': i}
            self.emit('results', data)
            log.debug("Emitting results: %s" % data)
            sleep(0.01)
            if self.should_stop():
                log.warning("Caught the stop flag in the procedure")
                break

if __name__ == "__main__":
    console_log(log)

    log.info("Constructing a SimpleProcedure")
    procedure = SimpleProcedure()
    procedure.iterations = 100

    data_filename = 'example.csv'
    log.info("Constructing the Results with a data file: %s" % data_filename)
    results = Results(procedure, data_filename)

    log.info("Constructing the Worker")
    worker = Worker(results)
    worker.start()
    log.info("Started the Worker")

    log.info("Joining with the worker in at most 1 hr")
    worker.join(timeout=3600) # wait at most 1 hr (3600 sec)
    log.info("Finished the measurement")
```

First, we have imported the Python logging module and grabbed the logger using the `__name__` argument. This gives us logging information specific to the current file. Conversely, we could use the `''` argument to get all logs, including those of pymeasure. We use the `console_log` function to conveniently output the log to the console. Further details on how to use the logger are addressed in the Python logging documentation.

Modifying our script

Now that you have a background on how to use the different features of the Procedure class, and how they are run, we will revisit our IV characteristic measurement using Procedures. Below we present the modified version of our example script, now as a IVProcedure class.

```
# Import necessary packages
from pymeasure.instruments.keithley import Keithley2400
from pymeasure.experiment import Procedure
from pymeasure.experiment import IntegerParameter, FloatParameter
from time import sleep

class IVProcedure(Procedure):

    data_points = IntegerParameter('Data points', default=50)
    averages = IntegerParameter('Averages', default=50)
    max_current = FloatParameter('Maximum Current', unit='A', default=0.01)
    min_current = FloatParameter('Minimum Current', unit='A', default=-0.01)

    DATA_COLUMNS = ['Current (A)', 'Voltage (V)', 'Voltage Std (V)']

    def startup(self):
        log.info("Connecting and configuring the instrument")
        self.sourcemeter = Keithley2400("GPIB::4")
        self.sourcemeter.reset()
        self.sourcemeter.use_front_terminals()
        self.sourcemeter.measure_voltage()
        self.sourcemeter.config_current_source()
        sleep(0.1) # wait here to give the instrument time to react
        self.sourcemeter.set_buffer(averages)

    def execute(self):
        currents = np.linspace(
            self.min_current,
            self.max_current,
            num=self.data_points
        )

        # Loop through each current point, measure and record the voltage
        for current in currents:
            log.info("Setting the current to %g A" % current)
            self.sourcemeter.current = current
            self.sourcemeter.reset_buffer()
            sleep(0.1)
            self.sourcemeter.start_buffer()
            log.info("Waiting for the buffer to fill with measurements")
            self.sourcemeter.wait_for_buffer()

            self.emit('results', {
                'Current (A)': current,
                'Voltage (V)': self.sourcemeter.means,
                'Voltage Std (V)': self.sourcemeter.standard_devs
            })
            sleep(0.01)
            if self.should_stop():
                log.info("User aborted the procedure")
                break
```

(continues on next page)

(continued from previous page)

```

def shutdown(self):
    self.sourcemeter.shutdown()
    log.info("Finished measuring")

if __name__ == "__main__":
    console_log(log)

    log.info("Constructing an IVProcedure")
    procedure = IVProcedure()
    procedure.data_points = 100
    procedure.averages = 50
    procedure.max_current = -0.01
    procedure.min_current = 0.01

    data_filename = 'example.csv'
    log.info("Constructing the Results with a data file: %s" % data_filename)
    results = Results(procedure, data_filename)

    log.info("Constructing the Worker")
    worker = Worker(results)
    worker.start()
    log.info("Started the Worker")

    log.info("Joining with the worker in at most 1 hr")
    worker.join(timeout=3600) # wait at most 1 hr (3600 sec)
    log.info("Finished the measurement")

```

At this point, you are familiar with how to construct a Procedure sub-class. The next section shows how to put these procedures to work in a graphical environment, where will have live-plotting of the data and the ability to easily queue up a number of experiments in sequence. All of these features come from using the Procedure object.

3.3 Using a graphical interface

In the previous tutorial we measured the IV characteristic of a sample to show how we can set up a simple experiment in PyMeasure. The real power of PyMeasure comes when we also use the graphical tools that are included to turn our simple example into a full-fledged user interface.

3.3.1 Using the Plotter

While it lacks the nice features of the ManagedWindow, the Plotter object is the simplest way of getting live-plotting. The Plotter takes a Results object and plots the data at a regular interval, grabbing the latest data each time from the file.

Let's extend our SimpleProcedure with a RandomProcedure, which generates random numbers during our loop. This example does not include instruments to provide a simpler example.

```

import logging
log = logging.getLogger(__name__)
log.addHandler(logging.NullHandler())

import random
from time import sleep
from pymeasure.log import console_log

```

(continues on next page)

```
from pymeasure.display import Plotter
from pymeasure.experiment import Procedure, Results, Worker
from pymeasure.experiment import IntegerParameter, FloatParameter, Parameter

class RandomProcedure(Procedure):

    iterations = IntegerParameter('Loop Iterations')
    delay = FloatParameter('Delay Time', units='s', default=0.2)
    seed = Parameter('Random Seed', default='12345')

    DATA_COLUMNS = ['Iteration', 'Random Number']

    def startup(self):
        log.info("Setting the seed of the random number generator")
        random.seed(self.seed)

    def execute(self):
        log.info("Starting the loop of %d iterations" % self.iterations)
        for i in range(self.iterations):
            data = {
                'Iteration': i,
                'Random Number': random.random()
            }
            self.emit('results', data)
            log.debug("Emitting results: %s" % data)
            sleep(self.delay)
            if self.should_stop():
                log.warning("Caught the stop flag in the procedure")
                break

if __name__ == "__main__":
    console_log(log)

    log.info("Constructing a RandomProcedure")
    procedure = RandomProcedure()
    procedure.iterations = 100

    data_filename = 'random.csv'
    log.info("Constructing the Results with a data file: %s" % data_filename)
    results = Results(procedure, data_filename)

    log.info("Constructing the Plotter")
    plotter = Plotter(results)
    plotter.start()
    log.info("Started the Plotter")


    log.info("Constructing the Worker")
    worker = Worker(results)
    worker.start()
    log.info("Started the Worker")

    log.info("Joining with the worker in at most 1 hr")
    worker.join(timeout=3600) # wait at most 1 hr (3600 sec)
    log.info("Finished the measurement")
```

The important addition is the construction of the Plotter from the Results object.

```
plotter = Plotter(results)
plotter.start()
```

The Plotter is started in a different process so that it can be run on a separate CPU for higher performance. The Plotter launches a Qt graphical interface using pyqtgraph which allows the Results data to be viewed based on the columns in the data.



3.3.2 Using the ManagedWindow

The ManagedWindow is the most convenient tool for running measurements with your Procedure. This has the major advantage of accepting the input parameters graphically. From the parameters, a graphical form is automatically generated that allows the inputs to be typed in. With this feature, measurements can be started dynamically, instead of defined in a script.

Another major feature of the ManagedWindow is its support for running measurements in a sequential queue. This allows you to set up a number of measurements with different input parameters, and watch them unfold on the live-plot. This is especially useful for long running measurements. The ManagedWindow achieves this through the Manager object, which coordinates which Procedure the Worker should run and keeps track of its status as the Worker progresses.

Below we adapt our previous example to use a ManagedWindow.

```
import logging
log = logging.getLogger(__name__)
log.addHandler(logging.NullHandler())

import sys
import tempfile
import random
from time import sleep
from pymeasure.log import console_log
from pymeasure.display.Qt import QtGui
from pymeasure.display.windows import ManagedWindow
from pymeasure.experiment import Procedure, Results
from pymeasure.experiment import IntegerParameter, FloatParameter, Parameter

class RandomProcedure(Procedure):

    iterations = IntegerParameter('Loop Iterations')
    delay = FloatParameter('Delay Time', units='s', default=0.2)
    seed = Parameter('Random Seed', default='12345')

    DATA_COLUMNS = ['Iteration', 'Random Number']

    def startup(self):
        log.info("Setting the seed of the random number generator")
        random.seed(self.seed)

    def execute(self):
        log.info("Starting the loop of %d iterations" % self.iterations)
        for i in range(self.iterations):
            data = {
                'Iteration': i,
                'Random Number': random.random()
            }
            self.emit('results', data)
            log.debug("Emitting results: %s" % data)
            sleep(self.delay)
            if self.should_stop():
                log.warning("Caught the stop flag in the procedure")
                break

class MainWindow(ManagedWindow):

    def __init__(self):
        super(MainWindow, self).__init__(
            procedure_class=RandomProcedure,
            inputs=['iterations', 'delay', 'seed'],
            displays=['iterations', 'delay', 'seed'],
            x_axis='Iteration',
            y_axis='Random Number'
        )
        self.setWindowTitle('GUI Example')

    def queue(self):
        filename = tempfile.mktemp()

        procedure = self.make_procedure()
```

(continues on next page)

(continued from previous page)

```


        results = Results(procedure, filename)
        experiment = self.new_experiment(results)

        self.manager.queue(experiment)

if __name__ == "__main__":
    app = QtGui.QApplication(sys.argv)
    window = MainWindow()
    window.show()
    sys.exit(app.exec_())

```


This results in the following graphical display.




In the code, the `MainWindow` class is a sub-class of the `ManagedWindow` class. We override the constructor to provide information about the procedure class and its options. The `inputs` are a list of `Parameters` class-variable names, which the display will generate graphical fields for. The `displays` is a similar list, which instead defines the parameters to display in the browser window. This browser keeps track of the experiments being run in the sequential queue.

The `queue` method establishes how the `Procedure` object is constructed. We use the `self.make_procedure` method to create a `Procedure` based on the graphical input fields. Here we are free to modify the procedure before


putting it on the queue. In this context, the Manager uses an Experiment object to keep track of the Procedure, Results, and its associated graphical representations in the browser and live-graph. This is then given to the Manager to queue the experiment.



By default the Manager starts a measurement when its procedure is queued. The abort button can be pressed to stop an experiment. In the Procedure, the `self.should_stop` call will catch the abort event and halt the measurement. It is important to check this value, or the Procedure will not be responsive to the abort event.



If you abort a measurement, the resume button must be pressed to continue the next measurement. This allows you to adjust anything, which is presumably why the abort was needed.



Now that you have learned about the `ManagedWindow`, you have all of the basics to get up and running quickly with a measurement and produce an easy to use graphical interface with PyMeasure.

3.3.3 Customising the plot options

For both the `PlotterWindow` and `ManagedWindow`, plotting is provided by the `pyqtgraph` library. This library allows you to change various plot options, as you might expect: axis ranges (by default auto-ranging), logarithmic and semilogarithmic axes, downsampling, grid display, FFT display, etc. There are two main ways you can do this:

1. You can right click on the plot to manually change any available options. This is also a good way of getting an overview of what options are available in `pyqtgraph`. Option changes will, of course, not persist across a restart of your program.
2. You can programmatically set these options using `pyqtgraph`'s `PlotItem` API, so that the window will open with these display options already set, as further explained below.

For `Plotter`, you can make a sub-class that overrides the `setup_plot()` method. This method will be called when the `Plotter` constructs the window. As an example

```
class LogPlotter(Plotter):
    def setup_plot(self, plot):
```

(continues on next page)

(continued from previous page)

```
# use logarithmic x-axis (e.g. for frequency sweeps)
plot.setLogMode(x=True)
```

For *ManagedWindow*, Similarly to the *Plotter*, the *setup_plot()* method can be overridden by your sub-class in order to do the set-up

```
class MainWindow(ManagedWindow):

    # ...

    def setup_plot(self, plot):
        # use logarithmic x-axis (e.g. for frequency sweeps)
        plot.setLogMode(x=True)

    # ...
```

It is also possible to access the *plot* attribute while outside of your sub-class, for example we could modify the previous section's example

```
if __name__ == "__main__":
    app = QtGui.QApplication(sys.argv)
    window = MainWindow()
    window.plot.setLogMode(x=True) # use logarithmic x-axis (e.g. for frequency_
↪sweeps)
    window.show()
    sys.exit(app.exec_())
```

See *pyqtgraph*'s API documentation on [PlotItem](#) for further details.

PYMEASURE.ADAPTERS

The adapter classes allow the instruments to be independent of the communication method used.

Adapters for specific instruments should be grouped in an `adapters.py` file in the corresponding manufacturer's folder of `pymeasure.instruments`. For example, the adapter for communicating with LakeShore instruments over USB, `LakeShoreUSBAdapter`, is found in `pymeasure.instruments.lakeshore.adapters`.

4.1 Adapter base class

class `pymeasure.adapters.Adapter`

Base class for Adapter child classes, which adapt between the Instrument object and the connection, to allow flexible use of different connection techniques.

This class should only be inherited from.

ask (*command*)

Writes the command to the instrument and returns the resulting ASCII response

Parameters **command** – SCPI command string to be sent to the instrument

Returns String ASCII response of the instrument

binary_values (*command, header_bytes=0, dtype=<class 'numpy.float32'>*)

Returns a numpy array from a query for binary data

Parameters

- **command** – SCPI command to be sent to the instrument
- **header_bytes** – Integer number of bytes to ignore in header
- **dtype** – The NumPy data type to format the values with

Returns NumPy array of values

read ()

Reads until the buffer is empty and returns the resulting ASCII response

Returns String ASCII response of the instrument.

values (*command, separator=',', cast=<class 'float'>*)

Writes a command to the instrument and returns a list of formatted values from the result

Parameters

- **command** – SCPI command to be sent to the instrument
- **separator** – A separator character to split the string into a list

- **cast** – A type to cast the result

Returns A list of the desired type, or strings where the casting fails

write (*command*)

Writes a command to the instrument

Parameters **command** – SCPI command string to be sent to the instrument

4.2 Fake adapter

class `pymeasure.adapters.FakeAdapter`

Bases: `pymeasure.adapters.adapter.Adapter`

Provides a fake adapter for debugging purposes, which bounces back the command so that arbitrary values testing is possible.

```
a = FakeAdapter()
assert a.read() == ""
a.write("5")
assert a.read() == "5"
assert a.read() == ""
assert a.ask("10") == "10"
assert a.values("10") == [10]
```

ask (*command*)

Writes the command to the instrument and returns the resulting ASCII response

Parameters **command** – SCPI command string to be sent to the instrument

Returns String ASCII response of the instrument

binary_values (*command, header_bytes=0, dtype=<class 'numpy.float32'>*)

Returns a numpy array from a query for binary data

Parameters

- **command** – SCPI command to be sent to the instrument
- **header_bytes** – Integer number of bytes to ignore in header
- **dtype** – The NumPy data type to format the values with

Returns NumPy array of values

read ()

Returns the last commands given after the last read call.

values (*command, separator=',', cast=<class 'float'>*)

Writes a command to the instrument and returns a list of formatted values from the result

Parameters

- **command** – SCPI command to be sent to the instrument
- **separator** – A separator character to split the string into a list
- **cast** – A type to cast the result

Returns A list of the desired type, or strings where the casting fails

write (*command*)

Writes the command to a buffer, so that it can be read back.

4.3 Serial adapter

class `pymeasure.adapters.SerialAdapter` (*port*, ***kwargs*)

Bases: `pymeasure.adapters.adapter.Adapter`

Adapter class for using the Python Serial package to allow serial communication to instrument

Parameters

- **port** – Serial port
- **kwargs** – Any valid key-word argument for `serial.Serial`

ask (*command*)

Writes the command to the instrument and returns the resulting ASCII response

Parameters **command** – SCPI command string to be sent to the instrument

Returns String ASCII response of the instrument

binary_values (*command*, *header_bytes=0*, *dtype=<class 'numpy.float32'>*)

Returns a numpy array from a query for binary data

Parameters

- **command** – SCPI command to be sent to the instrument
- **header_bytes** – Integer number of bytes to ignore in header
- **dtype** – The NumPy data type to format the values with

Returns NumPy array of values

read ()

Reads until the buffer is empty and returns the resulting ASCII response

Returns String ASCII response of the instrument.

values (*command*, *separator=' '*, *cast=<class 'float'>*)

Writes a command to the instrument and returns a list of formatted values from the result

Parameters

- **command** – SCPI command to be sent to the instrument
- **separator** – A separator character to split the string into a list
- **cast** – A type to cast the result

Returns A list of the desired type, or strings where the casting fails

write (*command*)

Writes a command to the instrument

Parameters **command** – SCPI command string to be sent to the instrument

4.4 Prologix adapter

class `pymeasure.adapters.PrologixAdapter` (*port*, *address=None*, *rw_delay=None*, *serial_timeout=0.5*, ***kwargs*)

Bases: `pymeasure.adapters.serial.SerialAdapter`

Encapsulates the additional commands necessary to communicate over a Prologix GPIB-USB Adapter, using the `SerialAdapter`.

Each `PrologixAdapter` is constructed based on a serial port or connection and the GPIB address to be communicated to. Serial connection sharing is achieved by using the `gpiib()` method to spawn new `PrologixAdapters` for different GPIB addresses.

Parameters

- **port** – The Serial port name or a `serial.Serial` object
- **address** – Integer GPIB address of the desired instrument
- **rw_delay** – An optional delay to set between a write and read call for slow to respond instruments.
- **kwargs** – Key-word arguments if constructing a new serial object

Variables **address** – Integer GPIB address of the desired instrument

To allow user access to the Prologix adapter in Linux, create the file: `/etc/udev/rules.d/51-prologix.rules`, with contents:

```
SUBSYSTEMS=="usb",ATTRS{idVendor}=="0403",ATTRS{idProduct}=="6001",MODE="0666"
```

Then reload the udev rules with:

```
sudo udevadm control --reload-rules
sudo udevadm trigger
```

ask (*command*)

Ask the Prologix controller, include a forced delay for some instruments.

Parameters **command** – SCPI command string to be sent to instrument

binary_values (*command*, *header_bytes=0*, *dtype=<class 'numpy.float32'>*)

Returns a numpy array from a query for binary data

Parameters

- **command** – SCPI command to be sent to the instrument
- **header_bytes** – Integer number of bytes to ignore in header
- **dtype** – The NumPy data type to format the values with

Returns NumPy array of values

gpiib (*address*, *rw_delay=None*)

Returns and `PrologixAdapter` object that references the GPIB address specified, while sharing the Serial connection with other calls of this function

Parameters

- **address** – Integer GPIB address of the desired instrument
- **rw_delay** – Set a custom Read/Write delay for the instrument

Returns `PrologixAdapter` for specific GPIB address

read()

Reads the response of the instrument until timeout

Returns String ASCII response of the instrument

set_defaults()

Sets up the default behavior of the Prologix-GPIB adapter

values (*command*, *separator*=' ', *cast*=<class 'float'>)

Writes a command to the instrument and returns a list of formatted values from the result

Parameters

- **command** – SCPI command to be sent to the instrument
- **separator** – A separator character to split the string into a list
- **cast** – A type to cast the result

Returns A list of the desired type, or strings where the casting fails

wait_for_srq (*timeout*=25, *delay*=0.1)

Blocks until a SRQ, and leaves the bit high

Parameters

- **timeout** – Timeout duration in seconds
- **delay** – Time delay between checking SRQ in seconds

write (*command*)

Writes the command to the GPIB address stored in the *address*

Parameters **command** – SCPI command string to be sent to the instrument

4.5 VISA adapter

class `pymeaure.adapters.VISAAdapter` (*resourceName*, *visa_library*="", ***kwargs*)

Bases: `pymeaure.adapters.adapter.Adapter`

Adapter class for the VISA library using PyVISA to communicate with instruments.

Parameters

- **resource** – VISA resource name that identifies the address
- **visa_library** – VisaLibrary Instance, path of the VISA library or VisaLibrary spec string (@py or @ni). if not given, the default for the platform will be used.
- **kwargs** – Any valid key-word arguments for constructing a PyVISA instrument

ask (*command*)

Writes the command to the instrument and returns the resulting ASCII response

Parameters **command** – SCPI command string to be sent to the instrument

Returns String ASCII response of the instrument

ask_values (*command*)

Writes a command to the instrument and returns a list of formatted values from the result. The format of the return is configured by `self.config()`.

Parameters **command** – SCPI command to be sent to the instrument

Returns Formatted response of the instrument.

binary_values (*command*, *header_bytes=0*, *dtype=<class 'numpy.float32'>*)

Returns a numpy array from a query for binary data

Parameters

- **command** – SCPI command to be sent to the instrument
- **header_bytes** – Integer number of bytes to ignore in header
- **dtype** – The NumPy data type to format the values with

Returns NumPy array of values

config (*is_binary=False*, *datatype='str'*, *container=<built-in function array>*, *converter='s'*, *separator=', '*, *is_big_endian=False*)

Configure the format of data transfer to and from the instrument.

Parameters

- **is_binary** – If True, data is in binary format, otherwise ASCII.
- **datatype** – Data type.
- **container** – Return format. Any callable/type that takes an iterable.
- **converter** – String converter, used in dealing with ASCII data.
- **separator** – Delimiter of a series of data in ASCII.
- **is_big_endian** – Endianness.

static has_supported_version ()

Returns True if the PyVISA version is greater than 1.8

read ()

Reads until the buffer is empty and returns the resulting ASCII response

Returns String ASCII response of the instrument.

read_bytes (*size*)

Reads specified number of bytes from the buffer and returns the resulting ASCII response

Parameters **size** – Number of bytes to read from the buffer

Returns String ASCII response of the instrument.

values (*command*, *separator=', '*, *cast=<class 'float'>*)

Writes a command to the instrument and returns a list of formatted values from the result

Parameters

- **command** – SCPI command to be sent to the instrument
- **separator** – A separator character to split the string into a list
- **cast** – A type to cast the result

Returns A list of the desired type, or strings where the casting fails

wait_for_srq (*timeout=25*, *delay=0.1*)

Blocks until a SRQ, and leaves the bit high

Parameters

- **timeout** – Timeout duration in seconds
- **delay** – Time delay between checking SRQ in seconds

write (*command*)

Writes a command to the instrument

Parameters **command** – SCPI command string to be sent to the instrument

4.6 VXI-11 adapter

PYMEASURE.EXPERIMENT

This section contains specific documentation on the classes and methods of the package.

5.1 Experiment class

The Experiment class is intended for use in the Jupyter notebook environment.

class `pymeasure.experiment.experiment.Experiment` (*title*, *procedure*, *analyse*=<function *Experiment.<lambda>>*)

Bases: `object`

Class which starts logging and creates/runs the results and worker processes.

```
procedure = Procedure()
experiment = Experiment(title, procedure)
experiment.start()
experiment.plot_live('x', 'y', style='.-')
```

for a multi-subplot graph:

```
import pylab as pl
ax1 = pl.subplot(121)
experiment.plot('x', 'y', ax=ax1)
ax2 = pl.subplot(122)
experiment.plot('x', 'z', ax=ax2)
experiment.plot_live()
```

Variables `value` – The value of the parameter

Parameters

- **title** – The experiment title
- **procedure** – The procedure object
- **analyse** – Post-analysis function, which takes a pandas dataframe as input and returns it with added (analysed) columns. The analysed results are accessible via `experiment.data`, as opposed to `experiment.results.data` for the ‘raw’ data.
- **_data_timeout** – Time limit for how long live plotting should wait for datapoints.

clear_plot ()

Clear the figures and plot lists.

property data

Data property which returns analysed data, if an analyse function is defined, otherwise returns the raw data.

pcolor (*xname, yname, zname, *args, **kwargs*)

Plot the results from the experiment.data pandas dataframe in a pcolor graph. Store the plots in a plots list attribute.

plot (**args, **kwargs*)

Plot the results from the experiment.data pandas dataframe. Store the plots in a plots list attribute.

plot_live (**args, **kwargs*)

Live plotting loop for jupyter notebook, which automatically updates (an) in-line matplotlib graph(s). Will create a new plot as specified by input arguments, or will update (an) existing plot(s).

start ()

Start the worker

update_line (*ax, hl, xname, yname*)

Update a line in a matplotlib graph with new data.

update_pcolor (*ax, xname, yname, zname*)

Update a pcolor graph with new data.

update_plot ()

Update the plots in the plots list with new data from the experiment.data pandas dataframe.

wait_for_data ()

Wait for the data attribute to fill with datapoints.

pymeaure.experiment.experiment.create_filename (*title*)

Create a new filename according to the style defined in the config file. If no config is specified, create a temporary file.

pymeaure.experiment.experiment.get_array (*start, stop, step*)

Returns a numpy array from start to stop

pymeaure.experiment.experiment.get_array_steps (*start, stop, numsteps*)

Returns a numpy array from start to stop in numsteps

pymeaure.experiment.experiment.get_array_zero (*maxval, step*)

Returns a numpy array from 0 to maxval to -maxval to 0

5.2 Listener class

class **pymeaure.experiment.listeners.Listener** (*port, topic="", timeout=0.01*)

Bases: **pymeaure.thread.StoppableThread**

Base class for Threads that need to listen for messages on a ZMQ TCP port and can be stopped by a thread-safe method call

message_waiting ()**receive** (*flags=0*)**class** **pymeaure.experiment.listeners.Monitor** (*results, queue*)

Bases: **pymeaure.log.QueueListener**

class **pymeaure.experiment.listeners.Recorder** (*results, queue, **kwargs*)

Bases: **pymeaure.log.QueueListener**

Recorder loads the initial Results for a filepath and appends data by listening for it over a queue. The queue ensures that no data is lost between the Recorder and Worker.

5.3 Procedure class

class `pymeaure.experiment.procedure.Procedure` (**kwargs)

Provides the base class of a procedure to organize the experiment execution. Procedures should be run by Workers to ensure that asynchronous execution is properly managed.

```
procedure = Procedure()
results = Results(procedure, data_filename)
worker = Worker(results, port)
worker.start()
```

Inheriting classes should define the startup, execute, and shutdown methods as needed. The shutdown method is called even with a software exception or abort event during the execute method.

If keyword arguments are provided, they are added to the object as attributes.

check_parameters ()

Raises an exception if any parameter is missing before calling the associated function. Ensures that each value can be set and got, which should cast it into the right format. Used as a decorator `@check_parameters` on the startup method

execute ()

Performs the commands needed for the measurement itself. During execution the shutdown method will always be run following this method. This includes when Exceptions are raised.

gen_measurement ()

Create MEASURE and DATA_COLUMNS variables for get_datapoint method.

parameter_objects ()

Returns a dictionary of all the Parameter objects and grabs any current values that are not in the default definitions

parameter_values ()

Returns a dictionary of all the Parameter values and grabs any current values that are not in the default definitions

parameters_are_set ()

Returns True if all parameters are set

refresh_parameters ()

Enforces that all the parameters are re-cast and updated in the meta dictionary

set_parameters (parameters, except_missing=True)

Sets a dictionary of parameters and raises an exception if additional parameters are present if `except_missing` is True

shutdown ()

Executes the commands necessary to shut down the instruments and leave them in a safe state. This method is always run at the end.

startup ()

Executes the commands needed at the start-up of the measurement

class `pymeaure.experiment.procedure.UnknownProcedure` (parameters)

Handles the case when a `Procedure` object can not be imported during loading in the `Results` class

startup()

Executes the commands needed at the start-up of the measurement

5.4 Parameter classes

The parameter classes are used to define input variables for a *Procedure*. They each inherit from the *Parameter* base class.

```
class pymeasure.experiment.parameters.BooleanParameter (name, default=None,
                                                         ui_class=None)
```

Parameter sub-class that uses the boolean type to store the value.

Variables **value** – The boolean value of the parameter

Parameters

- **name** – The parameter name
- **default** – The default boolean value
- **ui_class** – A Qt class to use for the UI of this parameter

```
class pymeasure.experiment.parameters.FloatParameter (name, units=None, minimum=-
                                                         1000000000.0, maximum=
                                                         1000000000.0,
                                                         **kwargs)
```

Parameter sub-class that uses the floating point type to store the value.

Variables **value** – The floating point value of the parameter

Parameters

- **name** – The parameter name
- **units** – The units of measure for the parameter
- **minimum** – The minimum allowed value (default: -1e9)
- **maximum** – The maximum allowed value (default: 1e9)
- **default** – The default floating point value
- **ui_class** – A Qt class to use for the UI of this parameter

```
class pymeasure.experiment.parameters.IntegerParameter (name, units=None,
                                                         minimum=-1000000000.0,
                                                         maximum=1000000000.0,
                                                         **kwargs)
```

Parameter sub-class that uses the integer type to store the value.

Variables **value** – The integer value of the parameter

Parameters

- **name** – The parameter name
- **units** – The units of measure for the parameter
- **minimum** – The minimum allowed value (default: -1e9)
- **maximum** – The maximum allowed value (default: 1e9)
- **default** – The default integer value
- **ui_class** – A Qt class to use for the UI of this parameter

class `pymeaure.experiment.parameters.ListParameter` (*name*, *choices=None*,
units=None, ***kwargs*)

Parameter sub-class that stores the value as a list.

Parameters

- **name** – The parameter name
- **choices** – An explicit list of choices, which is disregarded if None
- **units** – The units of measure for the parameter
- **default** – The default value
- **ui_class** – A Qt class to use for the UI of this parameter

property choices

Returns an immutable iterable of choices, or None if not set.

class `pymeaure.experiment.parameters.Measurable` (*name*, *fget=None*, *units=None*,
measure=True, *default=None*,
***kwargs*)

Encapsulates the information for a measurable experiment parameter with information about the name, fget function and units if supplied. The value property is called when the procedure retrieves a datapoint and calls the fget function. If no fget function is specified, the value property will return the latest set value of the parameter (or default if never set).

Variables **value** – The value of the parameter

Parameters

- **name** – The parameter name
- **fget** – The parameter fget function (e.g. an instrument parameter)
- **default** – The default value

class `pymeaure.experiment.parameters.Parameter` (*name*, *default=None*, *ui_class=None*)

Encapsulates the information for an experiment parameter with information about the name, and units if supplied.

Variables **value** – The value of the parameter

Parameters

- **name** – The parameter name
- **default** – The default value
- **ui_class** – A Qt class to use for the UI of this parameter

is_set ()

Returns True if the Parameter value is set

class `pymeaure.experiment.parameters.PhysicalParameter` (*name*, *uncertaintyType='absolute'*,
***kwargs*)

VectorParameter sub-class of 2 dimensions to store a value and its uncertainty.

Variables **value** – The value of the parameter as a list of 2 floating point numbers

Parameters

- **name** – The parameter name
- **uncertainty_type** – Type of uncertainty, 'absolute', 'relative' or 'percentage'

- **units** – The units of measure for the parameter
- **default** – The default value
- **ui_class** – A Qt class to use for the UI of this parameter

class `pymeasure.experiment.parameters.VectorParameter` (*name, length=3, units=None, **kwargs*)

Parameter sub-class that stores the value in a vector format.

Variables **value** – The value of the parameter as a list of floating point numbers

Parameters

- **name** – The parameter name
- **length** – The integer dimensions of the vector
- **units** – The units of measure for the parameter
- **default** – The default value
- **ui_class** – A Qt class to use for the UI of this parameter

5.5 Worker class

class `pymeasure.experiment.workers.Worker` (*results, log_queue=None, log_level=20, port=None*)

Bases: `pymeasure.thread.StoppableThread`

Worker runs the procedure and emits information about the procedure and its status over a ZMQ TCP port. In a child thread, a Recorder is run to write the results to

emit (*topic, record*)

Emits data of some topic over TCP

handle_abort ()

handle_error ()

join (*timeout=0*)

Joins the current thread and forces it to stop after the timeout if necessary

Parameters **timeout** – Timeout duration in seconds

run ()

Method representing the thread's activity.

You may override this method in a subclass. The standard `run()` method invokes the callable object passed to the object's constructor as the target argument, if any, with sequential and keyword arguments taken from the `args` and `kwargs` arguments, respectively.

shutdown ()

update_status (*status*)

5.6 Results class

class `pymeaure.experiment.results.CSVFormatter` (*columns, delimiter=' , '*)
 Formatter of data results

format (*record*)

Formats a record as csv.

Parameters **record** (*dict*) – record to format.

Returns a string

class `pymeaure.experiment.results.Results` (*procedure, data_filename*)

The Results class provides a convenient interface to reading and writing data in connection with a *Procedure* object.

Variables

- **COMMENT** – The character used to identify a comment (default: #)
- **DELIMITER** – The character used to delimit the data (default: ,)
- **LINE_BREAK** – The character used for line breaks (default n)
- **CHUNK_SIZE** – The length of the data chunk that is read

Parameters

- **procedure** – Procedure object
- **data_filename** – The data filename where the data is or should be stored

format (*data*)

Returns a formatted string containing the data to be written to a file

header ()

Returns a text header to accompany a datafile so that the procedure can be reconstructed

labels ()

Returns the columns labels as a string to be written to the file

static load (*data_filename, procedure_class=None*)

Returns a Results object with the associated Procedure object and data

parse (*line*)

Returns a dictionary containing the data from the line

static parse_header (*header, procedure_class=None*)

Returns a Procedure object with the parameters as defined in the header text.

reload ()

Performs a full reloading of the file data, neglecting any changes in the comments

`pymeaure.experiment.results.unique_filename` (*directory, prefix='DATA', suffix='', ext='csv', dated_folder=False, index=True, datetimeformat='%Y-%m-%d'*)

Returns a unique filename based on the directory and prefix

PYMEASURE.DISPLAY

This section contains specific documentation on the classes and methods of the package.

6.1 Browser classes

class `pymeasure.display.browser.Browser` (*procedure_class*, *display_parameters*, *measured_quantities*, *sort_by_filename=False*, *parent=None*)

Bases: `PyQt5.QtWidgets.QTreeWidget`

Graphical list view of *Experiment* objects allowing the user to view the status of queued Experiments as well as loading and displaying data from previous runs.

In order that different Experiments be displayed within the same Browser, they must have entries in *DATA_COLUMNS* corresponding to the *measured_quantities* of the Browser.

add (*experiment*)

Add a *Experiment* object to the Browser. This function checks to make sure that the Experiment measures the appropriate quantities to warrant its inclusion, and then adds a *BrowserItem* to the Browser, filling all relevant columns with Parameter data.

class `pymeasure.display.browser.BrowserItem` (*results*, *curve*, *parent=None*)

Bases: `PyQt5.QtWidgets.QTreeWidgetItem`

6.2 Curves classes

class `pymeasure.display.curves.BufferCurve` (*errors=False*, ***kwargs*)

Bases: `pyqtgraph.graphicsItems.PlotDataItem.PlotDataItem`

Creates a curve based on a predefined buffer size and allows data to be added dynamically, in addition to supporting error bars

append (*x*, *y*, *xError=None*, *yError=None*)

Appends data to the curve with optional errors

prepare (*size*, *dtype=<class 'numpy.float32'>*)

Prepares the buffer based on its size, data type

class `pymeasure.display.curves.Crosshairs` (*plot*, *pen=None*)

Bases: `PyQt5.QtCore.QObject`

Attaches crosshairs to the a plot and provides a signal with the x and y graph coordinates

mouseMoved (*event=None*)

Updates the mouse position upon mouse movement

update ()

Updates the mouse position based on the data in the plot. For dynamic plots, this is called each time the data changes to ensure the x and y values correspond to those on the display.

class `pymeasure.display.curves.ResultsCurve` (*results, x, y, xerr=None, yerr=None, force_reload=False, **kwargs*)

Bases: `pyqtgraph.graphicsItems.PlotDataItem.PlotDataItem`

Creates a curve loaded dynamically from a file through the Results object and supports error bars. The data can be forced to fully reload on each update, useful for cases when the data is changing across the full file instead of just appending.

update ()

Updates the data by polling the results

6.3 Inputs classes

class `pymeasure.display.inputs.BooleanInput` (*parameter, parent=None, **kwargs*)

Bases: `PyQt5.QtWidgets.QCheckBox`, `pymeasure.display.inputs.Input`

Checkbox for boolean values, connected to a `BooleanParameter`.

set_parameter (*parameter*)

Connects a new parameter to the input box, and initializes the box value.

Parameters *parameter* – parameter to connect.

class `pymeasure.display.inputs.FloatInput` (*parameter, parent=None, **kwargs*)

Bases: `PyQt5.QtWidgets.QDoubleSpinBox`, `pymeasure.display.inputs.Input`

Spin input box for floating-point values, connected to a `FloatParameter`.

See also:

Class `ScientificInput` For inputs in scientific notation.

set_parameter (*parameter*)

Connects a new parameter to the input box, and initializes the box value.

Parameters *parameter* – parameter to connect.

class `pymeasure.display.inputs.Input` (*parameter, **kwargs*)

Bases: `object`

Mix-in class that connects a `Parameter` object to a GUI input box.

Parameters *parameter* – The parameter to connect to this input box.

Attr *parameter* Read-only property to access the associated parameter.

property *parameter*

The connected parameter object. Read-only property; see `set_parameter()`.

Note that reading this property will have the side-effect of updating its value from the GUI input box.

set_parameter (*parameter*)

Connects a new parameter to the input box, and initializes the box value.

Parameters *parameter* – parameter to connect.

update_parameter ()

Update the parameter value with the Input GUI element's current value.

class `pymeasure.display.inputs.IntegerInput` (*parameter*, *parent=None*, ***kwargs*)

Bases: `PyQt5.QtWidgets.QSpinBox`, `pymeasure.display.inputs.Input`

Spin input box for integer values, connected to a `IntegerParameter`.

set_parameter (*parameter*)

Connects a new parameter to the input box, and initializes the box value.

Parameters *parameter* – parameter to connect.

class `pymeasure.display.inputs.ListInput` (*parameter*, *parent=None*, ***kwargs*)

Bases: `PyQt5.QtWidgets.QComboBox`, `pymeasure.display.inputs.Input`

Dropdown for list values, connected to a `ListParameter`.

set_parameter (*parameter*)

Connects a new parameter to the input box, and initializes the box value.

Parameters *parameter* – parameter to connect.

class `pymeasure.display.inputs.ScientificInput` (*parameter*, *parent=None*, ***kwargs*)

Bases: `PyQt5.QtWidgets.QDoubleSpinBox`, `pymeasure.display.inputs.Input`

Spinner input box for floating-point values, connected to a `FloatParameter`. This box will display and accept values in scientific notation when appropriate.

See also:

Class `FloatInput` For a non-scientific floating-point input box.

set_parameter (*parameter*)

Connects a new parameter to the input box, and initializes the box value.

Parameters *parameter* – parameter to connect.

stepEnabled (*self*) → `QAbstractSpinBox.StepEnabled`

textFromValue (*self*, *float*) → `str`

validate (*self*, *str*, *int*) → `Tuple[QValidator.State, str, int]`

valueFromText (*self*, *str*) → `float`

class `pymeasure.display.inputs.StringInput` (*parameter*, *parent=None*, ***kwargs*)

Bases: `PyQt5.QtWidgets.QLineEdit`, `pymeasure.display.inputs.Input`

String input box connected to a `Parameter`. Parameter subclasses that are string-based may also use this input, but non-string parameters should use more specialised input classes.

6.4 Listeners classes

class `pymeasure.display.listeners.Monitor` (*queue*)

Bases: `PyQt5.QtCore.QThread`

Monitor listens for status and progress messages from a `Worker` through a queue to ensure no messages are lost

run (*self*)

class `pymeasure.display.listeners.QListener` (*port*, *topic=""*, *timeout=0.01*)
Bases: `pymeasure.display.thread.StoppableQThread`

Base class for QThreads that need to listen for messages on a ZMQ TCP port and can be stopped by a thread- and process-safe method call

6.5 Log classes

class `pymeasure.display.log.LogHandler` (*parent=None*)
Bases: `PyQt5.QtCore.QObject`, `logging.Handler`

emit (*record*)

Do whatever it takes to actually log the specified logging record.

This version is intended to be implemented by subclasses and so raises a `NotImplementedError`.

6.6 Manager classes

class `pymeasure.display.manager.Experiment` (*results*, *curve*, *browser_item*, *parent=None*)
Bases: `PyQt5.QtCore.QObject`

The `Experiment` class helps group the *Procedure*, *Results*, and their display functionality. Its function is only a convenient container.

Parameters

- **results** – *Results* object
- **curve** – *ResultsCurve* object
- **browser_item** – *BrowserItem* object

class `pymeasure.display.manager.ExperimentQueue`
Bases: `PyQt5.QtCore.QObject`

Represents a Queue of Experiments and allows queries to be easily preformed

has_next ()

Returns True if another item is on the queue

next ()

Returns the next experiment on the queue

class `pymeasure.display.manager.Manager` (*plot*, *browser*, *port=5888*, *log_level=20*, *parent=None*)
Bases: `PyQt5.QtCore.QObject`

Controls the execution of *Experiment* classes by implementing a queue system in which Experiments are added, removed, executed, or aborted. When instantiated, the Manager is linked to a *Browser* and a PyQt-Graph *PlotItem* within the user interface, which are updated in accordance with the execution status of the Experiments.

abort ()

Aborts the currently running Experiment, but raises an exception if there is no running experiment

clear ()

Remove all Experiments

is_running()
Returns True if a procedure is currently running

load(*experiment*)
Load a previously executed Experiment

next()
Initiates the start of the next experiment in the queue as long as no other experiments are currently running and there is a procedure in the queue.

queue(*experiment*)
Adds an experiment to the queue.

remove(*experiment*)
Removes an Experiment

resume()
Resume processing of the queue.

6.7 Plotter class

class `pymeasure.display.plotter.Plotter` (*results, refresh_time=0.1*)

Bases: `pymeasure.thread.StoppableThread`

Plotter dynamically plots data from a file through the Results object and supports error bars.

See also:

Tutorial *Using the Plotter* A tutorial and example on using the Plotter and PlotterWindow.

run()
Method representing the thread's activity.

You may override this method in a subclass. The standard run() method invokes the callable object passed to the object's constructor as the target argument, if any, with sequential and keyword arguments taken from the args and kwargs arguments, respectively.

setup_plot(*plot*)
This method does nothing by default, but can be overridden by the child class in order to set up custom options for the plot window, via its `PlotItem`.

Parameters `plot` – This window's `PlotItem` instance.

6.8 Qt classes

All Qt imports should reference `pymeasure.display.Qt`, for consistent importing from either PySide or PyQt4.

`Qt.fromUi` (***kwargs*)

Returns a Qt object constructed using `loadUiType` based on its arguments. All `QWidget` objects in the form class are set in the returned object for easy accessibility.

6.9 Thread classes

class `pymeasure.display.thread.StoppableQThread` (*parent=None*)

Bases: `PyQt5.QtCore.QThread`

Base class for QThreads which require the ability to be stopped by a thread-safe method call

join (*timeout=0*)

Joins the current thread and forces it to stop after the timeout if necessary

Parameters `timeout` – Timeout duration in seconds

6.10 Widget classes

class `pymeasure.display.widgets.BrowserWidget` (**args, parent=None*)

Bases: `PyQt5.QtWidgets.QWidget`

class `pymeasure.display.widgets.InputsWidget` (*procedure_class, inputs=(), parent=None*)

Bases: `PyQt5.QtWidgets.QWidget`

get_procedure ()

Returns the current procedure

class `pymeasure.display.widgets.LogWidget` (*parent=None*)

Bases: `PyQt5.QtWidgets.QWidget`

class `pymeasure.display.widgets.PlotFrame` (*x_axis=None, y_axis=None, refresh_time=0.2, check_status=True, parent=None*)

Bases: `PyQt5.QtWidgets.QFrame`

Combines a PyQtGraph Plot with Crosshairs. Refreshes the plot based on the `refresh_time`, and allows the axes to be changed on the fly, which updates the plotted data

parse_axis (*axis*)

Returns the units of an axis by searching the string

class `pymeasure.display.widgets.PlotWidget` (*columns, x_axis=None, y_axis=None, refresh_time=0.2, check_status=True, parent=None*)

Bases: `PyQt5.QtWidgets.QWidget`

Extends the `PlotFrame` to allow different columns of the data to be dynamically chosen

sizeHint (*self*) → `QSize`

class `pymeasure.display.widgets.ResultsDialog` (*columns, x_axis=None, y_axis=None, parent=None*)

Bases: `PyQt5.QtWidgets.QFileDialog`

6.11 Windows classes

```
class pymeasure.display.windows.ManagedWindow (procedure_class, inputs=(), displays=(), x_axis=None, y_axis=None, log_channel="", log_level=20, parent=None)
```

Bases: PyQt5.QtWidgets.QMainWindow

Abstract base class.

The `ManagedWindow` provides an interface for inputting experiment parameters, running several experiments (*Procedure*), plotting result curves, and listing the experiments conducted during a session.

The `ManagedWindow` uses a `Manager` to control `Workers` in a `Queue`, and provides a simple interface. The `queue()` method must be overridden by the child class.

See also:

Tutorial *Using the ManagedWindow* A tutorial and example on the basic configuration and usage of `ManagedWindow`.

plot

The `pyqtgraph.PlotItem` object for this window. Can be accessed to further customise the plot view programmatically, e.g., display log-log or semi-log axes by default, change axis range, etc.

queue()

Abstract method, which must be overridden by the child class.

Implementations must call `self.manager.queue(experiment)` and pass an experiment (*Experiment*) object which contains the *Results* and *Procedure* to be run.

For example:

```
def queue(self):
    filename = unique_filename('results', prefix="data") # from pymeasure.
    ↪experiment

    procedure = self.make_procedure() # Procedure class was passed at_
    ↪construction
    results = Results(procedure, filename)
    experiment = self.new_experiment(results)

    self.manager.queue(experiment)
```

set_parameters (*parameters*)

This method should be overwritten by the child class. The `parameters` argument is a dictionary of `Parameter` objects. The `Parameters` should overwrite the GUI values so that a user can click “Queue” to capture the same parameters.

setup_plot (*plot*)

This method does nothing by default, but can be overridden by the child class in order to set up custom options for the plot

This method is called during the constructor, after all other set up has been completed, and is provided as a convenience method to parallel `Plotter`.

Parameters `plot` – This window’s `PlotItem` instance.

```
class pymeasure.display.windows.PlotterWindow (plotter, refresh_time=0.1, parent=None)
```

Bases: PyQt5.QtWidgets.QMainWindow

A window for plotting experiment results. Should not be instantiated directly, but only via the *Plotter* class.

See also:

Tutorial *Using the Plotter* A tutorial and example code for using the Plotter and PlotterWindow.

check_stop ()

Checks if the Plotter should stop and exits the Qt main loop if so

PYMEASURE.INSTRUMENTS

This section contains documentation on the instrument classes.

7.1 Instrument classes

class `pymeasure.instruments.Instrument` (*adapter, name, includeSCPI=True, **kwargs*)

This provides the base class for all Instruments, which is independent of the particular Adapter used to connect for communication to the instrument. It provides basic SCPI commands by default, but can be toggled with `includeSCPI`.

Parameters

- **adapter** – An *Adapter* object
- **name** – A string name
- **includeSCPI** – A boolean, which toggles the inclusion of standard SCPI commands

ask (*command*)

Writes the command to the instrument through the adapter and returns the read response.

Parameters **command** – command string to be sent to the instrument

check_errors ()

Return any accumulated errors. Must be reimplemented by subclasses.

clear ()

Clears the instrument status byte

static control (*get_command, set_command, docs, validator=<function Instrument.<lambda>>, values=(), map_values=False, get_process=<function Instrument.<lambda>>, set_process=<function Instrument.<lambda>>, check_set_errors=False, check_get_errors=False, **kwargs*)

Returns a property for the class based on the supplied commands. This property may be set and read from the instrument.

Parameters

- **get_command** – A string command that asks for the value
- **set_command** – A string command that writes the value
- **docs** – A docstring that will be included in the documentation
- **validator** – A function that takes both a value and a group of valid values and returns a valid value, while it otherwise raises an exception

- **values** – A list, tuple, range, or dictionary of valid values, that can be used as to map values if `map_values` is `True`.
- **map_values** – A boolean flag that determines if the values should be interpreted as a map
- **get_process** – A function that take a value and allows processing before value mapping, returning the processed value
- **set_process** – A function that takes a value and allows processing before value mapping, returning the processed value
- **check_set_errors** – Toggles checking errors after setting
- **check_get_errors** – Toggles checking errors after getting

property id

Requests and returns the identification of the instrument.

static measurement (*get_command*, *docs*, *values=()*, *map_values=None*, *get_process=<function Instrument.<lambda>>*, *command_process=<function Instrument.<lambda>>*, *check_get_errors=False*, ***kwargs*)

Returns a property for the class based on the supplied commands. This is a measurement quantity that may only be read from the instrument, not set.

Parameters

- **get_command** – A string command that asks for the value
- **docs** – A docstring that will be included in the documentation
- **values** – A list, tuple, range, or dictionary of valid values, that can be used as to map values if `map_values` is `True`.
- **map_values** – A boolean flag that determines if the values should be interpreted as a map
- **get_process** – A function that take a value and allows processing before value mapping, returning the processed value
- **command_process** – A function that take a command and allows processing before executing the command, for both getting and setting
- **check_get_errors** – Toggles checking errors after getting

read()

Reads from the instrument through the adapter and returns the response.

reset()

Resets the instrument.

static setting (*set_command*, *docs*, *validator=<function Instrument.<lambda>>*, *values=()*, *map_values=False*, *set_process=<function Instrument.<lambda>>*, *check_set_errors=False*, ***kwargs*)

Returns a property for the class based on the supplied commands. This property may be set, but raises an exception when being read from the instrument.

Parameters

- **set_command** – A string command that writes the value
- **docs** – A docstring that will be included in the documentation
- **validator** – A function that takes both a value and a group of valid values and returns a valid value, while it otherwise raises an exception

- **values** – A list, tuple, range, or dictionary of valid values, that can be used as to map values if `map_values` is `True`.
- **map_values** – A boolean flag that determines if the values should be interpreted as a map
- **set_process** – A function that takes a value and allows processing before value mapping, returning the processed value
- **check_set_errors** – Toggles checking errors after setting

shutdown ()

Brings the instrument to a safe and stable state

values (*command*, ***kwargs*)

Reads a set of values from the instrument through the adapter, passing on any key-word arguments.

write (*command*)

Writes the command to the instrument through the adapter.

Parameters **command** – command string to be sent to the instrument

class `pymeasure.instruments.Mock` (*wait=0.1*, ***kwargs*)

Bases: `pymeasure.instruments.instrument.Instrument`

Mock instrument for testing.

get_time ()

Get elapsed time

get_voltage ()

Get the voltage.

get_wave ()

Get wave.

reset_time ()

Reset the timer to 0 s.

set_output_voltage (*value*)

Set the voltage.

set_time (*value*)

Wait for the timer to reach the specified time. If `value = 0`, reset.

property **time**

Get elapsed time

property **voltage**

Get the voltage.

property **wave**

Get wave.

7.2 Validator functions

Validators are used in conjunction with the `Instrument.control` function to allow properties with complex restrictions for valid values. They are described in more detail in the *Advanced properties* section.

`pymeasure.instruments.validators.discreteTruncate` (*number*, *discreteSet*)

Truncates the number to the closest element in the positive discrete set. Returns False if the number is larger than the maximum value or negative.

`pymeasure.instruments.validators.joined_validators` (**validators*)

Join a list of validators together as a single. Expects a list of validator functions and values.

Parameters *validators* – an iterable of other validators

`pymeasure.instruments.validators.modular_range` (*value*, *values*)

Provides a validator function that returns the value if it is in the range. Otherwise it returns the value, modulo the max of the range.

Parameters

- **value** – a value to test
- **values** – A set of values that are valid

`pymeasure.instruments.validators.modular_range_bidirectional` (*value*, *values*)

Provides a validator function that returns the value if it is in the range. Otherwise it returns the value, modulo the max of the range. Allows negative values.

Parameters

- **value** – a value to test
- **values** – A set of values that are valid

`pymeasure.instruments.validators.strict_discrete_range` (*value*, *values*, *step*)

Provides a validator function that returns the value if its value is less than the maximum and greater than the minimum of the range and is a multiple of step. Otherwise it raises a ValueError.

Parameters

- **value** – A value to test
- **values** – A range of values (range, list, etc.)
- **step** – Minimum stepsize (resolution limit)

Raises ValueError if the value is out of the range

`pymeasure.instruments.validators.strict_discrete_set` (*value*, *values*)

Provides a validator function that returns the value if it is in the discrete set. Otherwise it raises a ValueError.

Parameters

- **value** – A value to test
- **values** – A set of values that are valid

Raises ValueError if the value is not in the set

`pymeasure.instruments.validators.strict_range` (*value*, *values*)

Provides a validator function that returns the value if its value is less than the maximum and greater than the minimum of the range. Otherwise it raises a ValueError.

Parameters

- **value** – A value to test

- **values** – A range of values (range, list, etc.)

Raises ValueError if the value is out of the range

`pymeasure.instruments.validators.truncated_discrete_set` (*value, values*)

Provides a validator function that returns the value if it is in the discrete set. Otherwise, it returns the smallest value that is larger than the value.

Parameters

- **value** – A value to test
- **values** – A set of values that are valid

`pymeasure.instruments.validators.truncated_range` (*value, values*)

Provides a validator function that returns the value if it is in the range. Otherwise it returns the closest range bound.

Parameters

- **value** – A value to test
- **values** – A set of values that are valid

7.3 Comedi data acquisition

The Comedi libraries provide a convenient method for interacting with data acquisition cards, but are restricted to Linux compatible operating systems.

`pymeasure.instruments.comedi.getAI` (*device, channel, range=None*)

Returns the analog input channel as specified for a given device

`pymeasure.instruments.comedi.getAO` (*device, channel, range=None*)

Returns the analog output channel as specified for a given device

`pymeasure.instruments.comedi.readAI` (*device, channel, range=None, count=1*)

Reads a single measurement (*count=1*) from the analog input channel of the device specified. Multiple readings can be preformed with *count* not equal to one, which are seperated by an arbitrary time

`pymeasure.instruments.comedi.writeAO` (*device, channel, voltage, range=None*)

Writes a single voltage to the analog output channel of the device specified

7.4 Resource Manager

The `list_resources` function provides an interface to check connected instruments interactively.

`pymeasure.instruments.list_resources` ()

Prints the available resources, and returns a list of VISA resource names

```
resources = list_resources()
#prints (e.g.)
#0 : GPIB0::22::INSTR : Agilent Technologies,34410A,*****
#1 : GPIB0::26::INSTR : Keithley Instruments Inc., Model 2612, *****
dmm = Agilent34410(resources[0])
```

Instruments by manufacturer:

7.5 Advantest

This section contains specific documentation on the Advantest instruments that are implemented. If you are interested in an instrument not included, please consider *adding the instrument*.

7.6 Agilent

This section contains specific documentation on the Agilent instruments that are implemented. If you are interested in an instrument not included, please consider *adding the instrument*.

7.6.1 Agilent 8257D Signal Generator

class `pymeasure.instruments.agilent.Agilent8257D` (*adapter*, *delay=0.02*, ***kwargs*)

Bases: `pymeasure.instruments.instrument.Instrument`

Represents the Agilent 8257D Signal Generator and provides a high-level interface for interacting with the instrument.

```
generator = Agilent8257D("GPIB::1")

generator.power = 0           # Sets the output power to 0 dBm
generator.frequency = 5      # Sets the output frequency to 5 GHz
generator.enable()          # Enables the output
```

property `amplitude_depth`

A floating point property that controls the amplitude modulation in percent, which can take values from 0 to 100 %.

property `amplitude_source`

A string property that controls the source of the amplitude modulation signal, which can take the values: 'internal', 'internal 2', 'external', and 'external 2'.

property `center_frequency`

A floating point property that represents the center frequency in Hz. This property can be set.

config `amplitude_modulation` (*frequency=1000.0*, *depth=100.0*, *shape='sine'*)

Configures the amplitude modulation of the output signal.

Parameters

- **frequency** – A modulation frequency for the internal oscillator
- **depth** – A linear depth percentage
- **shape** – A string that describes the shape for the internal oscillator

config `low_freq_out` (*source='internal'*, *amplitude=3*)

Configures the low-frequency output signal.

Parameters

- **source** – The source for the low-frequency output signal.
- **amplitude** – Amplitude of the low-frequency output

config `pulse_modulation` (*frequency=1000.0*, *input='square'*)

Configures the pulse modulation of the output signal.

Parameters

- **frequency** – A pulse rate frequency in Hertz
- **input** – A string that describes the internal pulse input

config_step_sweep ()

Configures a step sweep through frequency

disable ()

Disables the output of the signal.

disable_amplitude_modulation ()

Disables amplitude modulation of the output signal.

disable_low_freq_out ()

Disables low frequency output

disable_modulation ()

Disables the signal modulation.

disable_pulse_modulation ()

Disables pulse modulation of the output signal.

property dwell_time

A floating point property that represents the settling time in seconds at the current frequency or power setting. This property can be set.

enable ()

Enables the output of the signal.

enable_amplitude_modulation ()

Enables amplitude modulation of the output signal.

enable_low_freq_out ()

Enables low frequency output

enable_pulse_modulation ()

Enables pulse modulation of the output signal.

property frequency

A floating point property that represents the output frequency in Hz. This property can be set.

property has_amplitude_modulation

Reads a boolean value that is True if the amplitude modulation is enabled.

property has_modulation

Reads a boolean value that is True if the modulation is enabled.

property has_pulse_modulation

Reads a boolean value that is True if the pulse modulation is enabled.

property internal_frequency

A floating point property that controls the frequency of the internal oscillator in Hertz, which can take values from 0.5 Hz to 1 MHz.

property internal_shape

A string property that controls the shape of the internal oscillations, which can take the values: 'sine', 'triangle', 'square', 'ramp', 'noise', 'dual-sine', and 'swept-sine'.

property is_enabled

Reads a boolean value that is True if the output is on.

property low_freq_out_amplitude

A floating point property that controls the peak voltage (amplitude) of the low frequency output in volts, which can take values from 0-3.5V

property low_freq_out_source

A string property which controls the source of the low frequency output, which can take the values 'internal [2]' for the internal source, or 'function [2]' for an internal function generator which can be configured.

property power

A floating point property that represents the output power in dBm. This property can be set.

property pulse_frequency

A floating point property that controls the pulse rate frequency in Hertz, which can take values from 0.1 Hz to 10 MHz.

property pulse_input

A string property that controls the internally generated modulation input for the pulse modulation, which can take the values: 'square', 'free-run', 'triggered', 'doublet', and 'gated'.

property pulse_source

A string property that controls the source of the pulse modulation signal, which can take the values: 'internal', 'external', and 'scalar'.

shutdown ()

Shuts down the instrument by disabling any modulation and the output signal.

property start_frequency

A floating point property that represents the start frequency in Hz. This property can be set.

property start_power

A floating point property that represents the stop power in dBm. This property can be set.

start_step_sweep ()

Starts a step sweep.

property step_points

An integer number of points in a step sweep. This property can be set.

property stop_frequency

A floating point property that represents the stop frequency in Hz. This property can be set.

stop_step_sweep ()

Stops a step sweep.

7.6.2 Agilent 8722ES Vector Network Analyzer

class `pymeasure.instruments.agilent.Agilent8722ES` (*resourceName*, ***kwargs*)

Bases: `pymeasure.instruments.instrument.Instrument`

Represents the Agilent8722ES Vector Network Analyzer and provides a high-level interface for taking scans of the scattering parameters.

property data

Returns the real and imaginary data from the last scan

disable_averaging ()

Disables averaging

property frequencies

Returns a list of frequencies from the last scan

is_averaging ()
Returns True if averaging is enabled

log_magnitude (*real, imaginary*)
Returns the magnitude in dB from a real and imaginary number or numpy arrays

magnitude (*real, imaginary*)
Returns the magnitude from a real and imaginary number or numpy arrays

phase (*real, imaginary*)
Returns the phase in degrees from a real and imaginary number or numpy arrays

scan (*averages=1, blocking=True, timeout=25, delay=0.1*)
Initiates a scan with the number of averages specified and blocks until the operation is complete if blocking is True

scan_continuous ()
Initiates a continuous scan

property scan_points
Gets the number of scan points

scan_single ()
Initiates a single scan

set_IF_bandwidth (*bandwidth*)
Sets the resolution bandwidth (IF bandwidth)

set_averaging (*averages*)
Turns on averaging of a specific number between 0 and 999

set_fixed_frequency (*frequency*)
Sets the scan to be of only one frequency in Hz

property start_frequency
A floating point property that represents the start frequency in Hz. This property can be set.

property stop_frequency
A floating point property that represents the stop frequency in Hz. This property can be set.

property sweep_time
A floating point property that represents the sweep time in seconds. This property can be set.

7.6.3 Agilent E4408B Spectrum Analyzer

class `pymeasure.instruments.agilent.AgilentE4408B` (*resourceName, **kwargs*)
Bases: `pymeasure.instruments.instrument.Instrument`

Represents the AgilentE4408B Spectrum Analyzer and provides a high-level interface for taking scans of high-frequency spectrums

property center_frequency
A floating point property that represents the center frequency in Hz. This property can be set.

property frequencies
Returns a numpy array of frequencies in Hz that correspond to the current settings of the instrument.

property frequency_points
An integer property that represents the number of frequency points in the sweep. This property can take values from 101 to 8192.

property frequency_step

A floating point property that represents the frequency step in Hz. This property can be set.

property start_frequency

A floating point property that represents the start frequency in Hz. This property can be set.

property stop_frequency

A floating point property that represents the stop frequency in Hz. This property can be set.

property sweep_time

A floating point property that represents the sweep time in seconds. This property can be set.

trace (*number=1*)

Returns a numpy array of the data for a particular trace based on the trace number (1, 2, or 3).

trace_df (*number=1*)

Returns a pandas DataFrame containing the frequency and peak data for a particular trace, based on the trace number (1, 2, or 3).

7.6.4 Agilent E4980 LCR Meter

class `pymeasure.instruments.agilent.AgilentE4980` (*adapter, **kwargs*)

Bases: `pymeasure.instruments.instrument.Instrument`

Represents LCR meter E4980A/AL

property ac_current

AC current level, in Amps

property ac_voltage

AC voltage level, in Volts

aperture (*time=None, averages=1*)

Set and get aperture.

Parameters

- **time** – integration time as string: SHORT, MED, LONG (case insensitive); if None, get values
- **averages** – number of averages, numeric

freq_sweep (*freq_list, return_freq=False*)

Run frequency list sweep using sequential trigger.

Parameters

- **freq_list** – list of frequencies
- **return_freq** – if True, returns the frequencies read from the instrument

Returns values as configured with `mode`

property frequency

AC frequency (range depending on model), in Hertz

property impedance

Measured data A and B, according to `mode`

property mode

Select quantities to be measured:

- CPD: Parallel capacitance [F] and dissipation factor [number]

- CPQ: Parallel capacitance [F] and quality factor [number]
- CPG: Parallel capacitance [F] and parallel conductance [S]
- CPRP: Parallel capacitance [F] and parallel resistance [Ohm]
- CSD: Series capacitance [F] and dissipation factor [number]
- CSQ: Series capacitance [F] and quality factor [number]
- CSRS: Series capacitance [F] and series resistance [Ohm]
- LPD: Parallel inductance [H] and dissipation factor [number]
- LPQ: Parallel inductance [H] and quality factor [number]
- LPG: Parallel inductance [H] and parallel conductance [S]
- LPRP: Parallel inductance [H] and parallel resistance [Ohm]
- LSD: Series inductance [H] and dissipation factor [number]
- LSQ: Series inductance [H] and quality factor [number]
- LSRS: Series inductance [H] and series resistance [Ohm]
- RX: Resistance [Ohm] and reactance [Ohm]
- ZTD: Impedance, magnitude [Ohm] and phase [deg]
- ZTR: Impedance, magnitude [Ohm] and phase [rad]
- GB: Conductance [S] and susceptance [S]
- YTD: Admittance, magnitude [Ohm] and phase [deg]
- YTR: Admittance magnitude [Ohm] and phase [rad]

property trigger_source

Select trigger source; accept the values:

- HOLD: manual
- INT: internal
- BUS: external bus (GPIB/LAN/USB)
- EXT: external connector

7.6.5 Agilent 34410A Multimeter

class pymeasure.instruments.agilent.**Agilent34410A** (*adapter, delay=0.02, **kwargs*)

Bases: pymeasure.instruments.instrument.Instrument

Represent the HP/Agilent/Keysight 34410A and related multimeters.

Implemented measurements: voltage_dc, voltage_ac, current_dc, current_ac, resistance, resistance_4w

property current_ac

AC current, in Amps

property current_dc

DC current, in Amps

property resistance

Resistance, in Ohms

property resistance_4w

Four-wires (remote sensing) resistance, in Ohms

property voltage_ac

AC voltage, in Volts

property voltage_dc

DC voltage, in Volts

7.6.6 Agilent 4155/4156 Semiconductor Parameter Analyzer

class pymeasure.instruments.agilent.agilent4156.**Agilent4156** (*resourceName*,
***kwargs*)

Bases: pymeasure.instruments.instrument.Instrument

Represents the Agilent 4155/4156 Semiconductor Parameter Analyzer and provides a high-level interface for taking current-voltage (I-V) measurements.

```

from pymeasure.instruments.agilent import Agilent4156

# explicitly define r/w terminations; set sufficiently large timeout or None.
smu = Agilent4156("GPIB0::25", read_termination = '\n', write_termination = '\n',
↳ timeout=None)

# reset the instrument
smu.reset ()

# define configuration file for instrument and load config
smu.configure("configuration_file.json")

# save data variables, some or all of which are defined in the json config file.
smu.save(['VC', 'IC', 'VB', 'IB'])

# take measurements
status = smu.measure ()

# measured data is a pandas dataframe and can be exported to csv.
data = smu.get_data(path='./t1.csv')

```

The JSON file is an ascii text configuration file that defines the settings of each channel on the instrument. The JSON file is used to configure the instrument using the convenience function `configure()` as shown in the example above. For example, the instrument setup for a bipolar transistor measurement is shown below.

```

{
  "SMU1": {
    "voltage_name" : "VC",
    "current_name" : "IC",
    "channel_function" : "VAR1",
    "channel_mode" : "V",
    "series_resistance" : "0OHM"
  },
  "SMU2": {
    "voltage_name" : "VB",
    "current_name" : "IB",

```

(continues on next page)

(continued from previous page)

```

        "channel_function" : "VAR2",
        "channel_mode" : "I",
        "series_resistance" : "0OHM"
    },

    "SMU3": {
        "voltage_name" : "VE",
        "current_name" : "IE",
        "channel_function" : "CONS",
        "channel_mode" : "V",
        "constant_value" : 0,
        "compliance" : 0.1
    },

    "SMU4": {
        "voltage_name" : "VS",
        "current_name" : "IS",
        "channel_function" : "CONS",
        "channel_mode" : "V",
        "constant_value" : 0,
        "compliance" : 0.1
    },

    "VAR1": {
        "start" : 1,
        "stop" : 2,
        "step" : 0.1,
        "spacing" : "LINEAR",
        "compliance" : 0.1
    },

    "VAR2": {
        "start" : 0,
        "step" : 10e-6,
        "points" : 3,
        "compliance" : 2
    }
}

```

property analyzer_mode

A string property that controls the instrument operating mode.

- Values: SWEEP, SAMPLING

```
smu.analyzer_mode = "SWEEP"
```

configure (*config_file*)

Convenience function to configure the channel setup and sweep using a JSON (JavaScript Object Notation) configuration file.

Parameters *config_file* – JSON file to configure instrument channels.

```
instr.configure('config.json')
```

property data_variables

Gets a string list of data variables for which measured data is available. This looks for all the variables

saved by the `save()` and `save_var()` methods and returns it. This is useful for creation of dataframe headers.

Returns List

```
header = instr.data_variables
```

property `delay_time`

A floating point property that measurement delay time in seconds, which can take the values from 0 to 65s in 0.1s steps.

```
instr.delay_time = 1 # delay time of 1-sec
```

`disable_all()`

Disables all channels in the instrument.

```
instr.disable_all()
```

`get_data(path=None)`

Gets the measurement data from the instrument after completion. If the measurement period is set to `INF` in the `measure()` method, then the measurement must be stopped using `stop()` before getting valid data.

Parameters `path` – Path for optional data export to CSV.

Returns Pandas Dataframe

```
df = instr.get_data(path='./datafolder/data1.csv')
```

property `hold_time`

A floating point property that measurement hold time in seconds, which can take the values from 0 to 655s in 1s steps.

```
instr.hold_time = 2 # hold time of 2-secs.
```

property `integration_time`

A string property that controls the integration time.

- Values: SHORT, MEDIUM, LONG

```
instr.integration_time = "MEDIUM"
```

`measure(period='INF', points=100)`

Performs a single measurement and waits for completion in sweep mode. In sampling mode, the measurement period and number of points can be specified.

Parameters

- **period** – Period of sampling measurement from 6E-6 to 1E11 seconds. Default setting is `INF`.
- **points** – Number of samples to be measured, from 1 to 10001. Default setting is 100.

`save(trace_list)`

Save the voltage or current in the instrument display list

Parameters `trace_list` – A list of channel variables whose measured data should be saved.

A maximum of 8 variables are allowed. If only one variable is being saved, a string can be specified.

```
instr.save(['IC', 'IB', 'VC', 'VB']) #for list of variables
instr.save('IC') #for single variable
```

save_var (*trace_list*)

Save the voltage or current in the instrument variable list. This is useful if one or two more variables need to be saved in addition to the 8 variables allowed by *save()*.

Parameters *trace_list* – A list of channel variables whose measured data should be saved. A maximum of 2 variables are allowed. If only one variable is being saved, a string can be specified.

```
instr.save_var(['VA', 'VB'])
```

stop ()

Stops the ongoing measurement

```
instr.stop()
```

class pymeasure.instruments.agilent.agilent4156.**SMU** (*resourceName*, *channel*, ***kwargs*)

Bases: pymeasure.instruments.instrument.Instrument

property channel_function

A string property that controls the SMU<n> channel function.

- Values: VAR1, VAR2, VARD or CONS.

```
instr.smul.channel_function = "VAR1"
```

property channel_mode

A string property that controls the SMU<n> channel mode.

- Values: V, I or COMM

VPULSE AND IPULSE are not yet supported.

```
instr.smul.channel_mode = "V"
```

property compliance

This command sets the *constant* compliance value of SMU<n>. If the SMU channel is setup as a variable (VAR1, VAR2, VARD) then compliance limits are set by the variable definition.

- Value: Voltage in (-200V, 200V) and current in (-1A, 1A) based

on *channel_mode()*.

```
instr.smul.compliance = 0.1
```

property constant_value

This command sets the constant source value of SMU<n>. You use this command only if *channel_function()* is CONS and also *channel_mode()* should not be COMM.

Parameters *const_value* – Voltage in (-200V, 200V) and current in (-1A, 1A). Voltage or current depends on if *channel_mode()* is set to V or I.

```
instr.smul.constant_value = 1
```

property current_name

Define the current name of the channel.

If input is greater than 6 characters long or starts with a number, the name is autocorrected and prepended with 'a'. Event is logged.

```
instr.smul.current_name = "Ibase"
```

property disable

This command deletes the settings of SMU<n>.

```
instr.smul.disable()
```

property series_resistance

This command controls the series resistance of SMU<n>.

- Values: 0OHM, 10KOHM, 100KOHM, or 1MOHM

```
instr.smul.series_resistance = "10KOHM"
```

property voltage_name

Define the voltage name of the channel.

If input is greater than 6 characters long or starts with a number, the name is autocorrected and prepended with 'a'. Event is logged.

```
instr.smul.voltage_name = "Vbase"
```

class `pymeasure.instruments.agilent.agilent4156.VAR1` (*resourceName*, ***kwargs*)

Bases: `pymeasure.instruments.agilent.agilent4156.VARX`

Class to handle all the specific definitions needed for VAR1. Most common methods are inherited from base class.

property spacing

This command selects the sweep type of VAR1.

- Values: LINEAR, LOG10, LOG25, LOG50.

class `pymeasure.instruments.agilent.agilent4156.VAR2` (*resourceName*, ***kwargs*)

Bases: `pymeasure.instruments.agilent.agilent4156.VARX`

Class to handle all the specific definitions needed for VAR2. Common methods are imported from base class.

property points

This command sets the number of sweep steps of VAR2. You use this command only if there is an SMU or VSU whose function (FCTN) is VAR2.

```
instr.var2.points = 10
```

class `pymeasure.instruments.agilent.agilent4156.VARD` (*resourceName*, ***kwargs*)

Bases: `pymeasure.instruments.instrument.Instrument`

Class to handle all the definitions needed for VARD. VARD is always defined in relation to VAR1.

property compliance

This command sets the sweep COMPLIANCE value of VARD.

```
instr.vard.compliance = 0.1
```

property offset

This command sets the OFFSET value of VARD. For each step of sweep, the output values of VAR1' are determined by the following equation: $VARD = VAR1 \times RATIO + OFFSET$ You use this command only if there is an SMU or VSU whose function is VARD.

```
instr.vard.offset = 1
```

property ratio

This command sets the RATIO of VAR1'. For each step of sweep, the output values of VAR1' are determined by the following equation: VAR1' = VAR1 * RATio + OFFSet You use this command only if there is an SMU or VSU whose function (FCTN) is VAR1'.

```
instr.vard.ratio = 1
```

```
class pymeasure.instruments.agilent.agilent4156.VARX(resourceName, var_name,
                                                    **kwargs)
```

Bases: pymeasure.instruments.instrument.Instrument

Base class to define sweep variable settings

property compliance

Sets the sweep COMPLIANCE value.

```
instr.var1.compliance = 0.1
```

property start

Sets the sweep START value.

```
instr.var1.start = 0
```

property step

Sets the sweep STEP value.

```
instr.var1.step = 0.1
```

property stop

Sets the sweep STOP value.

```
instr.var1.stop = 3
```

```
class pymeasure.instruments.agilent.agilent4156.VMU(resourceName, channel,
                                                    **kwargs)
```

Bases: pymeasure.instruments.instrument.Instrument

property channel_mode

A string property that controls the VMU<n> channel mode.

- Values: V, DVOL

property disable

This command disables the settings of VMU<n>.

```
instr.vmu1.disable()
```

property voltage_name

Define the voltage name of the VMU channel.

If input is greater than 6 characters long or starts with a number, the name is autocorrected and prepended with 'a'. Event is logged.

```
instr.vmu1.voltage_name = "Vanode"
```

```
class pymeasure.instruments.agilent.agilent4156.VSU(resourceName, channel,  
                                                **kwargs)
```

Bases: `pymeasure.instruments.instrument.Instrument`

property channel_function

A string property that controls the VSU channel function.

- Value: VAR1, VAR2, VARD or CONS.

property channel_mode

Get channel mode of VSU<n>.

property constant_value

This command sets the constant source value of VSU<n>.

```
instr.vsu1.constant_value = 0
```

property disable

This command deletes the settings of VSU<n>.

```
instr.vsu1.disable()
```

property voltage_name

Define the voltage name of the VSU channel

If input is greater than 6 characters long or starts with a number, the name is autocorrected and prepended with 'a'. Event is logged.

```
instr.vsu1.voltage_name = "Ve"
```

7.6.7 Agilent 33220A Arbitrary Waveform Generator

```
class pymeasure.instruments.agilent.Agilent33220A(adapter, **kwargs)
```

Bases: `pymeasure.instruments.instrument.Instrument`

Represents the Agilent 33220A Arbitrary Waveform Generator.

property amplitude

A floating point property that controls the voltage amplitude of the output waveform in V, from 10e-3 V to 10 V. Can be set.

property amplitude_unit

A string property that controls the units of the amplitude. Valid values are Vpp (default), Vrms, and dBm. Can be set.

beep()

Causes a system beep.

property beeper_state

A boolean property that controls the state of the beeper. Can be set.

property burst_mode

A string property that controls the burst mode. Valid values are: TRIG<GERED>, GAT<ED>. This setting can be set.

property burst_ncycles

An integer property that sets the number of cycles to be output when a burst is triggered. Valid values are 1 to 50000. This can be set.

property burst_state

A boolean property that controls whether the burst mode is on (True) or off (False). Can be set.

check_errors ()

Read all errors from the instrument.

property frequency

A floating point property that controls the frequency of the output waveform in Hz, from 1e-6 (1 uHz) to 20e+6 (20 MHz), depending on the specified function. Can be set.

property offset

A floating point property that controls the voltage offset of the output waveform in V, from 0 V to 4.995 V, depending on the set voltage amplitude (maximum offset = $(10 - \text{voltage}) / 2$). Can be set.

property output

A boolean property that turns on (True) or off (False) the output of the function generator. Can be set.

property pulse_dutycycle

A floating point property that controls the duty cycle of a pulse waveform function in percent. Can be set.

property pulse_hold

A string property that controls if either the pulse width or the duty cycle is retained when changing the period or frequency of the waveform. Can be set to: WIDT<H> or DCYC<LE>.

property pulse_period

A floating point property that controls the period of a pulse waveform function in seconds, ranging from 200 ns to 2000 s. Can be set and overwrites the frequency for *all* waveforms. If the period is shorter than the pulse width + the edge time, the edge time and pulse width will be adjusted accordingly.

property pulse_transition

A floating point property that controls the the edge time in seconds for both the rising and falling edges. It is defined as the time between 0.1 and 0.9 of the threshold. Valid values are between 5 ns to 100 ns. Can be set.

property pulse_width

A floating point property that controls the width of a pulse waveform function in seconds, ranging from 20 ns to 2000 s, within a set of restrictions depending on the period. Can be set.

property ramp_symmetry

A floating point property that controls the symmetry percentage for the ramp waveform. Can be set.

property remote_local_state

A string property that controls the remote/local state of the function generator. Valid values are: LOC<AL>, REM<OTE>, RWL<OCK>. This setting can only be set.

property shape

A string property that controls the output waveform. Can be set to: SIN<USOID>, SQU<ARE>, RAMP, PULS<E>, NOIS<E>, DC, USER.

property square_dutycycle

A floating point property that controls the duty cycle of a square waveform function in percent. Can be set.

trigger ()

Send a trigger signal to the function generator.

property trigger_source

A string property that controls the trigger source. Valid values are: IMM<EDIATE> (internal), EXT<ERNAL> (rear input), BUS (via trigger command). This setting can be set.

property trigger_state

A boolean property that controls whether the output is triggered (True) or not (False). Can be set.

property voltage_high

A floating point property that controls the upper voltage of the output waveform in V, from -4.990 V to 5 V (must be higher than low voltage). Can be set.

property voltage_low

A floating point property that controls the lower voltage of the output waveform in V, from -5 V to 4.990 V (must be lower than high voltage). Can be set.

wait_for_trigger (*timeout=3600, should_stop=<function Agilent33220A.<lambda>>*)

Wait until the triggering has finished or timeout is reached.

Parameters

- **timeout** – The maximum time the waiting is allowed to take. If timeout is exceeded, a `TimeoutError` is raised. If timeout is set to zero, no timeout will be used.
- **should_stop** – Optional function (returning a bool) to allow the waiting to be stopped before its end.

7.6.8 Agilent 33500 Function/Arbitrary Waveform Generator Family

class `pymeasure.instruments.agilent.Agilent33500` (*adapter, **kwargs*)

Bases: `pymeasure.instruments.instrument.Instrument`

Represents the Agilent 33500 Function/Arbitrary Waveform Generator family. Individual devices are represented by subclasses.

```
generator = Agilent33500("GPIB::1")

generator.shape = 'SIN'           # Sets the output signal shape to sine
generator.frequency = 1e3         # Sets the output frequency to 1 kHz
generator.amplitude = 5          # Sets the output amplitude to 5 Vpp
generator.output = 'on'         # Enables the output

generator.shape = 'ARB'          # Set shape to arbitrary
generator.arb_srate = 1e6        # Set sample rate to 1MSa/s

generator.data_volatile_clear()  # Clear volatile internal memory
generator.data_arb(              # Send data points of arbitrary waveform
    'test',
    range(-10000, 10000, +20),    # In this case a simple ramp
    data_format='DAC')           # Data format is set to 'DAC'
)
generator.arb_file = 'test'      # Select the transmitted waveform 'test'
```

property amplitude

A floating point property that controls the voltage amplitude of the output waveform in V, from 10e-3 V to 10 V. Depends on the output impedance. Can be set.

property amplitude_unit

A string property that controls the units of the amplitude. Valid values are VPP (default), VRMS, and DBM. Can be set.

property arb_advance

A string property that selects how the device advances from data point to data point. Can be set to 'TRIG<GER>' or 'SRAT<E>' (default).

property arb_file

A string property that selects the arbitrary signal from the volatile memory of the device. String has to match an existing arb signal in volatile memore (set by `data_arb()`). Can be set.

property arb_filter

A string property that selects the filter setting for arbitrary signals. Can be set to 'NORM<AL>', 'STEP' and 'OFF'.

property arb_srate

An floating point property that sets the sample rate of the currently selected arbitrary signal. Valid values are 1 μ Sa/s to 250 MSa/s (maximum range, can be lower depending on your device). This can be set.

beep ()

Causes a system beep.

property burst_mode

A string property that controls the burst mode. Valid values are: TRIG<GERED>, GAT<ED>. This setting can be set.

property burst_ncycles

An integer property that sets the number of cycles to be output when a burst is triggered. Valid values are 1 to 100000. This can be set.

property burst_period

A floating point property that controls the period of subsequent bursts. Has to follow the equation $\text{burst_period} > (\text{burst_ncycles} / \text{frequency}) + 1 \mu\text{s}$. Valid values are 1 μs to 8000 s. Can be set.

property burst_state

A boolean property that controls whether the burst mode is on (True) or off (False). Can be set.

check_errors ()

Read all errors from the instrument.

clear_display ()

Removes a text message from the display.

data_arb (arb_name, data_points, data_format='DAC')

Uploads an arbitrary trace into the volatile memory of the device. The `data_points` can be given as comma separated 16 bit DAC values (ranging from -32767 to +32767), as comma separated floating point values (ranging from -1.0 to +1.0) or as a binary data stream. Check the manual for more information. The storage depends on the device type and ranges from 8 Sa to 16 MSa (maximum). *TODO: Binary is not yet implemented*

Parameters

- **arb_name** – The name of the trace in the volatile memory. This is used to access the trace.
- **data_points** – Individual points of the trace. The format depends on the format parameter.
 - format = 'DAC' (default): Accepts list of integer values ranging from -32767 to +32767. Minimum of 8 a maximum of 65536 points.
 - format = 'float': Accepts list of floating point values ranging from -1.0 to +1.0. Minimum of 8 a maximum of 65536 points.
 - format = 'binary': Accepts a binary stream of 8 bit data.
- **data_format** – Defines the format of `data_points`. Can be 'DAC' (default), 'float' or 'binary'. See documentation on parameter `data_points` above.

data_volatile_clear()

Clear all arbitrary signals from the volatile memory. This should be done if the same name is used continuously to load different arbitrary signals into the memory, since an error will occur if a trace is loaded which already exists in the memory.

property display

A string property which is displayed on the front panel of the device. Can be set.

property ext_trig_out

A boolean property that controls whether the trigger out signal is active (True) or not (False). This signal is output from the Ext Trig connector on the rear panel in Burst and Wobbel mode. Can be set.

property frequency

A floating point property that controls the frequency of the output waveform in Hz, from 1 uHz to 120 MHz (maximum range, can be lower depending on your device), depending on the specified function. Can be set.

property id

Reads the instrument identification

property offset

A floating point property that controls the voltage offset of the output waveform in V, from 0 V to 4.995 V, depending on the set voltage amplitude (maximum offset = $(V_{max} - voltage) / 2$). Can be set.

property output

A boolean property that turns on (True, 'on') or off (False, 'off') the output of the function generator. Can be set.

property output_load

Sets the expected load resistance (should be the load impedance connected to the output. The output impedance is always 50 Ohm, this setting can be used to correct the displayed voltage for loads unmatched to 50 Ohm. Valid values are between 1 and 10 kOhm or INF for high impedance. No validator is used since both numeric and string inputs are accepted, thus a value outside the range will not return an error. Can be set.

property pulse_dutycycle

A floating point property that controls the duty cycle of a pulse waveform function in percent, from 0% to 100%. Can be set.

property pulse_hold

A string property that controls if either the pulse width or the duty cycle is retained when changing the period or frequency of the waveform. Can be set to: WIDT<H> or DCYC<LE>.

property pulse_period

A floating point property that controls the period of a pulse waveform function in seconds, ranging from 33 ns to 1e6 s. Can be set and overwrites the frequency for *all* waveforms. If the period is shorter than the pulse width + the edge time, the edge time and pulse width will be adjusted accordingly.

property pulse_transition

A floating point property that controls the edge time in seconds for both the rising and falling edges. It is defined as the time between the 10% and 90% thresholds of the edge. Valid values are between 8.4 ns to 1 μ s. Can be set.

property pulse_width

A floating point property that controls the width of a pulse waveform function in seconds, ranging from 16 ns to 1e6 s, within a set of restrictions depending on the period. Can be set.

property ramp_symmetry

A floating point property that controls the symmetry percentage for the ramp waveform, from 0.0% to 100.0% Can be set.

property shape

A string property that controls the output waveform. Can be set to: SIN<USOID>, SQU<ARE>, TRI<ANGLE>, RAMP, PULS<E>, PRBS, NOIS<E>, ARB, DC.

property square_dutycycle

A floating point property that controls the duty cycle of a square waveform function in percent, from 0.01% to 99.98%. The duty cycle is limited by the frequency and the minimal pulse width of 16 ns. See manual for more details. Can be set.

trigger ()

Send a trigger signal to the function generator.

property trigger_source

A string property that controls the trigger source. Valid values are: IMM<EDIATE> (internal), EXT<ERNAL> (rear input), BUS (via trigger command). This setting can be set.

property voltage_high

A floating point property that controls the upper voltage of the output waveform in V, from -4.990 V to 5 V (must be higher than low voltage by at least 1 mV). Can be set.

property voltage_low

A floating point property that controls the lower voltage of the output waveform in V, from -5 V to 4.990 V (must be lower than high voltage by at least 1 mV). Can be set.

wait_for_trigger (*timeout=3600, should_stop=<function Agilent33500.<lambda>>*)

Wait until the triggering has finished or timeout is reached.

Parameters

- **timeout** – The maximum time the waiting is allowed to take. If timeout is exceeded, a `TimeoutError` is raised. If timeout is set to zero, no timeout will be used.
- **should_stop** – Optional function (returning a bool) to allow the waiting to be stopped before its end.

7.6.9 Agilent 33521A Function/Arbitrary Waveform Generator

class `pymeasure.instruments.agilent.Agilent33521A` (*adapter, **kwargs*)

Bases: `pymeasure.instruments.agilent.agilent33500.Agilent33500`

Represents the Agilent 33521A Function/Arbitrary Waveform Generator. This documentation page shows only methods different from the parent class `Agilent33500`.

property arb_srate

An floating point property that sets the sample rate of the currently selected arbitrary signal. Valid values are 1 μ Sa/s to 250 MSa/s. This can be set.

property frequency

A floating point property that controls the frequency of the output waveform in Hz, from 1 uHz to 30 MHz, depending on the specified function. Can be set.

7.7 AMI

This section contains specific documentation on the AMI instruments that are implemented. If you are interested in an instrument not included, please consider *adding the instrument*.

7.7.1 AMI 430 Power Supply

class `pymessage.instruments.ami.AMI430` (*resourceName*, ***kwargs*)

Bases: `pymessage.instruments.instrument.Instrument`

Represents the AMI 430 Power supply and provides a high-level for interacting with the instrument.

```
magnet = AMI430("TCP/IP::web.address.com::7180::SOCKET")

magnet.coilconst = 1.182           # kGauss/A
magnet.voltage_limit = 2.2        # Sets the voltage limit in V

magnet.target_current = 10        # Sets the target current to 10 A
magnet.target_field = 1           # Sets target field to 1 kGauss

magnet.ramp_rate_current = 0.0357 # Sets the ramp rate in A/s
magnet.ramp_rate_field = 0.0422   # Sets the ramp rate in kGauss/s
magnet.ramp                       # Initiates the ramping
magnet.pause                       # Pauses the ramping
magnet.status                       # Returns the status of the magnet

magnet.ramp_to_current(5)         # Ramps the current to 5 A

magnet.shutdown()                 # Ramps the current to zero and disables_
↪output
```

property coilconst

A floating point property that sets the coil constant in kGauss/A.

disable_persistent_switch()

Disables the persistent switch.

enable_persistent_switch()

Enables the persistent switch.

property field

Reads the field in kGauss of the magnet.

has_persistent_switch_enabled()

Returns a boolean if the persistent switch is enabled.

property magnet_current

Reads the current in Amps of the magnet.

pause()

Pauses the ramping of the magnetic field.

ramp()

Initiates the ramping of the magnetic field to set current/field with ramping rate previously set.

property ramp_rate_current

A floating point property that sets the current ramping rate in A/s.

property ramp_rate_field

A floating point property that sets the field ramping rate in kGauss/s.

ramp_to_current (*current, rate*)

Heats up the persistent switch and ramps the current with set ramp rate.

ramp_to_field (*field, rate*)

Heats up the persistent switch and ramps the current with set ramp rate.

shutdown (*ramp_rate=0.0357*)

Turns on the persistent switch, ramps down the current to zero, and turns off the persistent switch.

property state

Reads the field in kGauss of the magnet.

property supply_current

Reads the current in Amps of the power supply.

property target_current

A floating point property that sets the target current in A for the magnet.

property target_field

A floating point property that sets the target field in kGauss for the magnet.

property voltage_limit

A floating point property that sets the voltage limit for charging/discharging the magnet.

wait_for_holding (*should_stop=<function AMI430.<lambda>>, timeout=800, interval=0.1*)**zero** ()

Initiates the ramping of the magnetic field to zero current/field with ramping rate previously set.

7.8 Anritsu

This section contains specific documentation on the Anritsu instruments that are implemented. If you are interested in an instrument not included, please consider [adding the instrument](#).

7.8.1 Anritsu MG3692C Signal Generator

class `pymessage.instruments.anritsu.AnritsuMG3692C` (*resourceName, **kwargs*)

Bases: `pymessage.instruments.instrument.Instrument`

Represents the Anritsu MG3692C Signal Generator

disable ()

Disables the signal output.

enable ()

Enables the signal output.

property frequency

A floating point property that represents the output frequency in Hz. This property can be set.

property output

A boolean property that represents the signal output state. This property can be set to control the output.

property power

A floating point property that represents the output power in dBm. This property can be set.

shutdown ()
Shuts down the instrument, putting it in a safe state.

7.8.2 Anritsu MS9710C Optical Spectrum Analyzer

class `pymeasure.instruments.anritsu.AnritsuMS9710C` (*adapter*, ***kwargs*)
Bases: `pymeasure.instruments.instrument.Instrument`
Anritsu MS9710C Optical Spectrum Analyzer.

property analysis
Analysis Control

property analysis_result
Read back analysis result from current scan.

property average_point
Number of averages to take on each point (2-1000), or OFF

property average_sweep
Number of averages to make on a sweep (2-1000) or OFF

center_at_peak (***kwargs*)
Center the spectrum at the measured peak.

property data_memory_a_condition
Returns the data condition of data memory register A. Starting wavelength, and a sampling point (11, 12, n).

property data_memory_a_size
Returns the number of points sampled in data memory register A.

property data_memory_a_values
Reads the binary data from memory register A.

property data_memory_b_condition
Returns the data condition of data memory register B. Starting wavelength, and a sampling point (11, 12, n).

property data_memory_b_size
Returns the number of points sampled in data memory register B.

property data_memory_b_values
Reads the binary data from memory register B.

property data_memory_select
Memory Data Select.

property dip_search
Dip Search Mode

property ese2
Extended Event Status Enable Register 2

property esr2
Extended Event Status Register 2

property level_lin
Level Linear Scale (/div)

property level_log
Level Log Scale (/div)

property level_opt_attn
Optical Attenuation Status (ON/OFF)

property level_scale
Current Level Scale

property measure_mode
Returns the current Measure Mode the OSA is in.

measure_peak ()
Measure the peak and return the trace marker.

property peak_search
Peak Search Mode

read_memory (slot='A')
Read the scan saved in a memory slot.

property resolution
Resolution (nm)

property resolution_actual
Resolution Actual (ON/OFF)

property resolution_vbw
Video Bandwidth Resolution

property sampling_points
Number of sampling points

single_sweep (kwargs)**
Perform a single sweep and wait for completion.

property trace_marker
Sets the trace marker with a wavelength. Returns the trace wavelength and power.

property trace_marker_center
Trace Marker at Center. Set to 1 or True to initiate command

wait (n=3, delay=1)
Query OPC Command and waits for appropriate response.

wait_for_sweep (n=20, delay=0.5)
Wait for a sweep to stop.

This is performed by checking bit 1 of the ESR2.

property wavelength_center
Center Wavelength of Spectrum Scan in nm.

property wavelength_marker_value
Wavelength Marker Value (wavelength or freq.?)

property wavelength_span
Wavelength Span of Spectrum Scan in nm.

property wavelength_start
Wavelength Start of Spectrum Scan in nm.

property wavelength_stop
Wavelength Stop of Spectrum Scan in nm.

property wavelength_value_in
Wavelength value in Vacuum or Air

property wavelengths

Return a numpy array of the current wavelengths of scans.

7.9 Danfysik

This section contains specific documentation on the Danfysik instruments that are implemented. If you are interested in an instrument not included, please consider *adding the instrument*.

7.9.1 Danfysik Serial Adapter

class `pymessage.instruments.danfysik.DanfysikAdapter` (*port*)

Bases: `pymessage.adapters.serial.SerialAdapter`

Provides a `SerialAdapter` with the specific baudrate and timeout for Danfysik serial communication.

Initiates the adapter to open serial communication over the supplied port.

Parameters `port` – A string representing the serial port

read ()

Overwrites the `SerialAdapter.read` method to automatically raise exceptions if errors are reported by the instrument.

Returns String ASCII response of the instrument

Raises An `Exception` if the Danfysik raises an error

write (*command*)

Overwrites the `SerialAdapter.write` method to automatically append a Unix-style linebreak at the end of the command.

Parameters `command` – SCPI command string to be sent to the instrument

7.9.2 Danfysik 8500 Power Supply

class `pymessage.instruments.danfysik.Danfysik8500` (*port*)

Bases: `pymessage.instruments.instrument.Instrument`

Represents the Danfysik 8500 Electromagnet Current Supply and provides a high-level interface for interacting with the instrument

To allow user access to the Prolific Technology PL2303 Serial port adapter in Linux, create the file: `/etc/udev/rules.d/50-danfysik.rules`, with contents:

```
SUBSYSTEMS=="usb",ATTRS{idVendor}=="067b",ATTRS{idProduct}=="2303",MODE="0666",
↳SYMLINK+="danfysik"
```

Then reload the udev rules with:

```
sudo udevadm control --reload-rules
sudo udevadm trigger
```

The device will be accessible through the port `/dev/danfysik`.

add_ramp_step (*current*)

Adds a current step to the ramp set.

Parameters current – A current in Amps

clear_ramp_set ()

Clears the ramp set.

clear_sequence (*stack*)

Clears the sequence by the stack number.

Parameters stack – A stack number between 0-15

property current

The actual current in Amps. This property can be set through `current_ppm`.

property current_ppm

The current in parts per million. This property can be set.

property current_setpoint

The setpoint for the current, which can deviate from the actual current (`current`) while the supply is in the process of setting the value.

disable ()

Disables the flow of current.

enable ()

Enables the flow of current.

property id

Reads the identification information.

is_current_stable ()

Returns True if the current is within 0.02 A of the setpoint value.

is_enabled ()

Returns True if the current supply is enabled.

is_ready ()

Returns True if the instrument is in the ready state.

is_sequence_running (*stack*)

Returns True if a sequence is running with a given stack number

Parameters stack – A stack number between 0-15

local ()

Sets the instrument in local mode, where the front panel can be used.

property polarity

The polarity of the current supply, being either -1 or 1. This property can be set by supplying one of these values.

ramp_to_current (*current, points, delay_time=1*)

Executes `set_ramp_to_current ()` and starts the ramp.

remote ()

Sets the instrument in remote mode, where the the front panel is disabled.

reset_interlocks ()

Resets the instrument interlocks.

set_ramp_delay (*time*)

Sets the ramp delay time in seconds.

Parameters time – The time delay time in seconds

set_ramp_to_current (*current, points, delay_time=1*)

Sets up a linear ramp from the initial current to a different current, with a number of points, and delay time.

Parameters

- **current** – The final current in Amps
- **points** – The number of linear points to traverse
- **delay_time** – A delay time in seconds

set_sequence (*stack, currents, times, multiplier=999999*)

Sets up an arbitrary ramp profile with a list of currents (Amps) and a list of interval times (seconds) on the specified stack number (0-15)

property slew_rate

The slew rate of the current sweep.

start_ramp ()

Starts the current ramp.

start_sequence (*stack*)

Starts a sequence by the stack number.

Parameters stack – A stack number between 0-15

property status

A list of human-readable strings that contain the instrument status information, based on `status_hex`.

property status_hex

The status in hexadecimal. This value is parsed in `status` into a human-readable list.

stop_ramp ()

Stops the current ramp.

stop_sequence ()

Stops the currently running sequence.

sync_sequence (*stack, delay=0*)

Arms the ramp sequence to be triggered by a hardware input to pin P33 1&2 (10 to 24 V) or a TS command. If a delay is provided, the sequence will start after the delay.

Parameters

- **stack** – A stack number between 0-15
- **delay** – A delay time in seconds

wait_for_current (*has_aborted=<function Danfysik8500.<lambda>>, delay=0.01*)

Blocks the process until the current has stabilized. A provided function `has_aborted` can be supplied, which is checked after each delay time (in seconds) in addition to the stability check. This allows an abort feature to be integrated.

Parameters

- **has_aborted** – A function that returns True if the process should stop waiting
- **delay** – The delay time in seconds between each check for stability

wait_for_ready (*has_aborted=<function Danfysik8500.<lambda>>, delay=0.01*)

Blocks the process until the instrument is ready. A provided function `has_aborted` can be supplied, which is checked after each delay time (in seconds) in addition to the readiness check. This allows an abort feature to be integrated.

Parameters

- **has_aborted** – A function that returns True if the process should stop waiting
- **delay** – The delay time in seconds between each check for readiness

7.10 F.W. Bell

This section contains specific documentation on the F.W. Bell instruments that are implemented. If you are interested in an instrument not included, please consider *adding the instrument*.

7.10.1 F.W. Bell 5080 Handheld Gaussmeter

class `pymeasure.instruments.fwbell.FWBell5080` (*port*)
 Bases: `pymeasure.instruments.instrument.Instrument`

Represents the F.W. Bell 5080 Handheld Gaussmeter and provides a high-level interface for interacting with the instrument

Parameters `port` – The serial port of the instrument

```
meter = FWBell5080('/dev/ttyUSB0')      # Connects over serial port /dev/ttyUSB0
↳ (Linux)

meter.units = 'gauss'                  # Sets the measurement units to Gauss
meter.range = 3e3                       # Sets the range to 3 kG
print(meter.field)                     # Reads and prints a field measurement in
↳ G

fields = meter.fields(100)              # Samples 100 field measurements
print(fields.mean(), fields.std())      # Prints the mean and standard deviation
↳ of the samples
```

ask (*command*)

Overwrites the `Instrument.ask` method to remove the last 2 characters from the output.

auto_range ()

Enables the auto range functionality.

property field

Reads a floating point value of the field in the appropriate units.

fields (*samples=1*)

Returns a numpy array of field samples for a given sample number.

Parameters `samples` – The number of samples to preform

property id

Reads the identification information.

property range

A floating point property that controls the maximum field range in the active units. This can take the values of 300 G, 3 kG, and 30 kG for Gauss, 30 mT, 300 mT, and 3 T for Tesla, and 23.88 kAm, 238.8 kAm, and 2388 kAm for Amp-meter.

read ()

Overwrites the `Instrument.read` method to remove the last 2 characters from the output.

reset ()

Resets the instrument.

property units

A string property that controls the field units, which can take the values: 'gauss', 'gauss ac', 'tesla', 'tesla ac', 'amp-meter', and 'amp-meter ac'. The AC versions configure the instrument to measure AC.

values (*command*)

Overwrites the `Instrument.values` method to remove the lastv2 characters from the output.

7.11 Hewlett Packard

This section contains specific documentation on the Hewlett Packard instruments that are implemented. If you are interested in an instrument not included, please consider *adding the instrument*.

7.11.1 HP 33120A Arbitrary Waveform Generator

class `pymeasure.instruments.hp.HP33120A` (*resourceName*, ***kwargs*)

Bases: `pymeasure.instruments.instrument.Instrument`

Represents the Hewlett Packard 33120A Arbitrary Waveform Generator and provides a high-level interface for interacting with the instrument.

property amplitude

A floating point property that controls the voltage amplitude of the output signal. The default units are in peak-to-peak Volts, but can be controlled by `amplitude_units`. The allowed range depends on the waveform shape and can be queried with `max_amplitude` and `min_amplitude`.

property amplitude_units

A string property that controls the units of the amplitude, which can take the values Vpp, Vrms, dBm, and default.

beep ()

Causes a system beep.

property frequency

A floating point property that controls the frequency of the output in Hz. The allowed range depends on the waveform shape and can be queried with `max_frequency` and `min_frequency`.

property max_amplitude

Reads the maximum amplitude in Volts for the given shape

property max_frequency

Reads the maximum frequency in Hz for the given shape

property max_offset

Reads the maximum offset in Volts for the given shape

property min_amplitude

Reads the minimum amplitude in Volts for the given shape

property min_frequency

Reads the minimum frequency in Hz for the given shape

property min_offset

Reads the minimum offset in Volts for the given shape

property offset

A floating point property that controls the amplitude voltage offset in Volts. The allowed range depends on the waveform shape and can be queried with `max_offset` and `min_offset`.

property shape

A string property that controls the shape of the wave, which can take the values: sinusoid, square, triangle, ramp, noise, dc, and user.

7.12 Keithley

This section contains specific documentation on the Keithley instruments that are implemented. If you are interested in an instrument not included, please consider *adding the instrument*.

7.12.1 Keithley 2000 Multimeter

class `pymeasure.instruments.keithley.Keithley2000` (*adapter*, ***kwargs*)

Bases: `pymeasure.instruments.instrument.Instrument`, `pymeasure.instruments.keithley.buffer.KeithleyBuffer`

Represents the Keithley 2000 Multimeter and provides a high-level interface for interacting with the instrument.

```
meter = Keithley2000("GPIB::1")
meter.measure_voltage()
print(meter.voltage)
```

acquire_reference (*mode=None*)

Sets the active value as the reference for the active mode, or can set another mode by its name.

Parameters *mode* – A valid mode name, or None for the active mode

auto_range (*mode=None*)

Sets the active mode to use auto-range, or can set another mode by its name.

Parameters *mode* – A valid mode name, or None for the active mode

beep (*frequency, duration*)

Sounds a system beep.

Parameters

- **frequency** – A frequency in Hz between 65 Hz and 2 MHz
- **duration** – A time in seconds between 0 and 7.9 seconds

property beep_state

A string property that enables or disables the system status beeper, which can take the values: `:code:'enabled'` and `:code:'disabled'`.

property buffer_data

Returns a numpy array of values from the buffer.

property buffer_points

An integer property that controls the number of buffer points. This does not represent actual points in the buffer, but the configuration value instead.

check_errors ()

Read all errors from the instrument.

config_buffer (*points=64, delay=0*)

Configures the measurement buffer for a number of points, to be taken with a specified delay.

Parameters

- **points** – The number of points in the buffer.
- **delay** – The delay time in seconds.

property current

Reads a DC or AC current measurement in Amps, based on the active mode.

property current_ac_bandwidth

A floating point property that sets the AC current detector bandwidth in Hz, which can take the values 3, 30, and 300 Hz.

property current_ac_digits

An integer property that controls the number of digits in the AC current readings, which can take values from 4 to 7.

property current_ac_nplc

A floating point property that controls the number of power line cycles (NPLC) for the AC current measurements, which sets the integration period and measurement speed. Takes values from 0.01 to 10, where 0.1, 1, and 10 are Fast, Medium, and Slow respectively.

property current_ac_range

A floating point property that controls the AC current range in Amps, which can take values from 0 to 3.1 A. Auto-range is disabled when this property is set.

property current_ac_reference

A floating point property that controls the AC current reference value in Amps, which can take values from -3.1 to 3.1 A.

property current_digits

An integer property that controls the number of digits in the DC current readings, which can take values from 4 to 7.

property current_nplc

A floating point property that controls the number of power line cycles (NPLC) for the DC current measurements, which sets the integration period and measurement speed. Takes values from 0.01 to 10, where 0.1, 1, and 10 are Fast, Medium, and Slow respectively.

property current_range

A floating point property that controls the DC current range in Amps, which can take values from 0 to 3.1 A. Auto-range is disabled when this property is set.

property current_reference

A floating point property that controls the DC current reference value in Amps, which can take values from -3.1 to 3.1 A.

disable_buffer ()

Disables the connection between measurements and the buffer, but does not abort the measurement process.

disable_filter (mode=None)

Disables the averaging filter for the active mode, or can set another mode by its name.

Parameters mode – A valid mode name, or None for the active mode

disable_reference (mode=None)

Disables the reference for the active mode, or can set another mode by its name.

Parameters mode – A valid mode name, or None for the active mode

enable_filter (mode=None, type='repeat', count=1)

Enables the averaging filter for the active mode, or can set another mode by its name.

Parameters

- **mode** – A valid mode name, or None for the active mode
- **type** – The type of averaging filter, either ‘repeat’ or ‘moving’.
- **count** – A number of averages, which can take values from 1 to 100

enable_reference (*mode=None*)

Enables the reference for the active mode, or can set another mode by its name.

Parameters mode – A valid mode name, or None for the active mode

property frequency

Reads a frequency measurement in Hz, based on the active mode.

property frequency_aperture

A floating point property that controls the frequency aperture in seconds, which sets the integration period and measurement speed. Takes values from 0.01 to 1.0 s.

property frequency_digits

An integer property that controls the number of digits in the frequency readings, which can take values from 4 to 7.

property frequency_reference

A floating point property that controls the frequency reference value in Hz, which can take values from 0 to 15 MHz.

property frequency_threshold

A floating point property that controls the voltage signal threshold level in Volts for the frequency measurement, which can take values from 0 to 1010 V.

property id

Requests and returns the identification of the instrument.

is_buffer_full ()

Returns True if the buffer is full of measurements.

local ()

Returns control to the instrument panel, and enables the panel if disabled.

measure_continuity ()

Configures the instrument to perform continuity testing.

measure_current (*max_current=0.01, ac=False*)

Configures the instrument to measure current, based on a maximum current to set the range, and a boolean flag to determine if DC or AC is required.

Parameters

- **max_current** – A current in Volts to set the current range
- **ac** – False for DC current, and True for AC current

measure_diode ()

Configures the instrument to perform diode testing.

measure_frequency ()

Configures the instrument to measure the frequency.

measure_period ()

Configures the instrument to measure the period.

measure_resistance (*max_resistance=10000000.0, wires=2*)

Configures the instrument to measure voltage, based on a maximum voltage to set the range, and a boolean flag to determine if DC or AC is required.

Parameters

- **max_voltage** – A voltage in Volts to set the voltage range
- **ac** – False for DC voltage, and True for AC voltage

measure_temperature ()

Configures the instrument to measure the temperature.

measure_voltage (*max_voltage=1, ac=False*)

Configures the instrument to measure voltage, based on a maximum voltage to set the range, and a boolean flag to determine if DC or AC is required.

Parameters

- **max_voltage** – A voltage in Volts to set the voltage range
- **ac** – False for DC voltage, and True for AC voltage

property mode

A string property that controls the configuration mode for measurements, which can take the values: `'current'` (DC), `'current ac'`, `'voltage'` (DC), `'voltage ac'`, `'resistance'` (2-wire), `'resistance 4W'` (4-wire), `'period'`, `'frequency'`, `'temperature'`, `'diode'`, and `'frequency'`.

property period

Reads a period measurement in seconds, based on the active mode.

property period_aperture

A floating point property that controls the period aperture in seconds, which sets the integration period and measurement speed. Takes values from 0.01 to 1.0 s.

property period_digits

An integer property that controls the number of digits in the period readings, which can take values from 4 to 7.

property period_reference

A floating point property that controls the period reference value in seconds, which can take values from 0 to 1 s.

property period_threshold

A floating point property that controls the voltage signal threshold level in Volts for the period measurement, which can take values from 0 to 1010 V.

remote ()

Places the instrument in the remote state, which is does not need to be explicitly called in general.

remote_lock ()

Disables and locks the front panel controls to prevent changes during remote operations. This is disabled by calling `local()`.

reset ()

Resets the instrument state.

reset_buffer ()

Resets the buffer.

property resistance

Reads a resistance measurement in Ohms for both 2-wire and 4-wire configurations, based on the active mode.

property resistance_4W_digits

An integer property that controls the number of digits in the 4-wire resistance readings, which can take values from 4 to 7.

property resistance_4W_nplc

A floating point property that controls the number of power line cycles (NPLC) for the 4-wire resistance measurements, which sets the integration period and measurement speed. Takes values from 0.01 to 10, where 0.1, 1, and 10 are Fast, Medium, and Slow respectively.

property resistance_4W_range

A floating point property that controls the 4-wire resistance range in Ohms, which can take values from 0 to 120 MOhms. Auto-range is disabled when this property is set.

property resistance_4W_reference

A floating point property that controls the 4-wire resistance reference value in Ohms, which can take values from 0 to 120 MOhms.

property resistance_digits

An integer property that controls the number of digits in the 2-wire resistance readings, which can take values from 4 to 7.

property resistance_nplc

A floating point property that controls the number of power line cycles (NPLC) for the 2-wire resistance measurements, which sets the integration period and measurement speed. Takes values from 0.01 to 10, where 0.1, 1, and 10 are Fast, Medium, and Slow respectively.

property resistance_range

A floating point property that controls the 2-wire resistance range in Ohms, which can take values from 0 to 120 MOhms. Auto-range is disabled when this property is set.

property resistance_reference

A floating point property that controls the 2-wire resistance reference value in Ohms, which can take values from 0 to 120 MOhms.

shutdown ()

Brings the instrument to a safe and stable state

start_buffer ()

Starts the buffer.

stop_buffer ()

Aborts the buffering measurement, by stopping the measurement arming and triggering sequence. If possible, a Selected Device Clear (SDC) is used.

property temperature

Reads a temperature measurement in Celsius, based on the active mode.

property temperature_digits

An integer property that controls the number of digits in the temperature readings, which can take values from 4 to 7.

property temperature_nplc

A floating point property that controls the number of power line cycles (NPLC) for the temperature measurements, which sets the integration period and measurement speed. Takes values from 0.01 to 10, where 0.1, 1, and 10 are Fast, Medium, and Slow respectively.

property temperature_reference

A floating point property that controls the temperature reference value in Celsius, which can take values from -200 to 1372 C.

property trigger_count

An integer property that controls the trigger count, which can take values from 1 to 9,999.

property trigger_delay

A floating point property that controls the trigger delay in seconds, which can take values from 1 to 9,999,999.999 s.

property voltage

Reads a DC or AC voltage measurement in Volts, based on the active mode.

property voltage_ac_bandwidth

A floating point property that sets the AC voltage detector bandwidth in Hz, which can take the values 3, 30, and 300 Hz.

property voltage_ac_digits

An integer property that controls the number of digits in the AC voltage readings, which can take values from 4 to 7.

property voltage_ac_nplc

A floating point property that controls the number of power line cycles (NPLC) for the AC voltage measurements, which sets the integration period and measurement speed. Takes values from 0.01 to 10, where 0.1, 1, and 10 are Fast, Medium, and Slow respectively.

property voltage_ac_range

A floating point property that controls the AC voltage range in Volts, which can take values from 0 to 757.5 V. Auto-range is disabled when this property is set.

property voltage_ac_reference

A floating point property that controls the AC voltage reference value in Volts, which can take values from -757.5 to 757.5 Volts.

property voltage_digits

An integer property that controls the number of digits in the DC voltage readings, which can take values from 4 to 7.

property voltage_nplc

A floating point property that controls the number of power line cycles (NPLC) for the DC voltage measurements, which sets the integration period and measurement speed. Takes values from 0.01 to 10, where 0.1, 1, and 10 are Fast, Medium, and Slow respectively.

property voltage_range

A floating point property that controls the DC voltage range in Volts, which can take values from 0 to 1010 V. Auto-range is disabled when this property is set.

property voltage_reference

A floating point property that controls the DC voltage reference value in Volts, which can take values from -1010 to 1010 V.

wait_for_buffer (*should_stop=<function KeithleyBuffer.<lambda>>, timeout=60, interval=0.1*)

Blocks the program, waiting for a full buffer. This function returns early if the `should_stop` function returns True or the timeout is reached before the buffer is full.

Parameters

- **should_stop** – A function that returns True when this function should return early
- **timeout** – A time in seconds after which this function should return early
- **interval** – A time in seconds for how often to check if the buffer is full

7.12.2 Keithley 2400 SourceMeter

class pymeasure.instruments.keithley.**Keithley2400** (*adapter*, ***kwargs*)

Bases: pymeasure.instruments.instrument.Instrument, pymeasure.instruments.keithley.buffer.KeithleyBuffer

Represents the Keithley 2400 SourceMeter and provides a high-level interface for interacting with the instrument.

```
keithley = Keithley2400("GPIB::1")

keithley.apply_current()           # Sets up to source current
keithley.source_current_range = 10e-3 # Sets the source current range to 10 mA
keithley.compliance_voltage = 10    # Sets the compliance voltage to 10 V
keithley.source_current = 0         # Sets the source current to 0 mA
keithley.enable_source()           # Enables the source output

keithley.measure_voltage()         # Sets up to measure voltage

keithley.ramp_to_current(5e-3)     # Ramps the current to 5 mA
print(keithley.voltage)             # Prints the voltage in Volts

keithley.shutdown()                # Ramps the current to 0 mA and disables_
↔output
```

apply_current (*current_range=None, compliance_voltage=0.1*)

Configures the instrument to apply a source current, and uses an auto range unless a current range is specified. The compliance voltage is also set.

Parameters

- **compliance_voltage** – A float in the correct range for a compliance_voltage
- **current_range** – A current_range value or None

apply_voltage (*voltage_range=None, compliance_current=0.1*)

Configures the instrument to apply a source voltage, and uses an auto range unless a voltage range is specified. The compliance current is also set.

Parameters

- **compliance_current** – A float in the correct range for a compliance_current
- **voltage_range** – A voltage_range value or None

auto_range_source ()

Configures the source to use an automatic range.

beep (*frequency, duration*)

Sounds a system beep.

Parameters

- **frequency** – A frequency in Hz between 65 Hz and 2 MHz
- **duration** – A time in seconds between 0 and 7.9 seconds

property buffer_data

Returns a numpy array of values from the buffer.

property buffer_points

An integer property that controls the number of buffer points. This does not represent actual points in the buffer, but the configuration value instead.

check_errors ()

Logs any system errors reported by the instrument.

property compliance_current

A floating point property that controls the compliance current in Amps.

property compliance_voltage

A floating point property that controls the compliance voltage in Volts.

config_buffer (points=64, delay=0)

Configures the measurement buffer for a number of points, to be taken with a specified delay.

Parameters

- **points** – The number of points in the buffer.
- **delay** – The delay time in seconds.

property current

Reads the current in Amps, if configured for this reading.

property current_nplc

A floating point property that controls the number of power line cycles (NPLC) for the DC current measurements, which sets the integration period and measurement speed. Takes values from 0.01 to 10, where 0.1, 1, and 10 are Fast, Medium, and Slow respectively.

property current_range

A floating point property that controls the measurement current range in Amps, which can take values between -1.05 and +1.05 A. Auto-range is disabled when this property is set.

disable_buffer ()

Disables the connection between measurements and the buffer, but does not abort the measurement process.

disable_output_trigger ()

Disables the output trigger for the Trigger layer

disable_source ()

Disables the source of current or voltage depending on the configuration of the instrument.

enable_source ()

Enables the source of current or voltage depending on the configuration of the instrument.

property error

Returns a tuple of an error code and message from a single error.

property id

Requests and returns the identification of the instrument.

is_buffer_full ()

Returns True if the buffer is full of measurements.

property max_current

Returns the maximum current from the buffer

property max_resistance

Returns the maximum resistance from the buffer

property max_voltage

Returns the maximum voltage from the buffer

property maximums

Returns the calculated maximums for voltage, current, and resistance from the buffer data as a list.

property mean_current

Returns the mean current from the buffer

property mean_resistance

Returns the mean resistance from the buffer

property mean_voltage

Returns the mean voltage from the buffer

property means

Reads the calculated means (averages) for voltage, current, and resistance from the buffer data as a list.

measure_current (*nplc=1, current=0.000105, auto_range=True*)

Configures the measurement of current.

Parameters

- **nplc** – Number of power line cycles (NPLC) from 0.01 to 10
- **current** – Upper limit of current in Amps, from -1.05 A to 1.05 A
- **auto_range** – Enables auto_range if True, else uses the set current

measure_resistance (*nplc=1, resistance=210000.0, auto_range=True*)

Configures the measurement of resistance.

Parameters

- **nplc** – Number of power line cycles (NPLC) from 0.01 to 10
- **resistance** – Upper limit of resistance in Ohms, from -210 MOhms to 210 MOhms
- **auto_range** – Enables auto_range if True, else uses the set resistance

measure_voltage (*nplc=1, voltage=21.0, auto_range=True*)

Configures the measurement of voltage.

Parameters

- **nplc** – Number of power line cycles (NPLC) from 0.01 to 10
- **voltage** – Upper limit of voltage in Volts, from -210 V to 210 V
- **auto_range** – Enables auto_range if True, else uses the set voltage

property min_current

Returns the minimum current from the buffer

property min_resistance

Returns the minimum resistance from the buffer

property min_voltage

Returns the minimum voltage from the buffer

property minimums

Returns the calculated minimums for voltage, current, and resistance from the buffer data as a list.

output_trigger_on_external (*line=1, after='DEL'*)

Configures the output trigger on the specified trigger link line number, with the option of supplying the part of the measurement after which the trigger should be generated (default to delay, which is right before the measurement)

Parameters

- **line** – A trigger line from 1 to 4
- **after** – An event string that determines when to trigger

ramp_to_current (*target_current, steps=30, pause=0.02*)

Ramps to a target current from the set current value over a certain number of linear steps, each separated by a pause duration.

Parameters

- **target_current** – A current in Amps
- **steps** – An integer number of steps
- **pause** – A pause duration in seconds to wait between steps

ramp_to_voltage (*target_voltage, steps=30, pause=0.02*)

Ramps to a target voltage from the set voltage value over a certain number of linear steps, each separated by a pause duration.

Parameters

- **target_voltage** – A voltage in Amps
- **steps** – An integer number of steps
- **pause** – A pause duration in seconds to wait between steps

reset ()

Resets the instrument and clears the queue.

reset_buffer ()

Resets the buffer.

property resistance

Reads the resistance in Ohms, if configured for this reading.

property resistance_nplc

A floating point property that controls the number of power line cycles (NPLC) for the 2-wire resistance measurements, which sets the integration period and measurement speed. Takes values from 0.01 to 10, where 0.1, 1, and 10 are Fast, Medium, and Slow respectively.

property resistance_range

A floating point property that controls the resistance range in Ohms, which can take values from 0 to 210 MOhms. Auto-range is disabled when this property is set.

sample_continuously ()

Causes the instrument to continuously read samples and turns off any buffer or output triggering

set_timed_arm (*interval*)

Sets up the measurement to be taken with the internal trigger at a variable sampling rate defined by the interval in seconds between sampling points

set_trigger_counts (*arm, trigger*)

Sets the number of counts for both the sweeps (arm) and the points in those sweeps (trigger), where the total number of points can not exceed 2500

shutdown ()

Ensures that the current or voltage is turned to zero and disables the output.

property source_current

A floating point property that controls the source current in Amps.

property source_current_range

A floating point property that controls the source current range in Amps, which can take values between -1.05 and +1.05 A. Auto-range is disabled when this property is set.

property source_enabled

A boolean property that controls whether the source is enabled, takes values True or False. The convenience methods `enable_source()` and `disable_source()` can also be used.

property source_mode

A string property that controls the source mode, which can take the values 'current' or 'voltage'. The convenience methods `apply_current()` and `apply_voltage()` can also be used.

property source_voltage

A floating point property that controls the source voltage in Volts.

property source_voltage_range

A floating point property that controls the source voltage range in Volts, which can take values from -210 to 210 V. Auto-range is disabled when this property is set.

property standard_devs

Returns the calculated standard deviations for voltage, current, and resistance from the buffer data as a list.

start_buffer()

Starts the buffer.

property std_current

Returns the current standard deviation from the buffer

property std_resistance

Returns the resistance standard deviation from the buffer

property std_voltage

Returns the voltage standard deviation from the buffer

stop_buffer()

Aborts the buffering measurement, by stopping the measurement arming and triggering sequence. If possible, a Selected Device Clear (SDC) is used.

triad(*base_frequency*, *duration*)

Sounds a musical triad using the system beep.

Parameters

- **base_frequency** – A frequency in Hz between 65 Hz and 1.3 MHz
- **duration** – A time in seconds between 0 and 7.9 seconds

trigger()

Executes a bus trigger, which can be used when `trigger_on_bus()` is configured.

property trigger_count

An integer property that controls the trigger count, which can take values from 1 to 9,999.

property trigger_delay

A floating point property that controls the trigger delay in seconds, which can take values from 0 to 999.9999 s.

trigger_immediately()

Configures measurements to be taken with the internal trigger at the maximum sampling rate.

trigger_on_bus()

Configures the trigger to detect events based on the bus trigger, which can be activated by GET or *TRG.

trigger_on_external(*line=1*)

Configures the measurement trigger to be taken from a specific line of an external trigger

Parameters **line** – A trigger line from 1 to 4

use_front_terminals ()

Enables the front terminals for measurement, and disables the rear terminals.

use_rear_terminals ()

Enables the rear terminals for measurement, and disables the front terminals.

property voltage

Reads the voltage in Volts, if configured for this reading.

property voltage_nplc

A floating point property that controls the number of power line cycles (NPLC) for the DC voltage measurements, which sets the integration period and measurement speed. Takes values from 0.01 to 10, where 0.1, 1, and 10 are Fast, Medium, and Slow respectively.

property voltage_range

A floating point property that controls the measurement voltage range in Volts, which can take values from -210 to 210 V. Auto-range is disabled when this property is set.

wait_for_buffer (*should_stop*=<function KeithleyBuffer.<lambda>>, *timeout*=60, *interval*=0.1)

Blocks the program, waiting for a full buffer. This function returns early if the `should_stop` function returns True or the timeout is reached before the buffer is full.

Parameters

- **should_stop** – A function that returns True when this function should return early
- **timeout** – A time in seconds after which this function should return early
- **interval** – A time in seconds for how often to check if the buffer is full

property wires

An integer property that controls the number of wires in use for resistance measurements, which can take the value of 2 or 4.

7.13 Lake Shore Cryogenics

This section contains specific documentation on the Lake Shore Cryogenics instruments that are implemented. If you are interested in an instrument not included, please consider *adding the instrument*.

7.13.1 Lake Shore Adapters

class `pymeasure.instruments.lakeshore.LakeShoreUSBAdapter` (*port*)

Bases: `pymeasure.adapters.serial.SerialAdapter`

Provides a `SerialAdapter` with the specific baudrate, timeout, parity, and byte size for LakeShore USB communication.

Initiates the adapter to open serial communication over the supplied port.

Parameters `port` – A string representing the serial port

ask (*command*)

Writes the command to the instrument and returns the resulting ASCII response

Parameters `command` – SCPI command string to be sent to the instrument

Returns String ASCII response of the instrument

binary_values (*command*, *header_bytes*=0, *dtype*=<class 'numpy.float32'>)

Returns a numpy array from a query for binary data

Parameters

- **command** – SCPI command to be sent to the instrument
- **header_bytes** – Integer number of bytes to ignore in header
- **dtype** – The NumPy data type to format the values with

Returns NumPy array of values

read()

Reads until the buffer is empty and returns the resulting ASCII response

Returns String ASCII response of the instrument.

values (*command*, *separator*=' ', *cast*=<class 'float'>)

Writes a command to the instrument and returns a list of formatted values from the result

Parameters

- **command** – SCPI command to be sent to the instrument
- **separator** – A separator character to split the string into a list
- **cast** – A type to cast the result

Returns A list of the desired type, or strings where the casting fails

write (*command*)

Overwrites the `SerialAdapter.write` method to automatically append a Unix-style linebreak at the end of the command.

Parameters **command** – SCPI command string to be sent to the instrument

7.13.2 Lake Shore 331 Temperature Controller

class `pymeasure.instruments.lakeshore.LakeShore331` (*adapter*, ***kwargs*)

Bases: `pymeasure.instruments.instrument.Instrument`

Represents the Lake Shore 331 Temperature Controller and provides a high-level interface for interacting with the instrument.

```
controller = LakeShore331("GPIB::1")

print(controller.setpoint_1)      # Print the current setpoint for loop 1
controller.setpoint_1 = 50       # Change the setpoint to 50 K
controller.heater_range = 'low'  # Change the heater range to Low
controller.wait_for_temperature() # Wait for the temperature to stabilize
print(controller.temperature_A)  # Print the temperature at sensor A
```

disable_heater()

Turns the `heater_range` to off to disable the heater.

property heater_range

A string property that controls the heater range, which can take the values: off, low, medium, and high. These values correlate to 0, 0.5, 5 and 50 W respectively.

property setpoint_1

A floating point property that controls the setpoint temperature in Kelvin for Loop 1.

property setpoint_2

A floating point property that controls the setpoint temperature in Kelvin for Loop 2.

property temperature_A

Reads the temperature of the sensor A in Kelvin.

property temperature_B

Reads the temperature of the sensor B in Kelvin.

wait_for_temperature (*accuracy=0.1, interval=0.1, sensor='A', setpoint=1, timeout=360, should_stop=<function LakeShore331.<lambda>>*)

Blocks the program, waiting for the temperature to reach the setpoint within the accuracy (%), checking this each interval time in seconds.

Parameters

- **accuracy** – An acceptable percentage deviation between the setpoint and temperature
- **interval** – A time in seconds that controls the refresh rate
- **sensor** – The desired sensor to read, either A or B
- **setpoint** – The desired setpoint loop to read, either 1 or 2
- **timeout** – A timeout in seconds after which an exception is raised
- **should_stop** – A function that returns True if waiting should stop, by default this always returns False

7.13.3 Lake Shore 425 Gaussmeter

class pymeasure.instruments.lakeshore.LakeShore425 (*port*)

Bases: pymeasure.instruments.instrument.Instrument

Represents the LakeShore 425 Gaussmeter and provides a high-level interface for interacting with the instrument

To allow user access to the LakeShore 425 Gaussmeter in Linux, create the file: `/etc/udev/rules.d/52-lakeshore425.rules`, with contents:

```
SUBSYSTEMS=="usb",ATTRS{idVendor}=="1fb9",ATTRS{idProduct}=="0401",MODE="0666",
↳SYMLINK+="lakeshore425"
```

Then reload the udev rules with:

```
sudo udevadm control --reload-rules
sudo udevadm trigger
```

The device will be accessible through `/dev/lakeshore425`.

ac_mode (*wideband=True*)

Sets up a measurement of an oscillating (AC) field

auto_range ()

Sets the field range to automatically adjust

dc_mode (*wideband=True*)

Sets up a steady-state (DC) measurement of the field

property field

Returns the field in the current units

measure (*points, has_aborted=<function LakeShore425.<lambda>>, delay=0.001*)

Returns the mean and standard deviation of a given number of points while blocking

property range

A floating point property that controls the field range in units of Gauss, which can take the values 35, 350, 3500, and 35,000 G.

property unit

A string property that controls the units of the instrument, which can take the values of G, T, Oe, or A/m.

zero_probe ()

Initiates the zero field sequence to calibrate the probe

7.14 Newport

This section contains specific documentation on the Newport instruments that are implemented. If you are interested in an instrument not included, please consider *adding the instrument*.

7.14.1 ESP 300 Motion Controller

class `pymessage.instruments.newport.ESP300 (resourceName, **kwargs)`

Bases: `pymessage.instruments.instrument.Instrument`

Represents the Newport ESP 300 Motion Controller and provides a high-level for interacting with the instrument.

By default this instrument is constructed with x, y, and phi attributes that represent axes 1, 2, and 3. Custom implementations can overwrite this depending on the available axes. Axes are controlled through an *Axis* class.

property axes

A list of the *Axis* objects that are present.

clear_errors ()

Clears the error messages by checking until a 0 code is received.

disable ()

Disables all of the axes associated with this controller.

enable ()

Enables all of the axes associated with this controller.

property error

Reads an error code from the motion controller.

property errors

Returns a list of error Exceptions that can be later raised, or used to diagnose the situation.

shutdown ()

Shuts down the controller by disabling all of the axes.

class `pymessage.instruments.newport.esp300.Axis (axis, controller)`

Bases: `object`

Represents an axis of the Newport ESP300 Motor Controller, which can have independent parameters from the other axes.

define_position (position)

Overwrites the value of the current position with the given value.

disable ()

Disables motion for the axis.

enable ()

Enables motion for the axis.

property enabled

Returns a boolean value that is True if the motion for this axis is enabled.

home (type=1)

Drives the axis to the home position, which may be the negative hardware limit for some actuators (e.g. LTA-HS). type can take integer values from 0 to 6.

property left_limit

A floating point property that controls the left software limit of the axis.

property motion_done

Returns a boolean that is True if the motion is finished.

property position

A floating point property that controls the position of the axis. The units are defined based on the actuator. Use the `wait_for_stop()` method to ensure the position is stable.

property right_limit

A floating point property that controls the right software limit of the axis.

property units

A string property that controls the displacement units of the axis, which can take values of: encoder count, motor step, millimeter, micrometer, inches, milli-inches, micro-inches, degree, gradient, radian, milliradian, and microradian.

wait_for_stop (delay=0, interval=0.05)

Blocks the program until the motion is completed. A further delay can be specified in seconds.

zero ()

Resets the axis position to be zero at the current position.

class `pymeaure.instruments.newport.esp300.AxisError (code)`

Bases: Exception

Raised when a particular axis causes an error for the Newport ESP300.

class `pymeaure.instruments.newport.esp300.GeneralError (code)`

Bases: Exception

Raised when the Newport ESP300 has a general error.

7.15 Oxford Instruments

This section contains specific documentation on the Oxford Instruments instruments that are implemented. If you are interested in an instrument not included, please consider *adding the instrument*.

7.15.1 Oxford Instrument Intelligent Temperature Controller 503

class pymeasure.instruments.oxfordinstruments.**ITC503** (*resourceName*,
clear_buffer=True, ***kwargs*)

Bases: pymeasure.instruments.instrument.Instrument

Represents the Oxford Intelligent Temperature Controller 503.

```

itc = ITC503("GPIB::24")           # Default channel for the ITC503

itc.control_mode = "RU"           # Set the control mode to remote
itc.heater_gas_mode = "AUTO"     # Turn on auto heater and flow
itc.auto_pid = True               # Turn on auto-pid

print(itc.temperature_setpoint)   # Print the current set-point
itc.temperature_setpoint = 300    # Change the set-point to 300 K
itc.wait_for_temperature()        # Wait for the temperature to stabilize
print(itc.temperature_1)          # Print the temperature at sensor 1

```

property auto_pid

A boolean property that sets the Auto-PID mode on (True) or off (False).

property control_mode

A string property that sets the ITC in LOCAL or REMOTE and LOCKES, or UNLOCKES, the LOC/REM button. Allowed values are: LL: LOCAL & LOCKED RL: REMOTE & LOCKED LU: LOCAL & UNLOCKED RU: REMOTE & UNLOCKED.

property heater_gas_mode

A string property that sets the heater and gas flow control to AUTO or MANUAL. Allowed values are: MANUAL: HEATER MANUAL, GAS MANUAL AM: HEATER AUTO, GAS MANUAL MA: HEATER MANUAL, GAS AUTO AUTO: HEATER AUTO, GAS AUTO.

program_sweep (*temperatures*, *sweep_time*, *hold_time*, *steps=None*)

Program a temperature sweep in the controller. Stops any running sweep. After programming the sweep, it can be started using `OxfordITC503.sweep_status = 1`.

Parameters

- **temperatures** – An array containing the temperatures for the sweep
- **sweep_time** – The time (or an array of times) to sweep to a set-point in minutes (between 0 and 1339.9).
- **hold_time** – The time (or an array of times) to hold at a set-point in minutes (between 0 and 1339.9).
- **steps** – The number of steps in the sweep, if given, the temperatures, sweep_time and hold_time will be interpolated into (approximately) equal segments

property sweep_status

An integer property that sets the sweep status. Values are: 0: Sweep not running 1: Start sweep / sweeping to first set-point 2P - 1: Sweeping to set-point P 2P: Holding at set-point P.

property sweep_table

A property that sets values in the sweep table. Relies on the xpointer and ypointer to point at the location in the table that is to be set.

property temperature_1

Reads the temperature of the sensor 1 in Kelvin.

property temperature_2

Reads the temperature of the sensor 3 in Kelvin.

property temperature_error

Reads the difference between the set-point and the measured temperature in Kelvin. Positive when set-point is larger than measured.

property temperature_setpoint

A floating point property that controls the temperature set-point of the ITC in kelvin.

wait_for_temperature (*error=0.01, timeout=3600, check_interval=0.5, stability_interval=10, thermalize_interval=300, should_stop=<function ITC503.<lambda>>*)

Wait for the ITC to reach the set-point temperature.

Parameters

- **error** – The maximum error in Kelvin under which the temperature is considered at set-point
- **timeout** – The maximum time the waiting is allowed to take. If timeout is exceeded, a `TimeoutError` is raised. If timeout is set to zero, no timeout will be used.
- **check_interval** – The time between temperature queries to the ITC.
- **stability_interval** – The time over which the `temperature_error` is to be below error to be considered stable.
- **thermalize_interval** – The time to wait after stabilizing for the system to thermalize.
- **should_stop** – Optional function (returning a bool) to allow the waiting to be stopped before its end.

property xpointer

An integer property to set pointers into tables for loading and examining values in the table. For programming the sweep table values from 1 to 16 are allowed, corresponding to the maximum number of steps.

property ypointer

An integer property to set pointers into tables for loading and examining values in the table. For programming the sweep table the allowed values are: 1: Setpoint temperature, 2: Sweep-time to set-point, 3: Hold-time at set-point.

7.16 Parker

This section contains specific documentation on the Parker instruments that are implemented. If you are interested in an instrument not included, please consider [adding the instrument](#).

7.16.1 Parker GV6 Servo Motor Controller

class `pymeaure.instruments.parker.ParkerGV6` (*port*)

Bases: `pymeaure.instruments.instrument.Instrument`

Represents the Parker Gemini GV6 Servo Motor Controller and provides a high-level interface for interacting with the instrument

property angle

Returns the angle in degrees based on the position and whether relative or absolute positioning is enabled, returning `None` on error

property angle_error

Returns the angle error in degrees based on the position error, or returns `None` on error

disable ()
Disables the motor from moving

echo (*enable=False*)
Enables (True) or disables (False) the echoing of all commands that are sent to the instrument

enable ()
Enables the motor to move

is_moving ()
Returns True if the motor is currently moving

kill ()
Stops the motor

move ()
Initiates the motor to move to the setpoint

property position
Returns an integer number of counts that correspond to the angular position where 1 revolution equals 4000 counts

property position_error
Returns the error in the number of counts that corresponds to the error in the angular position where 1 revolution equals 4000 counts

read ()
Overwrites the Instrument.read command to provide the correct functionality

reset ()
Resets the motor controller while blocking and (CAUTION) resets the absolute position value of the motor

set_defaults ()
Sets up the default values for the motor, which is run upon construction

set_hardware_limits (*positive=True, negative=True*)
Enables (True) or disables (False) the hardware limits for the motor

set_software_limits (*positive, negative*)
Sets the software limits for motion based on the count unit where 4000 counts is 1 revolution

property status
Returns a list of the motor status in readable format

stop ()
Stops the motor during movement

use_absolute_position ()
Sets the motor to accept setpoints from an absolute zero position

use_relative_position ()
Sets the motor to accept setpoints that are relative to the last position

write (*command*)
Overwrites the Instrument.write command to provide the correct line break syntax

7.17 Signal Recovery

This section contains specific documentation on the Signal Recovery instruments that are implemented. If you are interested in an instrument not included, please consider *adding the instrument*.

7.17.1 DSP 7265 Lock-in Amplifier

class `pymeasure.instruments.signalrecovery.DSP7265` (*resourceName*, ***kwargs*)

Bases: `pymeasure.instruments.instrument.Instrument`

This is the class for the DSP 7265 lockin amplifier

property `adc1`

Reads the input value of ADC1 in Volts

property `adc2`

Reads the input value of ADC2 in Volts

property `dac1`

A floating point property that represents the output value on DAC1 in Volts. This property can be set.

property `dac2`

A floating point property that represents the output value on DAC2 in Volts. This property can be set.

property `dac3`

A floating point property that represents the output value on DAC3 in Volts. This property can be set.

property `dac4`

A floating point property that represents the output value on DAC4 in Volts. This property can be set.

property `frequency`

A floating point property that represents the lock-in frequency in Hz. This property can be set.

property `harmonic`

An integer property that represents the reference harmonic mode control, taking values from 1 to 65535. This property can be set.

property `id`

Reads the instrument identification

property `mag`

Reads the magnitude in Volts

property `phase`

A floating point property that represents the reference harmonic phase in degrees. This property can be set.

property `reference`

Controls the oscillator reference. Can be “internal”, “external rear” or “external front”

property `sensitivity`

A floating point property that controls the sensitivity range in Volts, which can take discrete values from 2 nV to 1 V. This property can be set.

setDifferentialMode (*lineFiltering=True*)

Sets lockin to differential mode, measuring A-B

shutdown ()

Brings the instrument to a safe and stable state

property slope

A integer property that controls the filter slope in dB/octave, which can take the values 6, 12, 18, or 24 dB/octave. This property can be set.

property time_constant

A floating point property that controls the time constant in seconds, which takes values from 10 microseconds to 50,000 seconds. This property can be set.

values (*command*)

Rewrite the method because of extra character in return string.

property voltage

A floating point property that represents the voltage in Volts. This property can be set.

property x

Reads the X value in Volts

property xy

Reads both the X and Y values in Volts

property y

Reads the Y value in Volts

7.18 Stanford Research Systems

This section contains specific documentation on the Stanford Research Systems (SRS) instruments that are implemented. If you are interested in an instrument not included, please consider [adding the instrument](#).

7.18.1 SR830 Lock-in Amplifier

class `pymeasure.instruments.srs.SR830` (*resourceName*, ***kwargs*)

Bases: `pymeasure.instruments.instrument.Instrument`

property adc1

Reads the Aux input 1 value in Volts with 1/3 mV resolution.

property adc2

Reads the Aux input 2 value in Volts with 1/3 mV resolution.

property adc3

Reads the Aux input 3 value in Volts with 1/3 mV resolution.

property adc4

Reads the Aux input 4 value in Volts with 1/3 mV resolution.

auto_offset (*channel*)

Offsets the channel (X, Y, or R) to zero

property aux_in_1

Reads the Aux input 1 value in Volts with 1/3 mV resolution.

property aux_in_2

Reads the Aux input 2 value in Volts with 1/3 mV resolution.

property aux_in_3

Reads the Aux input 3 value in Volts with 1/3 mV resolution.

property aux_in_4

Reads the Aux input 4 value in Volts with 1/3 mV resolution.

property aux_out_1

A floating point property that controls the output of Aux output 1 in Volts, taking values between -10.5 V and +10.5 V. This property can be set.

property aux_out_2

A floating point property that controls the output of Aux output 2 in Volts, taking values between -10.5 V and +10.5 V. This property can be set.

property aux_out_3

A floating point property that controls the output of Aux output 3 in Volts, taking values between -10.5 V and +10.5 V. This property can be set.

property aux_out_4

A floating point property that controls the output of Aux output 4 in Volts, taking values between -10.5 V and +10.5 V. This property can be set.

property channel1

A string property that represents the type of Channel 1, taking the values X, R, X Noise, Aux In 1, or Aux In 2. This property can be set.

property channel2

A string property that represents the type of Channel 2, taking the values Y, Theta, Y Noise, Aux In 3, or Aux In 4. This property can be set.

property dac1

A floating point property that controls the output of Aux output 1 in Volts, taking values between -10.5 V and +10.5 V. This property can be set.

property dac2

A floating point property that controls the output of Aux output 2 in Volts, taking values between -10.5 V and +10.5 V. This property can be set.

property dac3

A floating point property that controls the output of Aux output 3 in Volts, taking values between -10.5 V and +10.5 V. This property can be set.

property dac4

A floating point property that controls the output of Aux output 4 in Volts, taking values between -10.5 V and +10.5 V. This property can be set.

property filter_slope

An integer property that controls the filter slope, which can take on the values 6, 12, 18, and 24 dB/octave. Values are truncated to the next highest level if they are not exact.

property frequency

A floating point property that represents the lock-in frequency in Hz. This property can be set.

get_buffer (*channel=1, start=0, end=None*)

Acquires the 32 bit floating point data through binary transfer

get_scaling (*channel*)

Returns the offset present and the expansion term that are used to scale the channel in question

property harmonic

An integer property that controls the harmonic that is measured. Allowed values are 1 to 19999. Can be set.

property input_config

An string property that controls the input configuration. Allowed values are: ['A', 'A - B', 'I (1 MOhm)', 'I (100 MOhm)']

property input_coupling

An string property that controls the input coupling. Allowed values are: ['AC', 'DC']

property input_grounding

An string property that controls the input shield grounding. Allowed values are: ['Float', 'Ground']

property input_notch_config

An string property that controls the input line notch filter status. Allowed values are: ['None', 'Line', '2 x Line', 'Both']

is_out_of_range()

Returns True if the magnitude is out of range

property magnitude

Reads the magnitude in Volts.

output_conversion(channel)

Returns a function that can be used to determine the signal from the channel output (X, Y, or R)

property phase

A floating point property that represents the lock-in phase in degrees. This property can be set.

quick_range()

While the magnitude is out of range, increase the sensitivity by one setting

property reference_source

An string property that controls the reference source. Allowed values are: ['External', 'Internal']

property sample_frequency

Gets the sample frequency in Hz

property sensitivity

A floating point property that controls the sensitivity in Volts, which can take discrete values from 2 nV to 1 V. Values are truncated to the next highest level if they are not exact.

set_scaling(channel, percent, expand=0)

Sets the offset of a channel (X=1, Y=2, R=3) to a certain percent (-105% to 105%) of the signal, with an optional expansion term (0, 10=1, 100=2)

property sine_voltage

A floating point property that represents the reference sine-wave voltage in Volts. This property can be set.

property theta

Reads the theta value in degrees.

property time_constant

A floating point property that controls the time constant in seconds, which can take discrete values from 10 microseconds to 30,000 seconds. Values are truncated to the next highest level if they are not exact.

wait_for_buffer(count, has_aborted=<function SR830.<lambda>>, timeout=60, timestep=0.01)

Wait for the buffer to fill a certain count

property x

Reads the X value in Volts.

property y

Reads the Y value in Volts.

7.19 Tektronix

This section contains specific documentation on the Tektronix instruments that are implemented. If you are interested in an instrument not included, please consider *adding the instrument*.

7.19.1 TDS2000 Oscilloscope

class `pymeasure.instruments.tektronix.TDS2000` (*resourceName*, ***kwargs*)

Bases: `pymeasure.instruments.instrument.Instrument`

Represents the Tektronix TDS 2000 Oscilloscope and provides a high-level for interacting with the instrument

7.19.2 AFG3152C Arbitrary function generator

class `pymeasure.instruments.tektronix.AFG3152C` (*adapter*, ***kwargs*)

Bases: `pymeasure.instruments.instrument.Instrument`

Represents the Tektronix AFG 3000 series (one or two channels) arbitrary function generator and provides a high-level for interacting with the instrument.

```
afg=AFG3152C("GPIB::1") # AFG on GPIB 1
afg.reset() # Reset to default
afg.ch1.shape='sinusoidal' # Sinusoidal shape
afg.ch1.unit='VPP' # Sets CH1 unit to VPP
afg.ch1.amp_vpp=1 # Sets the CH1 level to 1 VPP
afg.ch1.frequency=1e3 # Sets the CH1 frequency to 1KHz
afg.ch1.enable() # Enables the output from CH1
```

7.20 Thorlabs

This section contains specific documentation on the Thorlabs instruments that are implemented. If you are interested in an instrument not included, please consider *adding the instrument*.

7.20.1 Thorlabs PM100USB Powermeter

class `pymeasure.instruments.thorlabs.ThorlabsPM100USB` (*adapter*, ***kwargs*)

Bases: `pymeasure.instruments.instrument.Instrument`

Represents Thorlabs PM100USB powermeter

measure_power (*wavelength*)

Set wavelength in nm and get power in W If wavelength is out of range it will be set to range limit

property power

Power, in Watts

sensor ()

Get sensor info

property wavelength

Wavelength in nm; not set outside of range

property wavelength_max

Get maximum wavelength, in nm

property wavelength_min

Get minimum wavelength, in nm

7.21 Yokogawa

This section contains specific documentation on the Yokogawa instruments that are implemented. If you are interested in an instrument not included, please consider *adding the instrument*.

7.21.1 Yokogawa 7651 Programmable Supply

class `pymessage.instruments.yokogawa.Yokogawa7651` (*adapter*, ***kwargs*)

Bases: `pymessage.instruments.instrument.Instrument`

Represents the Yokogawa 7651 Programmable DC Source and provides a high-level for interacting with the instrument.

```
yoko = Yokogawa7651("GPIB::1")

yoko.apply_current()           # Sets up to source current
yoko.source_current_range = 10e-3 # Sets the current range to 10 mA
yoko.compliance_voltage = 10    # Sets the compliance voltage to 10 V
yoko.source_current = 0         # Sets the source current to 0 mA

yoko.enable_source()          # Enables the current output
yoko.ramp_to_current(5e-3)    # Ramps the current to 5 mA

yoko.shutdown()              # Ramps the current to 0 mA and disables_
↪output
```

apply_current (*max_current=0.001*, *compliance_voltage=1*)

Configures the instrument to apply a source current, which can take optional parameters that defer to the `source_current_range` and `compliance_voltage` properties.

apply_voltage (*max_voltage=1*, *compliance_current=0.01*)

Configures the instrument to apply a source voltage, which can take optional parameters that defer to the `source_voltage_range` and `compliance_current` properties.

property compliance_current

A floating point property that sets the compliance current in Amps, which can take values from 5 to 120 mA.

property compliance_voltage

A floating point property that sets the compliance voltage in Volts, which can take values between 1 and 30 V.

disable_source ()

Disables the source of current or voltage depending on the configuration of the instrument.

enable_source ()

Enables the source of current or voltage depending on the configuration of the instrument.

property id

Returns the identification of the instrument

ramp_to_current (*current*, *steps=25*, *duration=0.5*)

Ramps the current to a value in Amps by traversing a linear spacing of current steps over a duration, defined in seconds.

Parameters

- **steps** – A number of linear steps to traverse

- **duration** – A time in seconds over which to ramp

ramp_to_voltage (*voltage*, *steps=25*, *duration=0.5*)

Ramps the voltage to a value in Volts by traversing a linear spacing of voltage steps over a duration, defined in seconds.

Parameters

- **steps** – A number of linear steps to traverse
- **duration** – A time in seconds over which to ramp

shutdown ()

Shuts down the instrument, and ramps the current or voltage to zero before disabling the source.

property source_current

A floating point property that controls the source current in Amps, if that mode is active.

property source_current_range

A floating point property that sets the current voltage range in Amps, which can take values: 1 mA, 10 mA, and 100 mA. Currents are truncated to an appropriate value if needed.

property source_enabled

Reads a boolean value that is True if the source is enabled, determined by checking if the 5th bit of the OC flag is a binary 1.

property source_mode

A string property that controls the source mode, which can take the values 'current' or 'voltage'. The convenience methods *apply_current()* and *apply_voltage()* can also be used.

property source_voltage

A floating point property that controls the source voltage in Volts, if that mode is active.

property source_voltage_range

A floating point property that sets the source voltage range in Volts, which can take values: 10 mV, 100 mV, 1 V, 10 V, and 30 V. Voltages are truncated to an appropriate value if needed.

CONTRIBUTING

Contributions to the instrument repository and the main code base are highly encouraged. This section outlines the basic work-flow for new contributors.

8.1 Using the development version

New features are added to the development version of PyMeasure, hosted on [GitHub](#). We use [Git version control](#) to track and manage changes to the source code. On Windows, we recommend using [GitHub Desktop](#). Make sure you have an appropriate version of Git (or GitHub Desktop) installed and that you have a GitHub account.

In order to add your feature, you need to first [fork](#) PyMeasure. This will create a copy of the repository under your GitHub account.

The instructions below assume that you have set up Anaconda, as described in the [Quick Start guide](#) and describe the terminal commands necessary. If you are using GitHub Desktop, take a look through [their documentation](#) to understand the corresponding steps.

Clone your fork of PyMeasure `your-github-username/pymeaure`. In the following terminal commands replace your desired path and GitHub username.

```
cd /path/for/code
git clone https://github.com/your-github-username/pymeaure.git
```

If you had already installed PyMeasure using `pip`, make sure to uninstall it before continuing.

```
pip uninstall pymeaure
```

Install PyMeasure in the editable mode.

```
cd /path/for/code/pymeaure
pip install -e .
```

This will allow you to edit the files of PyMeasure and see the changes reflected. Make sure to reset your notebook kernel or Python console when doing so. Now you have your own copy of the development version of PyMeasure installed!

8.2 Working on a new feature

We use branches in Git to allow multiple features to be worked on simultaneously, without causing conflicts. The master branch contains the stable development version. Instead of working on the master branch, you will create your own branch off the master and merge it back into the master when you are finished.

Create a new branch for your feature before editing the code. For example, if you want to add the new instrument “Extreme 5000” you will make a new branch “dev/extreme-5000”.

```
git branch dev/extreme-5000
```

You can also [make a new branch](#) on GitHub. If you do so, you will have to fetch these changes before the branch will show up on your local computer.

```
git fetch
```

Once you have created the branch, change your current branch to match the new one.

```
git checkout dev/extreme-5000
```

Now you are ready to write your new feature and make changes to the code. To ensure consistency, please follow the [coding standards for PyMeasure](#). Use `git status` to check on the files that have been changed. As you go, commit your changes and push them to your fork.

```
git add file-that-changed.py
git commit -m "A short description about what changed"
git push
```

8.3 Making a pull-request

While you are working, its helpful to start a pull-request (PR) on the master branch of `ralph-group/pymasure`. This will allow you to discuss your feature with other contributors. We encourage you to start this pull-request after your first commit.

Start a pull-request on the [PyMeasure GitHub page](#).

Your pull-request will be merged by the PyMeasure maintainers once it meets the coding standards and passes unit tests. You will notice that your pull-request is automatically checked with the unit tests.

8.4 Unit testing

Unit tests are run each time a new commit is made to a branch. The purpose is to catch changes that break the current functionality, by testing each feature unit. PyMeasure relies on `pytest` to preform these tests, which are run on TravisCI and Appveyor for Linux/macOS and Windows respectively.

Running the unit tests while you develop is highly encouraged. This will ensure that you have a working contribution when you create a pull request.

```
python setup.py test
```

If your feature can be tested, unit tests are required. This will ensure that your features keep working as new features are added.

Now you are familiar with all the pieces of the PyMeasure development work-flow. We look forward to seeing your pull-request!

REPORTING AN ERROR

Please report all errors to the [Issues section](#) of the PyMeasure GitHub repository. Use the search function to determine if there is an existing or resolved issued before posting.

ADDING INSTRUMENTS

You can make a significant contribution to PyMeasure by adding a new instrument to the `pymeasure.instruments` package. Even adding an instrument with a few features can help get the ball rolling, since its likely that others are interested in the same instrument.

Before getting started, become familiar with the *contributing work-flow* for PyMeasure, which steps through the process of adding a new feature (like an instrument) to the development version of the source code. This section will describe how to lay out your instrument code.

10.1 File structure

Your new instrument should be placed in the directory corresponding to the manufacturer of the instrument. For example, if you are going to add an “Extreme 5000” instrument you should add the following files assuming “Extreme” is the manufacturer. Use lowercase for all filenames to distinguish packages from CamelCase Python classes.

```
pymeasure/pymeasure/instruments/extreme/  
|--> __init__.py  
|--> extreme5000.py
```

10.1.1 Updating the init file

The `__init__.py` file in the manufacturer directory should import all of the instruments that correspond to the manufacturer, to allow the files to be easily imported. For a new manufacturer, the manufacturer should also be added to `pymeasure/pymeasure/instruments/__init__.py`. In this case, you also need to add the new package to the `pymeasure/setup.py` file in the `packages` argument.

10.1.2 Adding documentation

Documentation for each instrument is required, and helps others understand the features you have implemented. Add a new reStructuredText file to the documentation.

```
pymeasure/docs/api/instruments/extreme/  
|--> index.rst  
|--> extreme5000.rst
```

Copy an existing instrument documentation file, which will automatically generate the documentation for the instrument. The `index.rst` file should link to the `extreme5000` file. For a new manufacturer, the manufacturer should be also linked in `pymeasure/docs/api/instruments/index.rst`.

10.2 Instrument file

All standard instruments should be child class of *Instrument*. This provides the basic functionality for working with *Adapters*, which perform the actual communication.

The most basic instrument, for our “Extreme 5000” example starts like this:

```
#
# This file is part of the PyMeasure package.
#
# Copyright (c) 2013-2019 PyMeasure Developers
#
# Permission is hereby granted, free of charge, to any person obtaining a copy
# of this software and associated documentation files (the "Software"), to deal
# in the Software without restriction, including without limitation the rights
# to use, copy, modify, merge, publish, distribute, sublicense, and/or sell
# copies of the Software, and to permit persons to whom the Software is
# furnished to do so, subject to the following conditions:
#
# The above copyright notice and this permission notice shall be included in
# all copies or substantial portions of the Software.
#
# THE SOFTWARE IS PROVIDED "AS IS", WITHOUT WARRANTY OF ANY KIND, EXPRESS OR
# IMPLIED, INCLUDING BUT NOT LIMITED TO THE WARRANTIES OF MERCHANTABILITY,
# FITNESS FOR A PARTICULAR PURPOSE AND NONINFRINGEMENT. IN NO EVENT SHALL THE
# AUTHORS OR COPYRIGHT HOLDERS BE LIABLE FOR ANY CLAIM, DAMAGES OR OTHER
# LIABILITY, WHETHER IN AN ACTION OF CONTRACT, TORT OR OTHERWISE, ARISING FROM,
# OUT OF OR IN CONNECTION WITH THE SOFTWARE OR THE USE OR OTHER DEALINGS IN
# THE SOFTWARE.
#
# from pymeasure.instruments import Instrument
```

This is a minimal instrument definition:

```
class Extreme5000(Instrument):
    """ Represents the imaginary Extreme 5000 instrument.
    """

    def __init__(self, resourceName, **kwargs):
        super(Extreme5000, self).__init__(
            resourceName,
            "Extreme 5000",
            **kwargs
        )
```

Make sure to include the PyMeasure license to each file, and add yourself as an author to the `AUTHORS.txt` file.

In principle you are free to write any functions that are necessary for interacting with the instrument. When doing so, make sure to use the `self.ask(command)`, `self.write(command)`, and `self.read()` methods to issue command instead of calling the adapter directly.

In practice, we have developed a number of convenience functions for making instruments easy to write and maintain. The following sections detail these conveniences and are highly encouraged.

10.3 Writing properties

In PyMeasure, [Python properties](#) are the preferred method for dealing with variables that are read or set. PyMeasure comes with two convenience functions for making properties for classes. The `Instrument.measurement` function returns a property that issues a GPIB/SCPI requests when the value is used. For example, if our “Extreme 5000” has the `*IDN?` command we can write the following property to be added above the `def __init__` line in our above example class, or added to the class after the fact as in the code here:

```
Extreme5000.id = Instrument.measurement(
    "*IDN?", """ Reads the instrument identification """
)
```

You will notice that a documentation string is required, and should be descriptive and specific.

When we use this property we will get the identification information.

```
>>> extreme = Extreme5000("GPIB::1")
>>> extreme.id          # Reads "*IDN?"
'Extreme 5000 identification from instrument'
```

The `Instrument.control` function extends this behavior by creating a property that you can read and set. For example, if our “Extreme 5000” has the `:VOLT?` and `:VOLT <float>` commands that are in Volts, we can write the following property.

```
Extreme5000.voltage = Instrument.control(
    ":VOLT?", ":VOLT %g",
    """ A floating point property that controls the voltage
    in Volts. This property can be set.
    """
)
```

You will notice that we use the [Python string format](#) `%g` to pass through the floating point.

We can use this property to set the voltage to 100 mV, which will execute the command and then request the current voltage.

```
>>> extreme = Extreme5000("GPIB::1")
>>> extreme.voltage = 0.1          # Executes ":VOLT 0.1"
>>> extreme.voltage              # Reads ":VOLT?"
0.1
```

Using both of these functions, you can create a number of properties for basic measurements and controls. The next section details additional features of `Instrument.control` that allow you to write properties that cover specific ranges, or have to map between a real value to one used in the command.

10.4 Advanced properties

Many GPIB/SCIP commands are more restrictive than our basic examples above. The `Instrument.control` function has the ability to encode these restrictions using [validators](#). A validator is a function that takes a value and a set of values, and returns a valid value or raises an exception. There are a number of pre-defined validators in `pymeasure.instruments.validators` that should cover most situations. We will cover the four basic types here.

In the examples below we assume you have imported the validators.

10.4.1 In a restricted range

If you have a property with a restricted range, you can use the `strict_range` and `truncated_range` functions. For example, if our “Extreme 5000” can only support voltages from -1 V to 1 V, we can modify our previous example to use a strict validator over this range.

```
Extreme5000.voltage = Instrument.control(
    ":VOLT?", ":VOLT %g",
    """ A floating point property that controls the voltage
    in Volts, from -1 to 1 V. This property can be set. """ ,
    validator=strict_range,
    values=[-1, 1]
)
```

Now our voltage will raise a `ValueError` if the value is out of the range.

```
>>> extreme = Extreme5000("GPIB::1")
>>> extreme.voltage = 100
Traceback (most recent call last):
...
ValueError: Value of 100 is not in range [-1,1]
```

This is useful if you want to alert the programmer that they are using an invalid value. However, sometimes it can be nicer to truncate the value to be within the range.

```
Extreme5000.voltage = Instrument.control(
    ":VOLT?", ":VOLT %g",
    """ A floating point property that controls the voltage
    in Volts, from -1 to 1 V. Invalid voltages are truncated.
    This property can be set. """ ,
    validator=truncated_range,
    values=[-1, 1]
)
```

Now our voltage will not raise an error, and will truncate the value to the range bounds.

```
>>> extreme = Extreme5000("GPIB::1")
>>> extreme.voltage = 100           # Executes ":VOLT 1"
>>> extreme.voltage
1.0
```

10.4.2 In a discrete set

Often a control property should only take a few discrete values. You can use the `strict_discrete_set` and `truncated_discrete_set` functions to handle these situations. The strict version raises an error if the value is not in the set, as in the range examples above.

For example, if our “Extreme 5000” has a `:RANG <float>` command that sets the voltage range that can take values of 10 mV, 100 mV, and 1 V in Volts, then we can write a control as follows.

```
Extreme5000.voltage = Instrument.control(
    ":RANG?", ":RANG %g",
    """ A floating point property that controls the voltage
    range in Volts. This property can be set.
    """ ,
```

(continues on next page)

(continued from previous page)

```

        validator=truncated_discrete_set,
        values=[10e-3, 100e-3, 1]
    )

```

Now we can set the voltage range, which will automatically truncate to an appropriate value.

```

>>> extreme = Extreme5000("GPIB::1")
>>> extreme.voltage = 0.08
>>> extreme.voltage
0.1

```

10.4.3 Using maps

Now that you are familiar with the validators, you can additionally use maps to satisfy instruments which require non-physical values. The `map_values` argument of `Instrument.control` enables this feature.

If your set of values is a list, then the command will use the index of the list. For example, if our “Extreme 5000” instead has a `:RANG <integer>`, where 0, 1, and 2 correspond to 10 mV, 100 mV, and 1 V, then we can use the following control.

```

Extreme5000.voltage = Instrument.control(
    ":RANG?", ":RANG %d",
    """ A floating point property that controls the voltage
    range in Volts, which takes values of 10 mV, 100 mV and 1 V.
    This property can be set. """ ,
    validator=truncated_discrete_set,
    values=[10e-3, 100e-3, 1],
    map_values=True
)

```

Now the actual GPIB/SCIP command is “:RANG 1” for a value of 100 mV, since the index of 100 mV in the values list is 1.

```

>>> extreme = Extreme5000("GPIB::1")
>>> extreme.voltage = 100e-3
>>> extreme.read()
'1'
>>> extreme.voltage = 1
>>> extreme.voltage
1

```

Dictionaries provide a more flexible method for mapping between real-values and those required by the instrument. If instead the `:RANG <integer>` took 1, 2, and 3 to correspond to 10 mV, 100 mV, and 1 V, then we can replace our previous control with the following.

```

Extreme5000.voltage = Instrument.control(
    ":RANG?", ":RANG %d",
    """ A floating point property that controls the voltage
    range in Volts, which takes values of 10 mV, 100 mV and 1 V.
    This property can be set. """ ,
    validator=truncated_discrete_set,
    values={10e-3:1, 100e-3:2, 1:3},
    map_values=True
)

```

```
>>> extreme = Extreme5000("GPIB::1")
>>> extreme.voltage = 10e-3
>>> extreme.read()
'1'
>>> extreme.voltage = 100e-3
>>> extreme.voltage
0.1
```

The dictionary now maps the keys to specific values. The values and keys can be any type, so this can support properties that use strings:

```
Extreme5000.channel = Instrument.control(
    ":CHAN?", ":CHAN %d",
    """ A string property that controls the measurement channel,
        which can take the values X, Y, or Z.
    """,
    validator=strict_discrete_set,
    values={'X':1, 'Y':2, 'Z':3},
    map_values=True
)
```

```
>>> extreme = Extreme5000("GPIB::1")
>>> extreme.channel = 'X'
>>> extreme.read()
'1'
>>> extreme.channel = 'Y'
>>> extreme.channel
'Y'
```

As you have seen, the `Instrument.control` function can be significantly extended by using validators and maps.

CODING STANDARDS

In order to maintain consistency across the different instruments in the PyMeasure repository, we enforce the following standards.

11.1 Python style guides

Only Python 3 is used in PyMeasure. This prevents the maintainance overhead of supporting Python 2.7, which will lose official support in the future.

The [PEP8 style guide](#) and [PEP257 docstring conventions](#) should be followed.

Function and variable names should be lower case with underscores as needed to seperate words. CamelCase should only be used for class names, unless working with Qt, where its use is common.

11.2 Documentation

PyMeasure documents code using reStructuredText and the [Sphinx documentation generator](#). All functions, classes, and methods should be documented in the code using a [docstring](#).

11.3 Usage of getter and setter functions

Getter and setter functions are discouraged, since properties provide a more fluid experience. Given the extensive tools available for defining properties, detailed in the [Advanced properties](#) section, these types of properties are preferred.

AUTHORS

PyMeasure was started in 2013 by Colin Jermain and Graham Rowlands at Cornell University, when it became apparent that both were working on similar Python packages for scientific measurements. PyMeasure combined these efforts and continues to gain valuable contributions from other scientists who are interested in advancing measurement software.

The following developers have contributed to the PyMeasure package:

Colin Jermain
Graham Rowlands
Minh-Hai Nguyen
Guen Prawiro-Atmodjo
Tim van Boxtel
Davide Spirito
Marcos Guimaraes
Ghislain Antony Vaillant
Ben Feinstein
Neal Reynolds
Christoph Buchner
Julian Dlugosch
Vikram Sekar
Casper Schippers
Sumatran Tiger
Dennis Feng
Stefano Pirotta
Moritz Jung

LICENSE

Copyright (c) 2013-2019 PyMeasure Developers

Permission is hereby granted, free of charge, to any person obtaining a copy of this software and associated documentation files (the “Software”), to deal in the Software without restriction, including without limitation the rights to use, copy, modify, merge, publish, distribute, sublicense, and/or sell copies of the Software, and to permit persons to whom the Software is furnished to do so, subject to the following conditions:

The above copyright notice and this permission notice shall be included in all copies or substantial portions of the Software.

THE SOFTWARE IS PROVIDED “AS IS”, WITHOUT WARRANTY OF ANY KIND, EXPRESS OR IMPLIED, INCLUDING BUT NOT LIMITED TO THE WARRANTIES OF MERCHANTABILITY, FITNESS FOR A PARTICULAR PURPOSE AND NONINFRINGEMENT. IN NO EVENT SHALL THE AUTHORS OR COPYRIGHT HOLDERS BE LIABLE FOR ANY CLAIM, DAMAGES OR OTHER LIABILITY, WHETHER IN AN ACTION OF CONTRACT, TORT OR OTHERWISE, ARISING FROM, OUT OF OR IN CONNECTION WITH THE SOFTWARE OR THE USE OR OTHER DEALINGS IN THE SOFTWARE.

PYTHON MODULE INDEX

p

- `pymeasure.display.browser`, 41
- `pymeasure.display.curves`, 41
- `pymeasure.display.inputs`, 42
- `pymeasure.display.listeners`, 43
- `pymeasure.display.log`, 44
- `pymeasure.display.manager`, 44
- `pymeasure.display.plotter`, 45
- `pymeasure.display.thread`, 46
- `pymeasure.display.widgets`, 46
- `pymeasure.display.windows`, 47
- `pymeasure.experiment.experiment`, 33
- `pymeasure.experiment.listeners`, 34
- `pymeasure.experiment.parameters`, 36
- `pymeasure.experiment.procedure`, 35
- `pymeasure.experiment.results`, 39
- `pymeasure.experiment.workers`, 38
- `pymeasure.instruments`, 48
 - `pymeasure.instruments.advantest`, 53
 - `pymeasure.instruments.agilent`, 54
 - `pymeasure.instruments.agilent.agilent4156`, 60
 - `pymeasure.instruments.ami`, 71
 - `pymeasure.instruments.anritsu`, 73
 - `pymeasure.instruments.comedi`, 53
 - `pymeasure.instruments.danfysik`, 76
 - `pymeasure.instruments.fwbell`, 79
 - `pymeasure.instruments.hp`, 80
 - `pymeasure.instruments.keithley`, 81
 - `pymeasure.instruments.lakeshore`, 92
 - `pymeasure.instruments.newport`, 95
 - `pymeasure.instruments.oxfordinstruments`, 96
 - `pymeasure.instruments.parker`, 98
 - `pymeasure.instruments.signalrecovery`, 99
 - `pymeasure.instruments.srs`, 101
 - `pymeasure.instruments.tektronix`, 103
 - `pymeasure.instruments.thorlabs`, 104
 - `pymeasure.instruments.validators`, 51
 - `pymeasure.instruments.yokogawa`, 104

A

- abort () (pymeasure.display.manager.Manager method), 44
- ac_current () (pymeasure.instruments.agilent.AgilentE4980 property), 58
- ac_mode () (pymeasure.instruments.lakeshore.LakeShore425 method), 94
- ac_voltage () (pymeasure.instruments.agilent.AgilentE4980 property), 58
- acquire_reference () (pymeasure.instruments.keithley.Keithley2000 method), 81
- Adapter (class in pymeasure.adapters), 25
- adc1 () (pymeasure.instruments.signalrecovery.DSP7265 property), 100
- adc1 () (pymeasure.instruments.srs.SR830 property), 101
- adc2 () (pymeasure.instruments.signalrecovery.DSP7265 property), 100
- adc2 () (pymeasure.instruments.srs.SR830 property), 101
- adc3 () (pymeasure.instruments.srs.SR830 property), 101
- adc4 () (pymeasure.instruments.srs.SR830 property), 101
- add () (pymeasure.display.browser.Browser method), 41
- add_ramp_step () (pymeasure.instruments.danfysik.Danfysik8500 method), 76
- AFG3152C (class in pymeasure.instruments.tektronix), 104
- Agilent33220A (class in pymeasure.instruments.agilent), 66
- Agilent33500 (class in pymeasure.instruments.agilent), 68
- Agilent33521A (class in pymeasure.instruments.agilent), 71
- Agilent34410A (class in pymeasure.instruments.agilent), 59
- Agilent4156 (class in pymeasure.instruments.agilent.agilent4156), 60
- Agilent8257D (class in pymeasure.instruments.agilent), 54
- Agilent8722ES (class in pymeasure.instruments.agilent), 56
- AgilentE4408B (class in pymeasure.instruments.agilent), 57
- AgilentE4980 (class in pymeasure.instruments.agilent), 58
- AMI430 (class in pymeasure.instruments.ami), 72
- amplitude () (pymeasure.instruments.agilent.Agilent33220A property), 66
- amplitude () (pymeasure.instruments.agilent.Agilent33500 property), 68
- amplitude () (pymeasure.instruments.hp.HP33120A property), 80
- amplitude_depth () (pymeasure.instruments.agilent.Agilent8257D property), 54
- amplitude_source () (pymeasure.instruments.agilent.Agilent8257D property), 54
- amplitude_unit () (pymeasure.instruments.agilent.Agilent33220A property), 66
- amplitude_unit () (pymeasure.instruments.agilent.Agilent33500 property), 68
- amplitude_units () (pymeasure.instruments.hp.HP33120A property), 80
- analysis () (pymeasure.instruments.anritsu.AnritsuMS9710C property), 74
- analysis_result () (pymeasure.instruments.anritsu.AnritsuMS9710C property), 74
- analyzer_mode () (pymeasure.instruments.agilent.agilent4156.Agilent4156 property), 61

- angle() (*pymeasure.instruments.parker.ParkerGV6* property), 98
- angle_error() (*pymeasure.instruments.parker.ParkerGV6* property), 98
- AnritsuMG3692C (class in *pymeasure.instruments.anritsu*), 73
- AnritsuMS9710C (class in *pymeasure.instruments.anritsu*), 74
- aperture() (*pymeasure.instruments.agilent.AgilentE4980* method), 58
- append() (*pymeasure.display.curves.BufferCurve* method), 41
- apply_current() (*pymeasure.instruments.keithley.Keithley2400* method), 87
- apply_current() (*pymeasure.instruments.yokogawa.Yokogawa7651* method), 105
- apply_voltage() (*pymeasure.instruments.keithley.Keithley2400* method), 87
- apply_voltage() (*pymeasure.instruments.yokogawa.Yokogawa7651* method), 105
- arb_advance() (*pymeasure.instruments.agilent.Agilent33500* property), 68
- arb_file() (*pymeasure.instruments.agilent.Agilent33500* property), 68
- arb_filter() (*pymeasure.instruments.agilent.Agilent33500* property), 69
- arb_rate() (*pymeasure.instruments.agilent.Agilent33500* property), 69
- arb_rate() (*pymeasure.instruments.agilent.Agilent33521A* property), 71
- ask() (*pymeasure.adapters.Adapter* method), 25
- ask() (*pymeasure.adapters.FakeAdapter* method), 26
- ask() (*pymeasure.adapters.PrologixAdapter* method), 28
- ask() (*pymeasure.adapters.SerialAdapter* method), 27
- ask() (*pymeasure.adapters.VISAAdapter* method), 29
- ask() (*pymeasure.instruments.fwbell.FWBell5080* method), 79
- ask() (*pymeasure.instruments.Instrument* method), 49
- ask() (*pymeasure.instruments.lakeshore.LakeShoreUSBAdapter* method), 92
- ask_values() (*pymeasure.adapters.VISAAdapter* method), 29
- auto_offset() (*pymeasure.instruments.srs.SR830* method), 101
- auto_pid() (*pymeasure.instruments.oxfordinstruments.ITC503* property), 97
- auto_range() (*pymeasure.instruments.fwbell.FWBell5080* method), 79
- auto_range() (*pymeasure.instruments.keithley.Keithley2000* method), 81
- auto_range() (*pymeasure.instruments.lakeshore.LakeShore425* method), 94
- auto_range_source() (*pymeasure.instruments.keithley.Keithley2400* method), 87
- aux_in_1() (*pymeasure.instruments.srs.SR830* property), 101
- aux_in_2() (*pymeasure.instruments.srs.SR830* property), 101
- aux_in_3() (*pymeasure.instruments.srs.SR830* property), 101
- aux_in_4() (*pymeasure.instruments.srs.SR830* property), 101
- aux_out_1() (*pymeasure.instruments.srs.SR830* property), 101
- aux_out_2() (*pymeasure.instruments.srs.SR830* property), 102
- aux_out_3() (*pymeasure.instruments.srs.SR830* property), 102
- aux_out_4() (*pymeasure.instruments.srs.SR830* property), 102
- average_point() (*pymeasure.instruments.anritsu.AnritsuMS9710C* property), 74
- average_sweep() (*pymeasure.instruments.anritsu.AnritsuMS9710C* property), 74
- axes() (*pymeasure.instruments.newport.ESP300* property), 95
- Axis (class in *pymeasure.instruments.newport.esp300*), 95
- AxisError (class in *pymeasure.instruments.newport.esp300*), 96

B

- beep() (*pymeasure.instruments.agilent.Agilent33220A* method), 66
- beep() (*pymeasure.instruments.agilent.Agilent33500* method), 69
- beep() (*pymeasure.instruments.hp.HP33120A* method), 80

- beep () (*pymeasure.instruments.keithley.Keithley2000 method*), 81
- beep () (*pymeasure.instruments.keithley.Keithley2400 method*), 87
- beep_state () (*pymeasure.instruments.keithley.Keithley2000 property*), 81
- beeper_state () (*pymeasure.instruments.agilent.Agilent33220A property*), 66
- binary_values () (*pymeasure.adapters.Adapter method*), 25
- binary_values () (*pymeasure.adapters.FakeAdapter method*), 26
- binary_values () (*pymeasure.adapters.PrologixAdapter method*), 28
- binary_values () (*pymeasure.adapters.SerialAdapter method*), 27
- binary_values () (*pymeasure.adapters.VISAAdapter method*), 29
- binary_values () (*pymeasure.instruments.lakeshore.LakeShoreUSBAdapter method*), 92
- BooleanInput (*class in pymeasure.display.inputs*), 42
- BooleanParameter (*class in pymeasure.experiment.parameters*), 36
- Browser (*class in pymeasure.display.browser*), 41
- BrowserItem (*class in pymeasure.display.browser*), 41
- BrowserWidget (*class in pymeasure.display.widgets*), 46
- buffer_data () (*pymeasure.instruments.keithley.Keithley2000 property*), 81
- buffer_data () (*pymeasure.instruments.keithley.Keithley2400 property*), 87
- buffer_points () (*pymeasure.instruments.keithley.Keithley2000 property*), 81
- buffer_points () (*pymeasure.instruments.keithley.Keithley2400 property*), 87
- BufferCurve (*class in pymeasure.display.curves*), 41
- burst_mode () (*pymeasure.instruments.agilent.Agilent33220A property*), 66
- burst_mode () (*pymeasure.instruments.agilent.Agilent33500 property*), 69
- burst_ncycles () (*pymeasure.instruments.agilent.Agilent33220A property*), 66
- burst_ncycles () (*pymeasure.instruments.agilent.Agilent33500 property*), 69
- burst_period () (*pymeasure.instruments.agilent.Agilent33500 property*), 69
- burst_state () (*pymeasure.instruments.agilent.Agilent33220A property*), 66
- burst_state () (*pymeasure.instruments.agilent.Agilent33500 property*), 69
- ## C
- center_at_peak () (*pymeasure.instruments.anritsu.AnritsuMS9710C method*), 74
- center_frequency () (*pymeasure.instruments.agilent.Agilent8257D property*), 54
- center_frequency () (*pymeasure.instruments.agilent.AgilentE4408B property*), 57
- channel1 () (*pymeasure.instruments.srs.SR830 property*), 102
- channel2 () (*pymeasure.instruments.srs.SR830 property*), 102
- channel_function () (*pymeasure.instruments.agilent.agilent4156.SMU property*), 63
- channel_function () (*pymeasure.instruments.agilent.agilent4156.VSU property*), 66
- channel_mode () (*pymeasure.instruments.agilent.agilent4156.SMU property*), 63
- channel_mode () (*pymeasure.instruments.agilent.agilent4156.VMU property*), 65
- channel_mode () (*pymeasure.instruments.agilent.agilent4156.VSU property*), 66
- check_errors () (*pymeasure.instruments.agilent.Agilent33220A method*), 67
- check_errors () (*pymeasure.instruments.agilent.Agilent33500 method*), 69
- check_errors () (*pymeasure.instruments.Instrument method*), 49
- check_errors () (*pymeasure.instruments.keithley.Keithley2000 method*), 81
- check_errors () (*pymeasure.instruments.keithley.Keithley2400 method*), 81

- method), 87
- check_parameters() (pymeasure.experiment.procedure.Procedure method), 35
- check_stop() (pymeasure.display.windows.PlotterWindow method), 48
- choices() (pymeasure.experiment.parameters.ListParameter property), 37
- clear() (pymeasure.display.manager.Manager method), 44
- clear() (pymeasure.instruments.Instrument method), 49
- clear_display() (pymeasure.instruments.agilent.Agilent33500 method), 69
- clear_errors() (pymeasure.instruments.newport.ESP300 method), 95
- clear_plot() (pymeasure.experiment.experiment.Experiment method), 33
- clear_ramp_set() (pymeasure.instruments.danfysik.Danfysik8500 method), 77
- clear_sequence() (pymeasure.instruments.danfysik.Danfysik8500 method), 77
- coilconst() (pymeasure.instruments.ami.AMI430 property), 72
- compliance() (pymeasure.instruments.agilent.agilent4156.SMU property), 63
- compliance() (pymeasure.instruments.agilent.agilent4156.VARD property), 64
- compliance() (pymeasure.instruments.agilent.agilent4156.VARX property), 65
- compliance_current() (pymeasure.instruments.keithley.Keithley2400 property), 88
- compliance_current() (pymeasure.instruments.yokogawa.Yokogawa7651 property), 105
- compliance_voltage() (pymeasure.instruments.keithley.Keithley2400 property), 88
- compliance_voltage() (pymeasure.instruments.yokogawa.Yokogawa7651 property), 105
- config() (pymeasure.adapters.VISAAdapter method), 30
- config_amplitude_modulation() (pymeasure.instruments.agilent.Agilent8257D method), 54
- config_buffer() (pymeasure.instruments.keithley.Keithley2000 method), 81
- config_buffer() (pymeasure.instruments.keithley.Keithley2400 method), 88
- config_low_freq_out() (pymeasure.instruments.agilent.Agilent8257D method), 54
- config_pulse_modulation() (pymeasure.instruments.agilent.Agilent8257D method), 54
- config_step_sweep() (pymeasure.instruments.agilent.Agilent8257D method), 55
- configure() (pymeasure.instruments.agilent.agilent4156.Agilent4156 method), 61
- constant_value() (pymeasure.instruments.agilent.agilent4156.SMU property), 63
- constant_value() (pymeasure.instruments.agilent.agilent4156.VSU property), 66
- control() (pymeasure.instruments.Instrument static method), 49
- control_mode() (pymeasure.instruments.oxfordinstruments.ITC503 property), 97
- create_filename() (in module pymeasure.experiment.experiment), 34
- Crosshairs (class in pymeasure.display.curves), 41
- CSVFormatter (class in pymeasure.experiment.results), 39
- current() (pymeasure.instruments.danfysik.Danfysik8500 property), 77
- current() (pymeasure.instruments.keithley.Keithley2000 property), 82
- current() (pymeasure.instruments.keithley.Keithley2400 property), 88
- current_ac() (pymeasure.instruments.agilent.Agilent34410A property), 59
- current_ac_bandwidth() (pymeasure.instruments.keithley.Keithley2000 property), 82
- current_ac_digits() (pymeasure.instruments.keithley.Keithley2000 property), 82
- current_ac_nplc() (pymeasure.instruments.keithley.Keithley2000 property), 82

- `current_ac_range()` (*pymeasure.instruments.keithley.Keithley2000* property), 82
`current_ac_reference()` (*pymeasure.instruments.keithley.Keithley2000* property), 82
`current_dc()` (*pymeasure.instruments.agilent.Agilent34410A* property), 59
`current_digits()` (*pymeasure.instruments.keithley.Keithley2000* property), 82
`current_name()` (*pymeasure.instruments.agilent.agilent4156.SMU* property), 63
`current_nplc()` (*pymeasure.instruments.keithley.Keithley2000* property), 82
`current_nplc()` (*pymeasure.instruments.keithley.Keithley2400* property), 88
`current_ppm()` (*pymeasure.instruments.danfysik.Danfysik8500* property), 77
`current_range()` (*pymeasure.instruments.keithley.Keithley2000* property), 82
`current_range()` (*pymeasure.instruments.keithley.Keithley2400* property), 88
`current_reference()` (*pymeasure.instruments.keithley.Keithley2000* property), 82
`current_setpoint()` (*pymeasure.instruments.danfysik.Danfysik8500* property), 77
- D**
- `dac1()` (*pymeasure.instruments.signalrecovery.DSP7265* property), 100
`dac1()` (*pymeasure.instruments.srs.SR830* property), 102
`dac2()` (*pymeasure.instruments.signalrecovery.DSP7265* property), 100
`dac2()` (*pymeasure.instruments.srs.SR830* property), 102
`dac3()` (*pymeasure.instruments.signalrecovery.DSP7265* property), 100
`dac3()` (*pymeasure.instruments.srs.SR830* property), 102
`dac4()` (*pymeasure.instruments.signalrecovery.DSP7265* property), 100
`dac4()` (*pymeasure.instruments.srs.SR830* property), 102
- `Danfysik8500` (class in *pymeasure.instruments.danfysik*), 76
`DanfysikAdapter` (class in *pymeasure.instruments.danfysik*), 76
`data()` (*pymeasure.experiment.experiment.Experiment* property), 33
`data()` (*pymeasure.instruments.agilent.Agilent8722ES* property), 56
`data_arb()` (*pymeasure.instruments.agilent.Agilent33500* method), 69
`data_memory_a_condition()` (*pymeasure.instruments.anritsu.AnritsuMS9710C* property), 74
`data_memory_a_size()` (*pymeasure.instruments.anritsu.AnritsuMS9710C* property), 74
`data_memory_a_values()` (*pymeasure.instruments.anritsu.AnritsuMS9710C* property), 74
`data_memory_b_condition()` (*pymeasure.instruments.anritsu.AnritsuMS9710C* property), 74
`data_memory_b_size()` (*pymeasure.instruments.anritsu.AnritsuMS9710C* property), 74
`data_memory_b_values()` (*pymeasure.instruments.anritsu.AnritsuMS9710C* property), 74
`data_memory_select()` (*pymeasure.instruments.anritsu.AnritsuMS9710C* property), 74
`data_variables()` (*pymeasure.instruments.agilent.agilent4156.Agilent4156* property), 61
`data_volatile_clear()` (*pymeasure.instruments.agilent.Agilent33500* method), 69
`dc_mode()` (*pymeasure.instruments.lakeshore.LakeShore425* method), 94
`define_position()` (*pymeasure.instruments.newport.esp300.Axis* method), 95
`delay_time()` (*pymeasure.instruments.agilent.agilent4156.Agilent4156* property), 62
`dip_search()` (*pymeasure.instruments.anritsu.AnritsuMS9710C* property), 74
`disable()` (*pymeasure.instruments.agilent.agilent4156.SMU* property), 64
`disable()` (*pymeasure.instruments.agilent.agilent4156.VMU* property), 65
`disable()` (*pymeasure.instruments.agilent.agilent4156.VSU*

- property*), 66
- `disable()` (*pymeasure.instruments.agilent.Agilent8257D method*), 55
- `disable()` (*pymeasure.instruments.anritsu.AnritsuMG3692C method*), 73
- `disable()` (*pymeasure.instruments.danfysik.Danfysik8500SP7265 method*), 77
- `disable()` (*pymeasure.instruments.newport.ESP300 method*), 95
- `disable()` (*pymeasure.instruments.newport.esp300.Axis method*), 95
- `disable()` (*pymeasure.instruments.parker.ParkerGV6 method*), 99
- `disable_all()` (*pymeasure.instruments.agilent.agilent4156.Agilent4156 method*), 62
- `disable_amplitude_modulation()` (*pymeasure.instruments.agilent.Agilent8257D method*), 55
- `disable_averaging()` (*pymeasure.instruments.agilent.Agilent8722ES method*), 56
- `disable_buffer()` (*pymeasure.instruments.keithley.Keithley2000 method*), 82
- `disable_buffer()` (*pymeasure.instruments.keithley.Keithley2400 method*), 88
- `disable_filter()` (*pymeasure.instruments.keithley.Keithley2000 method*), 82
- `disable_heater()` (*pymeasure.instruments.lakeshore.LakeShore331 method*), 93
- `disable_low_freq_out()` (*pymeasure.instruments.agilent.Agilent8257D method*), 55
- `disable_modulation()` (*pymeasure.instruments.agilent.Agilent8257D method*), 55
- `disable_output_trigger()` (*pymeasure.instruments.keithley.Keithley2400 method*), 88
- `disable_persistent_switch()` (*pymeasure.instruments.ami.AMI430 method*), 72
- `disable_pulse_modulation()` (*pymeasure.instruments.agilent.Agilent8257D method*), 55
- `disable_reference()` (*pymeasure.instruments.keithley.Keithley2000 method*), 82
- `disable_source()` (*pymeasure.instruments.keithley.Keithley2400 method*), 88
- `disable_source()` (*pymeasure.instruments.yokogawa.Yokogawa7651 method*), 105
- `display()` (*pymeasure.instruments.agilent.Agilent33500 property*), 70
- `display()` (*class in pymeasure.instruments.signalrecovery*), 100
- `dwell_time()` (*pymeasure.instruments.agilent.Agilent8257D property*), 55
- ## E
- `echo()` (*pymeasure.instruments.parker.ParkerGV6 method*), 99
- `emit()` (*pymeasure.display.log.LogHandler method*), 44
- `emit()` (*pymeasure.experiment.workers.Worker method*), 38
- `enable()` (*pymeasure.instruments.agilent.Agilent8257D method*), 55
- `enable()` (*pymeasure.instruments.anritsu.AnritsuMG3692C method*), 73
- `enable()` (*pymeasure.instruments.danfysik.Danfysik8500 method*), 77
- `enable()` (*pymeasure.instruments.newport.ESP300 method*), 95
- `enable()` (*pymeasure.instruments.newport.esp300.Axis method*), 95
- `enable()` (*pymeasure.instruments.parker.ParkerGV6 method*), 99
- `enable_amplitude_modulation()` (*pymeasure.instruments.agilent.Agilent8257D method*), 55
- `enable_filter()` (*pymeasure.instruments.keithley.Keithley2000 method*), 82
- `enable_low_freq_out()` (*pymeasure.instruments.agilent.Agilent8257D method*), 55
- `enable_persistent_switch()` (*pymeasure.instruments.ami.AMI430 method*), 72
- `enable_pulse_modulation()` (*pymeasure.instruments.agilent.Agilent8257D method*), 55
- `enable_reference()` (*pymeasure.instruments.keithley.Keithley2000 method*), 83
- `enable_source()` (*pymeasure.instruments.keithley.Keithley2400 method*), 88
- `enable_source()` (*pymeasure.instruments.yokogawa.Yokogawa7651 method*), 105

enabled() (*pymeasure.instruments.newport.esp300.Axis* property), 96
 error() (*pymeasure.instruments.keithley.Keithley2400* property), 88
 error() (*pymeasure.instruments.newport.ESP300* property), 95
 errors() (*pymeasure.instruments.newport.ESP300* property), 95
 ese2() (*pymeasure.instruments.anritsu.AnritsuMS9710C* property), 74
 ESP300 (*class in pymeasure.instruments.newport*), 95
 esr2() (*pymeasure.instruments.anritsu.AnritsuMS9710C* property), 74
 execute() (*pymeasure.experiment.procedure.Procedure* method), 35
 Experiment (*class in pymeasure.display.manager*), 44
 Experiment (*class in pymeasure.experiment.experiment*), 33
 ExperimentQueue (*class in pymeasure.display.manager*), 44
 ext_trig_out() (*pymeasure.instruments.agilent.Agilent33500* property), 70
F
 FakeAdapter (*class in pymeasure.adapters*), 26
 field() (*pymeasure.instruments.ami.AMI430* property), 72
 field() (*pymeasure.instruments.fwbell.FWBell5080* property), 79
 field() (*pymeasure.instruments.lakeshore.LakeShore425* property), 94
 fields() (*pymeasure.instruments.fwbell.FWBell5080* method), 79
 filter_slope() (*pymeasure.instruments.srs.SR830* property), 102
 FloatInput (*class in pymeasure.display.inputs*), 42
 FloatParameter (*class in pymeasure.experiment.parameters*), 36
 format() (*pymeasure.experiment.results.CSVFormatter* method), 39
 format() (*pymeasure.experiment.results.Results* method), 39
 freq_sweep() (*pymeasure.instruments.agilent.AgilentE4980* method), 58
 frequencies() (*pymeasure.instruments.agilent.Agilent8722ES* property), 56
 frequencies() (*pymeasure.instruments.agilent.AgilentE4408B* property), 57
 frequency() (*pymeasure.instruments.agilent.Agilent33220A* property), 67
 frequency() (*pymeasure.instruments.agilent.Agilent33500* property), 70
 frequency() (*pymeasure.instruments.agilent.Agilent33521A* property), 71
 frequency() (*pymeasure.instruments.agilent.Agilent8257D* property), 55
 frequency() (*pymeasure.instruments.agilent.AgilentE4980* property), 58
 frequency() (*pymeasure.instruments.anritsu.AnritsuMG3692C* property), 73
 frequency() (*pymeasure.instruments.hp.HP33120A* property), 80
 frequency() (*pymeasure.instruments.keithley.Keithley2000* property), 83
 frequency() (*pymeasure.instruments.signalrecovery.DSP7265* property), 100
 frequency() (*pymeasure.instruments.srs.SR830* property), 102
 frequency_aperature() (*pymeasure.instruments.keithley.Keithley2000* property), 83
 frequency_digits() (*pymeasure.instruments.keithley.Keithley2000* property), 83
 frequency_points() (*pymeasure.instruments.agilent.AgilentE4408B* property), 57
 frequency_reference() (*pymeasure.instruments.keithley.Keithley2000* property), 83
 frequency_step() (*pymeasure.instruments.agilent.AgilentE4408B* property), 57
 frequency_threshold() (*pymeasure.instruments.keithley.Keithley2000* property), 83
 FWBell5080 (*class in pymeasure.instruments.fwbell*), 79
G
 gen_measurement() (*pymeasure.experiment.procedure.Procedure* method), 35
 GeneralError (*class in pymeasure.instruments.newport.esp300*), 96

- get_array() (in module *pymeasure.experiment.experiment*), 34
 get_array_steps() (in module *pymeasure.experiment.experiment*), 34
 get_array_zero() (in module *pymeasure.experiment.experiment*), 34
 get_buffer() (*pymeasure.instruments.srs.SR830* method), 102
 get_data() (*pymeasure.instruments.agilent.agilent4156.Agilent4156* method), 62
 get_procedure() (*pymeasure.display.widgets.InputsWidget* method), 46
 get_scaling() (*pymeasure.instruments.srs.SR830* method), 102
 get_time() (*pymeasure.instruments.Mock* method), 51
 get_voltage() (*pymeasure.instruments.Mock* method), 51
 get_wave() (*pymeasure.instruments.Mock* method), 51
 getAI() (in module *pymeasure.instruments.comedi*), 53
 getAO() (in module *pymeasure.instruments.comedi*), 53
 gpib() (*pymeasure.adapters.PrologixAdapter* method), 28
- ## H
- handle_abort() (*pymeasure.experiment.workers.Worker* method), 38
 handle_error() (*pymeasure.experiment.workers.Worker* method), 38
 harmonic() (*pymeasure.instruments.signalrecovery.DSP7265* property), 100
 harmonic() (*pymeasure.instruments.srs.SR830* property), 102
 has_amplitude_modulation() (*pymeasure.instruments.agilent.Agilent8257D* property), 55
 has_modulation() (*pymeasure.instruments.agilent.Agilent8257D* property), 55
 has_next() (*pymeasure.display.manager.ExperimentQueue* method), 44
 has_persistent_switch_enabled() (*pymeasure.instruments.ami.AMI430* method), 72
 has_pulse_modulation() (*pymeasure.instruments.agilent.Agilent8257D* property), 55
- has_supported_version() (*pymeasure.adapters.VISAAdapter* static method), 30
 header() (*pymeasure.experiment.results.Results* method), 39
 heater_gas_mode() (*pymeasure.instruments.oxfordinstruments.ITC503* property), 97
 heater_range() (*pymeasure.instruments.lakeshore.LakeShore331* property), 93
 hold_time() (*pymeasure.instruments.agilent.agilent4156.Agilent4156* property), 62
 home() (*pymeasure.instruments.newport.esp300.Axis* method), 96
 HP33120A (class in *pymeasure.instruments.hp*), 80
- ## I
- id() (*pymeasure.instruments.agilent.Agilent33500* property), 70
 id() (*pymeasure.instruments.danfysik.Danfysik8500* property), 77
 id() (*pymeasure.instruments.fwbell.FWBell5080* property), 79
 id() (*pymeasure.instruments.Instrument* property), 50
 id() (*pymeasure.instruments.keithley.Keithley2000* property), 83
 id() (*pymeasure.instruments.keithley.Keithley2400* property), 88
 id() (*pymeasure.instruments.signalrecovery.DSP7265* property), 100
 id() (*pymeasure.instruments.yokogawa.Yokogawa7651* property), 105
 impedance() (*pymeasure.instruments.agilent.AgilentE4980* property), 58
 Input (class in *pymeasure.display.inputs*), 42
 input_config() (*pymeasure.instruments.srs.SR830* property), 102
 input_coupling() (*pymeasure.instruments.srs.SR830* property), 102
 input_grounding() (*pymeasure.instruments.srs.SR830* property), 103
 input_notch_config() (*pymeasure.instruments.srs.SR830* property), 103
 InputsWidget (class in *pymeasure.display.widgets*), 46
 Instrument (class in *pymeasure.instruments*), 49
 IntegerInput (class in *pymeasure.display.inputs*), 43
 IntegerParameter (class in *pymeasure.experiment.parameters*), 36
 integration_time() (*pymeasure.instruments.agilent.agilent4156.Agilent4156*

- property*), 62
- `internal_frequency()` (*pymeasure.instruments.agilent.Agilent8257D property*), 55
- `internal_shape()` (*pymeasure.instruments.agilent.Agilent8257D property*), 55
- `is_averaging()` (*pymeasure.instruments.agilent.Agilent8722ES method*), 56
- `is_buffer_full()` (*pymeasure.instruments.keithley.Keithley2000 method*), 83
- `is_buffer_full()` (*pymeasure.instruments.keithley.Keithley2400 method*), 88
- `is_current_stable()` (*pymeasure.instruments.danfysik.Danfysik8500 method*), 77
- `is_enabled()` (*pymeasure.instruments.agilent.Agilent8257D property*), 55
- `is_enabled()` (*pymeasure.instruments.danfysik.Danfysik8500 method*), 77
- `is_moving()` (*pymeasure.instruments.parker.ParkerGV6 method*), 99
- `is_out_of_range()` (*pymeasure.instruments.srs.SR830 method*), 103
- `is_ready()` (*pymeasure.instruments.danfysik.Danfysik8500 method*), 77
- `is_running()` (*pymeasure.display.manager.Manager method*), 44
- `is_sequence_running()` (*pymeasure.instruments.danfysik.Danfysik8500 method*), 77
- `is_set()` (*pymeasure.experiment.parameters.Parameter method*), 37
- ITC503 (*class in pymeasure.instruments.oxfordinstruments*), 97
- ## J
- `join()` (*pymeasure.display.thread.StoppableQThread method*), 46
- `join()` (*pymeasure.experiment.workers.Worker method*), 38
- ## K
- Keithley2000 (*class in pymeasure.instruments.keithley*), 81
- Keithley2400 (*class in pymeasure.instruments.keithley*), 87
- `kill()` (*pymeasure.instruments.parker.ParkerGV6 method*), 99
- ## L
- `labels()` (*pymeasure.experiment.results.Results method*), 39
- LakeShore331 (*class in pymeasure.instruments.lakeshore*), 93
- LakeShore425 (*class in pymeasure.instruments.lakeshore*), 94
- LakeShoreUSBAdapter (*class in pymeasure.instruments.lakeshore*), 92
- `left_limit()` (*pymeasure.instruments.newport.esp300.Axis property*), 96
- `level_lin()` (*pymeasure.instruments.anritsu.AnritsuMS9710C property*), 74
- `level_log()` (*pymeasure.instruments.anritsu.AnritsuMS9710C property*), 74
- `level_opt_attn()` (*pymeasure.instruments.anritsu.AnritsuMS9710C property*), 74
- `level_scale()` (*pymeasure.instruments.anritsu.AnritsuMS9710C property*), 75
- `list_resources()` (*in module pymeasure.instruments*), 53
- Listener (*class in pymeasure.experiment.listeners*), 34
- ListInput (*class in pymeasure.display.inputs*), 43
- ListParameter (*class in pymeasure.experiment.parameters*), 36
- `load()` (*pymeasure.display.manager.Manager method*), 45
- `load()` (*pymeasure.experiment.results.Results static method*), 39
- `local()` (*pymeasure.instruments.danfysik.Danfysik8500 method*), 77
- `local()` (*pymeasure.instruments.keithley.Keithley2000 method*), 83
- `log_magnitude()` (*pymeasure.instruments.agilent.Agilent8722ES method*), 57
- LogHandler (*class in pymeasure.display.log*), 44
- LogWidget (*class in pymeasure.display.widgets*), 46
- `low_freq_out_amplitude()` (*pymeasure.instruments.agilent.Agilent8257D property*), 55
- `low_freq_out_source()` (*pymeasure.instruments.agilent.Agilent8257D property*), 56

M

- `mag()` (`pymeasure.instruments.signalrecovery.DSP7265` property), 100
- `magnet_current()` (`pymeasure.instruments.ami.AMI430` property), 72
- `magnitude()` (`pymeasure.instruments.agilent.Agilent8722ES` method), 57
- `magnitude()` (`pymeasure.instruments.srs.SR830` property), 103
- `ManagedWindow` (class in `pymeasure.display.windows`), 47
- `Manager` (class in `pymeasure.display.manager`), 44
- `max_amplitude()` (`pymeasure.instruments.hp.HP33120A` property), 80
- `max_current()` (`pymeasure.instruments.keithley.Keithley2400` property), 88
- `max_frequency()` (`pymeasure.instruments.hp.HP33120A` property), 80
- `max_offset()` (`pymeasure.instruments.hp.HP33120A` property), 80
- `max_resistance()` (`pymeasure.instruments.keithley.Keithley2400` property), 88
- `max_voltage()` (`pymeasure.instruments.keithley.Keithley2400` property), 88
- `maximums()` (`pymeasure.instruments.keithley.Keithley2400` property), 88
- `mean_current()` (`pymeasure.instruments.keithley.Keithley2400` property), 88
- `mean_resistance()` (`pymeasure.instruments.keithley.Keithley2400` property), 89
- `mean_voltage()` (`pymeasure.instruments.keithley.Keithley2400` property), 89
- `means()` (`pymeasure.instruments.keithley.Keithley2400` property), 89
- `Measurable` (class in `pymeasure.experiment.parameters`), 37
- `measure()` (`pymeasure.instruments.agilent.agilent4156.Agilent4156` method), 62
- `measure()` (`pymeasure.instruments.lakeshore.LakeShore425` method), 94
- `measure_continuity()` (`pymeasure.instruments.keithley.Keithley2000` method), 83
- `measure_current()` (`pymeasure.instruments.keithley.Keithley2000` method), 83
- `measure_current()` (`pymeasure.instruments.keithley.Keithley2400` method), 89
- `measure_diode()` (`pymeasure.instruments.keithley.Keithley2000` method), 83
- `measure_frequency()` (`pymeasure.instruments.keithley.Keithley2000` method), 83
- `measure_mode()` (`pymeasure.instruments.anritsu.AnritsuMS9710C` property), 75
- `measure_peak()` (`pymeasure.instruments.anritsu.AnritsuMS9710C` method), 75
- `measure_period()` (`pymeasure.instruments.keithley.Keithley2000` method), 83
- `measure_power()` (`pymeasure.instruments.thorlabs.ThorlabsPM100USB` method), 104
- `measure_resistance()` (`pymeasure.instruments.keithley.Keithley2000` method), 83
- `measure_resistance()` (`pymeasure.instruments.keithley.Keithley2400` method), 89
- `measure_temperature()` (`pymeasure.instruments.keithley.Keithley2000` method), 84
- `measure_voltage()` (`pymeasure.instruments.keithley.Keithley2000` method), 84
- `measure_voltage()` (`pymeasure.instruments.keithley.Keithley2400` method), 89
- `measurement()` (`pymeasure.instruments.Instrument` static method), 50
- `message_waiting()` (`pymeasure.experiment.listeners.Listener` method), 34
- `min_amplitude()` (`pymeasure.instruments.hp.HP33120A` property), 80
- `min_current()` (`pymeasure.instruments.keithley.Keithley2400` property), 89
- `min_frequency()` (`pymeasure.instruments.hp.HP33120A` property), 80
- `min_offset()` (`pymeasure.instruments.hp.HP33120A` property), 80

property), 80
 min_resistance() (*pymeasure.instruments.keithley.Keithley2400 property*), 89
 min_voltage() (*pymeasure.instruments.keithley.Keithley2400 property*), 89
 minimums() (*pymeasure.instruments.keithley.Keithley2400 property*), 89
 Mock (*class in pymeasure.instruments*), 51
 mode() (*pymeasure.instruments.agilent.AgilentE4980 property*), 58
 mode() (*pymeasure.instruments.keithley.Keithley2000 property*), 84
 Monitor (*class in pymeasure.display.listeners*), 43
 Monitor (*class in pymeasure.experiment.listeners*), 34
 motion_done() (*pymeasure.instruments.newport.esp300.Axis property*), 96
 mouseMoved() (*pymeasure.display.curves.Crosshairs method*), 41
 move() (*pymeasure.instruments.parker.ParkerGV6 method*), 99

N

next() (*pymeasure.display.manager.ExperimentQueue method*), 44
 next() (*pymeasure.display.manager.Manager method*), 45

O

offset() (*pymeasure.instruments.agilent.Agilent33220A property*), 67
 offset() (*pymeasure.instruments.agilent.Agilent33500 property*), 70
 offset() (*pymeasure.instruments.agilent.agilent4156.VARD property*), 64
 offset() (*pymeasure.instruments.hp.HP33120A property*), 80
 output() (*pymeasure.instruments.agilent.Agilent33220A property*), 67
 output() (*pymeasure.instruments.agilent.Agilent33500 property*), 70
 output() (*pymeasure.instruments.anritsu.AnritsuMG3692C property*), 73
 output_conversion() (*pymeasure.instruments.srs.SR830 method*), 103
 output_load() (*pymeasure.instruments.agilent.Agilent33500 property*), 70
 output_trigger_on_external() (*pymeasure.instruments.keithley.Keithley2400 method*), 89

P

Parameter (*class in pymeasure.experiment.parameters*), 37
 parameter() (*pymeasure.display.inputs.Input property*), 42
 parameter_objects() (*pymeasure.experiment.procedure.Procedure method*), 35
 parameter_values() (*pymeasure.experiment.procedure.Procedure method*), 35
 parameters_are_set() (*pymeasure.experiment.procedure.Procedure method*), 35
 ParkerGV6 (*class in pymeasure.instruments.parker*), 98
 parse() (*pymeasure.experiment.results.Results method*), 39
 parse_axis() (*pymeasure.display.widgets.PlotFrame method*), 46
 parse_header() (*pymeasure.experiment.results.Results static method*), 39
 pause() (*pymeasure.instruments.ami.AMI430 method*), 72
 pcolor() (*pymeasure.experiment.experiment.Experiment method*), 34
 peak_search() (*pymeasure.instruments.anritsu.AnritsuMS9710C property*), 75
 period() (*pymeasure.instruments.keithley.Keithley2000 property*), 84
 period_aperture() (*pymeasure.instruments.keithley.Keithley2000 property*), 84
 period_digits() (*pymeasure.instruments.keithley.Keithley2000 property*), 84
 period_reference() (*pymeasure.instruments.keithley.Keithley2000 property*), 84
 period_threshold() (*pymeasure.instruments.keithley.Keithley2000 property*), 84
 phase() (*pymeasure.instruments.agilent.Agilent8722ES method*), 57
 phase() (*pymeasure.instruments.signalrecovery.DSP7265 property*), 100
 phase() (*pymeasure.instruments.srs.SR830 property*), 103
 PhysicalParameter (*class in pymeasure.experiment.parameters*), 37
 plot (*pymeasure.display.windows.ManagedWindow attribute*), 47

[plot\(\)](#) (*pymeasure.experiment.experiment.Experiment method*), 34
[plot_live\(\)](#) (*pymeasure.experiment.experiment.Experiment method*), 34
[PlotFrame](#) (*class in pymeasure.display.widgets*), 46
[Plotter](#) (*class in pymeasure.display.plotter*), 45
[PlotterWindow](#) (*class in pymeasure.display.windows*), 47
[PlotWidget](#) (*class in pymeasure.display.widgets*), 46
[points\(\)](#) (*pymeasure.instruments.agilent.agilent4156.VAR2 property*), 64
[polarity\(\)](#) (*pymeasure.instruments.danfysik.Danfysik8500 property*), 77
[position\(\)](#) (*pymeasure.instruments.newport.esp300.Axis property*), 96
[position\(\)](#) (*pymeasure.instruments.parker.ParkerGV6 property*), 99
[position_error\(\)](#) (*pymeasure.instruments.parker.ParkerGV6 property*), 99
[power\(\)](#) (*pymeasure.instruments.agilent.Agilent8257D property*), 56
[power\(\)](#) (*pymeasure.instruments.anritsu.AnritsuMG3692 property*), 73
[power\(\)](#) (*pymeasure.instruments.thorlabs.ThorlabsPM100 property*), 104
[prepare\(\)](#) (*pymeasure.display.curves.BufferCurve method*), 41
[Procedure](#) (*class in pymeasure.experiment.procedure*), 35
[program_sweep\(\)](#) (*pymeasure.instruments.oxfordinstruments.ITC503 method*), 97
[PrologixAdapter](#) (*class in pymeasure.adapters*), 28
[pulse_dutycycle\(\)](#) (*pymeasure.instruments.agilent.Agilent33220A property*), 67
[pulse_dutycycle\(\)](#) (*pymeasure.instruments.agilent.Agilent33500 property*), 70
[pulse_frequency\(\)](#) (*pymeasure.instruments.agilent.Agilent8257D property*), 56
[pulse_hold\(\)](#) (*pymeasure.instruments.agilent.Agilent33220A property*), 67
[pulse_hold\(\)](#) (*pymeasure.instruments.agilent.Agilent33500 property*), 70
[pulse_input\(\)](#) (*pymeasure.instruments.agilent.Agilent8257D property*), 56
[pulse_period\(\)](#) (*pymeasure.instruments.agilent.Agilent33220A property*), 67
[pulse_period\(\)](#) (*pymeasure.instruments.agilent.Agilent33500 property*), 70
[pulse_source\(\)](#) (*pymeasure.instruments.agilent.Agilent8257D property*), 56
[pulse_transition\(\)](#) (*pymeasure.instruments.agilent.Agilent33220A property*), 67
[pulse_transition\(\)](#) (*pymeasure.instruments.agilent.Agilent33500 property*), 70
[pulse_width\(\)](#) (*pymeasure.instruments.agilent.Agilent33220A property*), 67
[pulse_width\(\)](#) (*pymeasure.instruments.agilent.Agilent33500 property*), 70
[pymeasure.display.browser](#) (*module*), 41
[pymeasure.display.curves](#) (*module*), 41
[pymeasure.display.inputs](#) (*module*), 42
[pymeasure.display.listeners](#) (*module*), 43
[pymeasure.display.log](#) (*module*), 44
[pymeasure.display.manager](#) (*module*), 44
[pymeasure.display.plotter](#) (*module*), 45
[pymeasure.display.thread](#) (*module*), 46
[pymeasure.display.widgets](#) (*module*), 46
[pymeasure.display.windows](#) (*module*), 47
[pymeasure.experiment.experiment](#) (*module*), 33
[pymeasure.experiment.listeners](#) (*module*), 34
[pymeasure.experiment.parameters](#) (*module*), 36
[pymeasure.experiment.procedure](#) (*module*), 35
[pymeasure.experiment.results](#) (*module*), 39
[pymeasure.experiment.workers](#) (*module*), 38
[pymeasure.instruments](#) (*module*), 48
[pymeasure.instruments.advantest](#) (*module*), 53
[pymeasure.instruments.agilent](#) (*module*), 54
[pymeasure.instruments.agilent.agilent4156](#) (*module*), 60
[pymeasure.instruments.ami](#) (*module*), 71
[pymeasure.instruments.anritsu](#) (*module*), 73
[pymeasure.instruments.comedi](#) (*module*), 53
[pymeasure.instruments.danfysik](#) (*module*), 76

- pymeasure.instruments.fwbell (*module*), 79
 pymeasure.instruments.hp (*module*), 80
 pymeasure.instruments.keithley (*module*), 81
 pymeasure.instruments.lakeshore (*module*), 92
 pymeasure.instruments.newport (*module*), 95
 pymeasure.instruments.oxfordinstruments (*module*), 96
 pymeasure.instruments.parker (*module*), 98
 pymeasure.instruments.signalrecovery (*module*), 99
 pymeasure.instruments.srs (*module*), 101
 pymeasure.instruments.tektronix (*module*), 103
 pymeasure.instruments.thorlabs (*module*), 104
 pymeasure.instruments.validators (*module*), 51
 pymeasure.instruments.yokogawa (*module*), 104
- ## Q
- QListener (*class in pymeasure.display.listeners*), 43
 queue () (*pymeasure.display.manager.Manager method*), 45
 queue () (*pymeasure.display.windows.ManagedWindow method*), 47
 quick_range () (*pymeasure.instruments.srs.SR830 method*), 103
- ## R
- ramp () (*pymeasure.instruments.ami.AMI430 method*), 72
 ramp_rate_current () (*pymeasure.instruments.ami.AMI430 property*), 72
 ramp_rate_field () (*pymeasure.instruments.ami.AMI430 property*), 72
 ramp_symmetry () (*pymeasure.instruments.agilent.Agilent33220A property*), 67
 ramp_symmetry () (*pymeasure.instruments.agilent.Agilent33500 property*), 70
 ramp_to_current () (*pymeasure.instruments.ami.AMI430 method*), 73
 ramp_to_current () (*pymeasure.instruments.danfysik.Danfysik8500 method*), 77
 ramp_to_current () (*pymeasure.instruments.keithley.Keithley2400 method*), 89
 ramp_to_current () (*pymeasure.instruments.yokogawa.Yokogawa7651 method*), 105
 ramp_to_field () (*pymeasure.instruments.ami.AMI430 method*), 73
 ramp_to_voltage () (*pymeasure.instruments.keithley.Keithley2400 method*), 90
 ramp_to_voltage () (*pymeasure.instruments.yokogawa.Yokogawa7651 method*), 106
 range () (*pymeasure.instruments.fwbell.FWBell5080 property*), 79
 range () (*pymeasure.instruments.lakeshore.LakeShore425 property*), 94
 ratio () (*pymeasure.instruments.agilent.agilent4156.VARD property*), 65
 read () (*pymeasure.adapters.Adapter method*), 25
 read () (*pymeasure.adapters.FakeAdapter method*), 26
 read () (*pymeasure.adapters.PrologixAdapter method*), 29
 read () (*pymeasure.adapters.SerialAdapter method*), 27
 read () (*pymeasure.adapters.VISAAdapter method*), 30
 read () (*pymeasure.instruments.danfysik.DanfysikAdapter method*), 76
 read () (*pymeasure.instruments.fwbell.FWBell5080 method*), 79
 read () (*pymeasure.instruments.Instrument method*), 50
 read () (*pymeasure.instruments.lakeshore.LakeShoreUSBAdapter method*), 93
 read () (*pymeasure.instruments.parker.ParkerGV6 method*), 99
 read_bytes () (*pymeasure.adapters.VISAAdapter method*), 30
 read_memory () (*pymeasure.instruments.anritsu.AnritsuMS9710C method*), 75
 readAI () (*in module pymeasure.instruments.comedi*), 53
 receive () (*pymeasure.experiment.listeners.Listener method*), 34
 Recorder (*class in pymeasure.experiment.listeners*), 34
 reference () (*pymeasure.instruments.signalrecovery.DSP7265 property*), 100
 reference_source () (*pymeasure.instruments.srs.SR830 property*), 103
 refresh_parameters () (*pymeasure.experiment.procedure.Procedure method*), 35
 reload () (*pymeasure.experiment.results.Results method*), 39
 remote () (*pymeasure.instruments.danfysik.Danfysik8500 method*), 77

- remote() (*pymeasure.instruments.keithley.Keithley2000 method*), 84
- remote_local_state() (*pymeasure.instruments.agilent.Agilent33220A property*), 67
- remote_lock() (*pymeasure.instruments.keithley.Keithley2000 method*), 84
- remove() (*pymeasure.display.manager.Manager method*), 45
- reset() (*pymeasure.instruments.fwbell.FWBell5080 method*), 79
- reset() (*pymeasure.instruments.Instrument method*), 50
- reset() (*pymeasure.instruments.keithley.Keithley2000 method*), 84
- reset() (*pymeasure.instruments.keithley.Keithley2400 method*), 90
- reset() (*pymeasure.instruments.parker.ParkerGV6 method*), 99
- reset_buffer() (*pymeasure.instruments.keithley.Keithley2000 method*), 84
- reset_buffer() (*pymeasure.instruments.keithley.Keithley2400 method*), 90
- reset_interlocks() (*pymeasure.instruments.danfysik.Danfysik8500 method*), 77
- reset_time() (*pymeasure.instruments.Mock method*), 51
- resistance() (*pymeasure.instruments.agilent.Agilent34410A property*), 59
- resistance() (*pymeasure.instruments.keithley.Keithley2000 property*), 84
- resistance() (*pymeasure.instruments.keithley.Keithley2400 property*), 90
- resistance_4w() (*pymeasure.instruments.agilent.Agilent34410A property*), 60
- resistance_4W_digits() (*pymeasure.instruments.keithley.Keithley2000 property*), 84
- resistance_4W_nplc() (*pymeasure.instruments.keithley.Keithley2000 property*), 85
- resistance_4W_range() (*pymeasure.instruments.keithley.Keithley2000 property*), 85
- resistance_4W_reference() (*pymeasure.instruments.keithley.Keithley2000 property*), 85
- resistance_digits() (*pymeasure.instruments.keithley.Keithley2000 property*), 85
- resistance_nplc() (*pymeasure.instruments.keithley.Keithley2000 property*), 85
- resistance_nplc() (*pymeasure.instruments.keithley.Keithley2400 property*), 90
- resistance_range() (*pymeasure.instruments.keithley.Keithley2000 property*), 85
- resistance_range() (*pymeasure.instruments.keithley.Keithley2400 property*), 90
- resistance_reference() (*pymeasure.instruments.keithley.Keithley2000 property*), 85
- resolution() (*pymeasure.instruments.anritsu.AnritsuMS9710C property*), 75
- resolution_actual() (*pymeasure.instruments.anritsu.AnritsuMS9710C property*), 75
- resolution_vbw() (*pymeasure.instruments.anritsu.AnritsuMS9710C property*), 75
- Results (*class in pymeasure.experiment.results*), 39
- ResultsCurve (*class in pymeasure.display.curves*), 42
- ResultsDialog (*class in pymeasure.display.widgets*), 46
- resume() (*pymeasure.display.manager.Manager method*), 45
- right_limit() (*pymeasure.instruments.newport.esp300.Axis property*), 96
- run() (*pymeasure.display.listeners.Monitor method*), 43
- run() (*pymeasure.display.plotter.Plotter method*), 45
- run() (*pymeasure.experiment.workers.Worker method*), 38
- ## S
- sample_continuously() (*pymeasure.instruments.keithley.Keithley2400 method*), 90
- sample_frequency() (*pymeasure.instruments.srs.SR830 property*), 103
- sampling_points() (*pymeasure.instruments.anritsu.AnritsuMS9710C property*), 75
- save() (*pymeasure.instruments.agilent.agilent4156.Agilent4156 method*), 62

`save_var()` (*pymeasure.instruments.agilent.agilent4156.Agilent4156* method), 63
`scan()` (*pymeasure.instruments.agilent.Agilent8722ES* method), 57
`scan_continuous()` (*pymeasure.instruments.agilent.Agilent8722ES* method), 57
`scan_points()` (*pymeasure.instruments.agilent.Agilent8722ES* property), 57
`scan_single()` (*pymeasure.instruments.agilent.Agilent8722ES* method), 57
`ScientificInput` (class in *pymeasure.display.inputs*), 43
`sensitivity()` (*pymeasure.instruments.signalrecovery.DSP7265* property), 100
`sensitivity()` (*pymeasure.instruments.srs.SR830* property), 103
`sensor()` (*pymeasure.instruments.thorlabs.ThorlabsPM100USB* method), 104
`SerialAdapter` (class in *pymeasure.adapters*), 27
`series_resistance()` (*pymeasure.instruments.agilent.agilent4156.SMU* property), 64
`set_averaging()` (*pymeasure.instruments.agilent.Agilent8722ES* method), 57
`set_defaults()` (*pymeasure.adapters.PrologixAdapter* method), 29
`set_defaults()` (*pymeasure.instruments.parker.ParkerGV6* method), 99
`set_fixed_frequency()` (*pymeasure.instruments.agilent.Agilent8722ES* method), 57
`set_hardware_limits()` (*pymeasure.instruments.parker.ParkerGV6* method), 99
`set_IF_bandwidth()` (*pymeasure.instruments.agilent.Agilent8722ES* method), 57
`set_output_voltage()` (*pymeasure.instruments.Mock* method), 51
`set_parameter()` (*pymeasure.display.inputs.BooleanInput* method), 42
`set_parameter()` (*pymeasure.display.inputs.FloatInput* method), 42
`set_parameter()` (*pymeasure.display.inputs.Input* method), 42
`set_parameter()` (*pymeasure.display.inputs.IntegerInput* method), 43
`set_parameter()` (*pymeasure.display.inputs.ListInput* method), 43
`set_parameter()` (*pymeasure.display.inputs.ScientificInput* method), 43
`set_parameters()` (*pymeasure.display.windows.ManagedWindow* method), 47
`set_parameters()` (*pymeasure.experiment.procedure.Procedure* method), 35
`set_ramp_delay()` (*pymeasure.instruments.danfysik.Danfysik8500* method), 77
`set_ramp_to_current()` (*pymeasure.instruments.danfysik.Danfysik8500* method), 77
`set_scaling()` (*pymeasure.instruments.srs.SR830* method), 103
`set_sequence()` (*pymeasure.instruments.danfysik.Danfysik8500* method), 78
`set_software_limits()` (*pymeasure.instruments.parker.ParkerGV6* method), 99
`set_time()` (*pymeasure.instruments.Mock* method), 51
`set_timed_arm()` (*pymeasure.instruments.keithley.Keithley2400* method), 90
`set_trigger_counts()` (*pymeasure.instruments.keithley.Keithley2400* method), 90
`setDifferentialMode()` (*pymeasure.instruments.signalrecovery.DSP7265* method), 100
`setpoint_1()` (*pymeasure.instruments.lakeshore.LakeShore331* property), 93
`setpoint_2()` (*pymeasure.instruments.lakeshore.LakeShore331* property), 93
`setting()` (*pymeasure.instruments.Instrument* static method), 50
`setup_plot()` (*pymeasure.display.plotter.Plotter* method), 45
`setup_plot()` (*pymeasure.display.windows.ManagedWindow* method), 47
`shape()` (*pymeasure.instruments.agilent.Agilent33220A* property), 67

- shape () (*pymeasure.instruments.agilent.Agilent33500* property), 70
- shape () (*pymeasure.instruments.hp.HP33120A* property), 80
- shutdown () (*pymeasure.experiment.procedure.Procedure* method), 35
- shutdown () (*pymeasure.experiment.workers.Worker* method), 38
- shutdown () (*pymeasure.instruments.agilent.Agilent8257D* method), 56
- shutdown () (*pymeasure.instruments.ami.AMI430* method), 73
- shutdown () (*pymeasure.instruments.anritsu.AnritsuMG3692C* method), 73
- shutdown () (*pymeasure.instruments.Instrument* method), 51
- shutdown () (*pymeasure.instruments.keithley.Keithley2000* method), 85
- shutdown () (*pymeasure.instruments.keithley.Keithley2400* method), 90
- shutdown () (*pymeasure.instruments.newport.ESP300* method), 95
- shutdown () (*pymeasure.instruments.signalrecovery.DSP7265* method), 100
- shutdown () (*pymeasure.instruments.yokogawa.Yokogawa7651* method), 106
- sine_voltage () (*pymeasure.instruments.srs.SR830* property), 103
- single_sweep () (*pymeasure.instruments.anritsu.AnritsuMS9710C* method), 75
- sizeHint () (*pymeasure.display.widgets.PlotWidget* method), 46
- slew_rate () (*pymeasure.instruments.danfysik.Danfysik8500* property), 78
- slope () (*pymeasure.instruments.signalrecovery.DSP7265* property), 100
- SMU (*class in pymeasure.instruments.agilent.agilent4156*), 63
- source_current () (*pymeasure.instruments.keithley.Keithley2400* property), 90
- source_current () (*pymeasure.instruments.yokogawa.Yokogawa7651* property), 106
- source_current_range () (*pymeasure.instruments.keithley.Keithley2400* property), 90
- source_current_range () (*pymeasure.instruments.yokogawa.Yokogawa7651* property), 106
- source_enabled () (*pymeasure.instruments.keithley.Keithley2400* property), 90
- source_enabled () (*pymeasure.instruments.yokogawa.Yokogawa7651* property), 106
- source_mode () (*pymeasure.instruments.keithley.Keithley2400* property), 91
- source_mode () (*pymeasure.instruments.yokogawa.Yokogawa7651* property), 106
- source_voltage () (*pymeasure.instruments.keithley.Keithley2400* property), 91
- source_voltage () (*pymeasure.instruments.yokogawa.Yokogawa7651* property), 106
- source_voltage_range () (*pymeasure.instruments.keithley.Keithley2400* property), 91
- source_voltage_range () (*pymeasure.instruments.yokogawa.Yokogawa7651* property), 106
- spacing () (*pymeasure.instruments.agilent.agilent4156.VARI* property), 64
- square_dutycycle () (*pymeasure.instruments.agilent.Agilent33220A* property), 67
- square_dutycycle () (*pymeasure.instruments.agilent.Agilent33500* property), 71
- SR830 (*class in pymeasure.instruments.srs*), 101
- standard_devs () (*pymeasure.instruments.keithley.Keithley2400* property), 91
- start () (*pymeasure.experiment.experiment.Experiment* method), 34
- start () (*pymeasure.instruments.agilent.agilent4156.VARX* property), 65
- start_buffer () (*pymeasure.instruments.keithley.Keithley2000* method), 85
- start_buffer () (*pymeasure.instruments.keithley.Keithley2400* method), 91
- start_frequency () (*pymeasure.instruments.agilent.Agilent8257D* property), 56

`start_frequency()` (`pymeasure.instruments.agilent.Agilent8722ES` property), 57
`start_frequency()` (`pymeasure.instruments.agilent.AgilentE4408B` property), 58
`start_power()` (`pymeasure.instruments.agilent.Agilent8257D` property), 56
`start_ramp()` (`pymeasure.instruments.danfysik.Danfysik8500` method), 78
`start_sequence()` (`pymeasure.instruments.danfysik.Danfysik8500` method), 78
`start_step_sweep()` (`pymeasure.instruments.agilent.Agilent8257D` method), 56
`startup()` (`pymeasure.experiment.procedure.Procedure` method), 35
`startup()` (`pymeasure.experiment.procedure.UnknownProcedure` method), 35
`state()` (`pymeasure.instruments.ami.AMI430` property), 73
`status()` (`pymeasure.instruments.danfysik.Danfysik8500` property), 78
`status()` (`pymeasure.instruments.parker.ParkerGV6` property), 99
`status_hex()` (`pymeasure.instruments.danfysik.Danfysik8500` property), 78
`std_current()` (`pymeasure.instruments.keithley.Keithley2400` property), 91
`std_resistance()` (`pymeasure.instruments.keithley.Keithley2400` property), 91
`std_voltage()` (`pymeasure.instruments.keithley.Keithley2400` property), 91
`step()` (`pymeasure.instruments.agilent.agilent4156.VARX` property), 65
`step_points()` (`pymeasure.instruments.agilent.Agilent8257D` property), 56
`stepEnabled()` (`pymeasure.display.inputs.ScientificInput` method), 43
`stop()` (`pymeasure.instruments.agilent.agilent4156.Agilent4156` method), 63
`stop()` (`pymeasure.instruments.agilent.agilent4156.VARX` property), 65
`stop()` (`pymeasure.instruments.parker.ParkerGV6` method), 99
`stop_buffer()` (`pymeasure.instruments.keithley.Keithley2000` method), 85
`stop_buffer()` (`pymeasure.instruments.keithley.Keithley2400` method), 91
`stop_frequency()` (`pymeasure.instruments.agilent.Agilent8257D` property), 56
`stop_frequency()` (`pymeasure.instruments.agilent.Agilent8722ES` property), 57
`stop_frequency()` (`pymeasure.instruments.agilent.AgilentE4408B` property), 58
`stop_ramp()` (`pymeasure.instruments.danfysik.Danfysik8500` method), 78
`stop_sequence()` (`pymeasure.instruments.danfysik.Danfysik8500` method), 78
`stop_step_sweep()` (`pymeasure.instruments.agilent.Agilent8257D` method), 56
`StoppableQThread` (class in `pymeasure.display.thread`), 46
`StringInput` (class in `pymeasure.display.inputs`), 43
`supply_current()` (`pymeasure.instruments.ami.AMI430` property), 73
`sweep_status()` (`pymeasure.instruments.oxfordinstruments.ITC503` property), 97
`sweep_table()` (`pymeasure.instruments.oxfordinstruments.ITC503` property), 97
`sweep_time()` (`pymeasure.instruments.agilent.Agilent8722ES` property), 57
`sweep_time()` (`pymeasure.instruments.agilent.AgilentE4408B` property), 58
`sync_sequence()` (`pymeasure.instruments.danfysik.Danfysik8500` method), 78

T

`target_current()` (`pymeasure.instruments.ami.AMI430` property), 73
`target_field()` (`pymeasure.instruments.ami.AMI430` property), 73

- TDS2000 (class in *pymeasure.instruments.tektronix*), 104
- temperature() (*pymeasure.instruments.keithley.Keithley2000* property), 85
- temperature_1() (*pymeasure.instruments.oxfordinstruments.ITC503* property), 97
- temperature_2() (*pymeasure.instruments.oxfordinstruments.ITC503* property), 97
- temperature_A() (*pymeasure.instruments.lakeshore.LakeShore331* property), 93
- temperature_B() (*pymeasure.instruments.lakeshore.LakeShore331* property), 94
- temperature_digits() (*pymeasure.instruments.keithley.Keithley2000* property), 85
- temperature_error() (*pymeasure.instruments.oxfordinstruments.ITC503* property), 98
- temperature_nplc() (*pymeasure.instruments.keithley.Keithley2000* property), 85
- temperature_reference() (*pymeasure.instruments.keithley.Keithley2000* property), 85
- temperature_setpoint() (*pymeasure.instruments.oxfordinstruments.ITC503* property), 98
- textFromValue() (*pymeasure.display.inputs.ScientificInput* method), 43
- theta() (*pymeasure.instruments.srs.SR830* property), 103
- ThorlabsPM100USB (class in *pymeasure.instruments.thorlabs*), 104
- time() (*pymeasure.instruments.Mock* property), 51
- time_constant() (*pymeasure.instruments.signalrecovery.DSP7265* property), 101
- time_constant() (*pymeasure.instruments.srs.SR830* property), 103
- trace() (*pymeasure.instruments.agilent.AgilentE4408B* method), 58
- trace_df() (*pymeasure.instruments.agilent.AgilentE4408B* method), 58
- trace_marker() (*pymeasure.instruments.anritsu.AnritsuMS9710C* property), 75
- trace_marker_center() (*pymeasure.instruments.anritsu.AnritsuMS9710C* property), 75
- triad() (*pymeasure.instruments.keithley.Keithley2400* method), 91
- trigger() (*pymeasure.instruments.agilent.Agilent33220A* method), 67
- trigger() (*pymeasure.instruments.agilent.Agilent33500* method), 71
- trigger() (*pymeasure.instruments.keithley.Keithley2400* method), 91
- trigger_count() (*pymeasure.instruments.keithley.Keithley2000* property), 85
- trigger_count() (*pymeasure.instruments.keithley.Keithley2400* property), 91
- trigger_delay() (*pymeasure.instruments.keithley.Keithley2000* property), 86
- trigger_delay() (*pymeasure.instruments.keithley.Keithley2400* property), 91
- trigger_immediately() (*pymeasure.instruments.keithley.Keithley2400* method), 91
- trigger_on_bus() (*pymeasure.instruments.keithley.Keithley2400* method), 91
- trigger_on_external() (*pymeasure.instruments.keithley.Keithley2400* method), 91
- trigger_source() (*pymeasure.instruments.agilent.Agilent33220A* property), 67
- trigger_source() (*pymeasure.instruments.agilent.Agilent33500* property), 71
- trigger_source() (*pymeasure.instruments.agilent.AgilentE4980* property), 59
- trigger_state() (*pymeasure.instruments.agilent.Agilent33220A* property), 67
- ## U
- unique_filename() (in module *pymeasure.experiment.results*), 39
- unit() (*pymeasure.instruments.lakeshore.LakeShore425* property), 95
- units() (*pymeasure.instruments.fwbell.FWBell5080* property), 79
- units() (*pymeasure.instruments.newport.esp300.Axis* property), 96

- UnknownProcedure (class in *pymeasure.experiment.procedure*), 35
- update() (*pymeasure.display.curves.Crosshairs* method), 42
- update() (*pymeasure.display.curves.ResultsCurve* method), 42
- update_line() (*pymeasure.experiment.experiment.Experiment* method), 34
- update_parameter() (*pymeasure.display.inputs.Input* method), 42
- update_pcolor() (*pymeasure.experiment.experiment.Experiment* method), 34
- update_plot() (*pymeasure.experiment.experiment.Experiment* method), 34
- update_status() (*pymeasure.experiment.workers.Worker* method), 38
- use_absolute_position() (*pymeasure.instruments.parker.ParkerGV6* method), 99
- use_front_terminals() (*pymeasure.instruments.keithley.Keithley2400* method), 91
- use_rear_terminals() (*pymeasure.instruments.keithley.Keithley2400* method), 92
- use_relative_position() (*pymeasure.instruments.parker.ParkerGV6* method), 99
- ## V
- validate() (*pymeasure.display.inputs.ScientificInput* method), 43
- valueFromText() (*pymeasure.display.inputs.ScientificInput* method), 43
- values() (*pymeasure.adapters.Adapter* method), 25
- values() (*pymeasure.adapters.FakeAdapter* method), 26
- values() (*pymeasure.adapters.PrologixAdapter* method), 29
- values() (*pymeasure.adapters.SerialAdapter* method), 27
- values() (*pymeasure.adapters.VISAAadapter* method), 30
- values() (*pymeasure.instruments.fwbell.FWBell5080* method), 80
- values() (*pymeasure.instruments.Instrument* method), 51
- values() (*pymeasure.instruments.lakeshore.LakeShoreUSBAdapter* method), 93
- values() (*pymeasure.instruments.signalrecovery.DSP7265* method), 101
- VAR1 (class in *pymeasure.instruments.agilent.agilent4156*), 64
- VAR2 (class in *pymeasure.instruments.agilent.agilent4156*), 64
- VAR3 (class in *pymeasure.instruments.agilent.agilent4156*), 64
- VAR4 (class in *pymeasure.instruments.agilent.agilent4156*), 64
- VAR5 (class in *pymeasure.instruments.agilent.agilent4156*), 65
- VectorParameter (class in *pymeasure.experiment.parameters*), 38
- VISAAadapter (class in *pymeasure.adapters*), 29
- VMU (class in *pymeasure.instruments.agilent.agilent4156*), 65
- voltage() (*pymeasure.instruments.keithley.Keithley2000* property), 86
- voltage() (*pymeasure.instruments.keithley.Keithley2400* property), 92
- voltage() (*pymeasure.instruments.Mock* property), 51
- voltage() (*pymeasure.instruments.signalrecovery.DSP7265* property), 101
- voltage_ac() (*pymeasure.instruments.agilent.Agilent34410A* property), 60
- voltage_ac_bandwidth() (*pymeasure.instruments.keithley.Keithley2000* property), 86
- voltage_ac_digits() (*pymeasure.instruments.keithley.Keithley2000* property), 86
- voltage_ac_nplc() (*pymeasure.instruments.keithley.Keithley2000* property), 86
- voltage_ac_range() (*pymeasure.instruments.keithley.Keithley2000* property), 86
- voltage_ac_reference() (*pymeasure.instruments.keithley.Keithley2000* property), 86
- voltage_dc() (*pymeasure.instruments.agilent.Agilent34410A* property), 60
- voltage_digits() (*pymeasure.instruments.keithley.Keithley2000* property), 86
- voltage_high() (*pymeasure.instruments.agilent.Agilent33220A* property), 67
- voltage_high() (*pymeasure.instruments.agilent.Agilent33500* property), 71
- voltage_limit() (*pymeasure.instruments.ami.AMI430* property),

- 73
- `voltage_low()` (*pymeasure.instruments.agilent.Agilent33220A property*), 68
- `voltage_low()` (*pymeasure.instruments.agilent.Agilent33500 property*), 71
- `voltage_name()` (*pymeasure.instruments.agilent.agilent4156.SMU property*), 64
- `voltage_name()` (*pymeasure.instruments.agilent.agilent4156.VMU property*), 65
- `voltage_name()` (*pymeasure.instruments.agilent.agilent4156.VSU property*), 66
- `voltage_nplc()` (*pymeasure.instruments.keithley.Keithley2000 property*), 86
- `voltage_nplc()` (*pymeasure.instruments.keithley.Keithley2400 property*), 92
- `voltage_range()` (*pymeasure.instruments.keithley.Keithley2000 property*), 86
- `voltage_range()` (*pymeasure.instruments.keithley.Keithley2400 property*), 92
- `voltage_reference()` (*pymeasure.instruments.keithley.Keithley2000 property*), 86
- VSU (class in *pymeasure.instruments.agilent.agilent4156*), 65
- W**
- `wait()` (*pymeasure.instruments.anritsu.AnritsuMS9710C method*), 75
- `wait_for_buffer()` (*pymeasure.instruments.keithley.Keithley2000 method*), 86
- `wait_for_buffer()` (*pymeasure.instruments.keithley.Keithley2400 method*), 92
- `wait_for_buffer()` (*pymeasure.instruments.srs.SR830 method*), 103
- `wait_for_current()` (*pymeasure.instruments.danfysik.Danfysik8500 method*), 78
- `wait_for_data()` (*pymeasure.experiment.experiment.Experiment method*), 34
- `wait_for_holding()` (*pymeasure.instruments.ami.AMI430 method*), 73
- `wait_for_ready()` (*pymeasure.instruments.danfysik.Danfysik8500 method*), 78
- `wait_for_srq()` (*pymeasure.adapters.PrologixAdapter method*), 29
- `wait_for_srq()` (*pymeasure.adapters.VISAAdapter method*), 30
- `wait_for_stop()` (*pymeasure.instruments.newport.esp300.Axis method*), 96
- `wait_for_sweep()` (*pymeasure.instruments.anritsu.AnritsuMS9710C method*), 75
- `wait_for_temperature()` (*pymeasure.instruments.lakeshore.LakeShore331 method*), 94
- `wait_for_temperature()` (*pymeasure.instruments.oxfordinstruments.ITC503 method*), 98
- `wait_for_trigger()` (*pymeasure.instruments.agilent.Agilent33220A method*), 68
- `wait_for_trigger()` (*pymeasure.instruments.agilent.Agilent33500 method*), 71
- `wave()` (*pymeasure.instruments.Mock property*), 51
- `wavelength()` (*pymeasure.instruments.thorlabs.ThorlabsPM100USB property*), 104
- `wavelength_center()` (*pymeasure.instruments.anritsu.AnritsuMS9710C property*), 75
- `wavelength_marker_value()` (*pymeasure.instruments.anritsu.AnritsuMS9710C property*), 75
- `wavelength_max()` (*pymeasure.instruments.thorlabs.ThorlabsPM100USB property*), 104
- `wavelength_min()` (*pymeasure.instruments.thorlabs.ThorlabsPM100USB property*), 104
- `wavelength_span()` (*pymeasure.instruments.anritsu.AnritsuMS9710C property*), 75
- `wavelength_start()` (*pymeasure.instruments.anritsu.AnritsuMS9710C property*), 75
- `wavelength_stop()` (*pymeasure.instruments.anritsu.AnritsuMS9710C property*), 75
- `wavelength_value_in()` (*pymeasure.instruments.anritsu.AnritsuMS9710C property*), 75

wavelengths() (*pymeasure.instruments.anritsu.AnritsuMS9710C* property), 75

wires() (*pymeasure.instruments.keithley.Keithley2400* property), 92

Worker (*class in pymeasure.experiment.workers*), 38

write() (*pymeasure.adapters.Adapter* method), 26

write() (*pymeasure.adapters.FakeAdapter* method), 26

write() (*pymeasure.adapters.PrologixAdapter* method), 29

write() (*pymeasure.adapters.SerialAdapter* method), 27

write() (*pymeasure.adapters.VISAAdapter* method), 30

write() (*pymeasure.instruments.danfysik.DanfysikAdapter* method), 76

write() (*pymeasure.instruments.Instrument* method), 51

write() (*pymeasure.instruments.lakeshore.LakeShoreUSBAdapter* method), 93

write() (*pymeasure.instruments.parker.ParkerGV6* method), 99

writeAO() (*in module pymeasure.instruments.comedi*), 53

X

x() (*pymeasure.instruments.signalrecovery.DSP7265* property), 101

x() (*pymeasure.instruments.srs.SR830* property), 103

xpointer() (*pymeasure.instruments.oxfordinstruments.ITC503* property), 98

xy() (*pymeasure.instruments.signalrecovery.DSP7265* property), 101

Y

y() (*pymeasure.instruments.signalrecovery.DSP7265* property), 101

y() (*pymeasure.instruments.srs.SR830* property), 103

Yokogawa7651 (*class in pymeasure.instruments.yokogawa*), 105

ypointer() (*pymeasure.instruments.oxfordinstruments.ITC503* property), 98

Z

zero() (*pymeasure.instruments.ami.AMI430* method), 73

zero() (*pymeasure.instruments.newport.esp300.Axis* method), 96

zero_probe() (*pymeasure.instruments.lakeshore.LakeShore425* method), 95